



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)

Cert. No. : CC-2239



CALIBRATION CERTIFICATE

Customer Name And Address	M/S: Ashok Nursing Home. F-3/13-16, Vijay Chowk, Krishan Nagar, Delhi-110051.	Certificate No.	CC223920600023524 F
		Service request No & date	P/01/20/08/2020
		Date of calibration	20/08/2020
		Valid up to	19/08/2021
		Date of issue	21/08/2020

Instrument Details

Instrument name	Digital Thermometer	S. No.	----
Make/Model	Multi-Thermometer	Location	Refrigerator
Range / Size	-50 to 300 °C	Accuracy	----
Least Count	0.1 °C	Visual Inspection	OK
I.D. No.	DTM-02		

Detail of reference standards & Major equipments used

Equipment Name	RTD With Digital Indicator		
Make	YCT / SBI		
Model / SR No.	11080368		
Certificate No.	TL/020/181.2.1		
Calibration Validity	01/03/2021		
Calibration by	Tempens		

Environmental Condition	Temperature	25±4 °C	Calibration Reference	IS:6274
	Relative Humidity	50±20 %	Calibration Procedure	CP-02T

Calibration Results

Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	-17.45	-16.6	±0.3 °C
02.	-15.32	-14.5	
03.	-10.22	-9.8	
04.	-8.15	-7.6	
05.	0.22	0.3	
06.	30.05	29.4	
07.	50.35	50.5	
08.	99.55	100.1	
09.	149.64	150.50	

Remarks:

<ul style="list-style-type: none"> ❖ This result of calibration refers to only to the particular item submitted for calibration. ❖ The above result are valid at the time of and under the stated conditions measurement. ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By
(Calibration Engg./TM)
(D. B. SAXENA)
Form No.-32



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End Of Report---



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)

Cert. No. : CC-2239



CALIBRATION CERTIFICATE

Customer Name And Address	M/S: Ashok Nursing Home. F-3/13-16, Vijay Chowk, Krishan Nagar, Delhi-110051.	Certificate No.	CC223920600023532 F
		Service request No & date	P/01/20/08/2020
		Date of calibration	20/08/2020
		Valid up to	19/08/2021
		Date of issue	21/08/2020

Instrument Details			
Instrument name	Digital Thermometer (Multi)	S. No.	-----
Make/Model	Multi-Thermometer	Location	Refrigerator (Path Lab)
Range / Size	-50 to 300 °C	Accuracy	-----
Least Count	0.1 °C	Visual Inspection	OK
I.D. No.	DTM-03		

Detail of reference standards & Major equipments used			
Equipment Name	RTD With Digital Indicator		
Make	YCT / SBI		
Model / SR No.	11080368		
Certificate No.	TL/020/181.2.1		
Calibration Validity	01/03/2021		
Calibration by	Tempsens		

Environmental Condition	Temperature	25±4 °C	Calibration Reference	IS:6274
	Relative Humidity	50±20 %	Calibration Procedure	CP-02T

Calibration Results			
Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	-17.35	-16.5	±0.3 °C
02.	-15.46	-14.5	
03.	-10.22	-9.0	
04.	0.10	0.5	
05.	50.25	50.0	
06.	99.44	100.0	
07.	149.56	150.0	
08.	192.56	190.0	

Remarks:

- ❖ This result of calibration refers to only to the particular item submitted for calibration.
- ❖ The above result are valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By
(Calibration Engg./TM)
(D.B. SAXENA)
Form No.-32



Approved By
(DM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End Of Report---