



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)

Cert. No. : CC-2239



Certificate No.
CC-2239

CALIBRATION CERTIFICATE

Customer Name And Address	M/S: Ashok Nursing Home. F-3/13-16, Vijay Chowk, Krishan Nagar, Delhi-110051.	Certificate No.	CC223920600023525 F
		Service request No & date	P/01/20/08/2020
		Date of calibration	20/08/2020
		Valid up to	19/08/2021
		Date of issue	21/08/2020

Instrument Details			
Instrument name	Digital Thermo Hygrometer	Sr. No.	-----
Make / Model No.	HTC-1	Location	Path Lab
Range / Size	-10 to 50 °C / 10 to 99 %RH	Accuracy	-----
Least Count	0.1 °C / 1 %RH	Visual Inspection	OK
I.D. No.	DTH-01		

Detail of reference standards & Major equipments used			
Equipment Name	RTD With Digital Indicator	Dig. Humidity Indicator With Sensor	
Make	YCT / SBI	Yudian / Rotronic	
Model / SR No.	11080368	11F183060/20274168	
Certificate No.	TL/020/181.2.1	FL/CAL/TH/07112019-001	
Calibration Validity	01/03/2021	11/11/2020	
Calibration by	Tempsens	FARELABS	

Environmental Condition	Temperature	25± 4 °C	Calibration Reference	ITS:90
	Relative Humidity	50±20%	Calibration Procedure	CP-04T

Calibration Results

Serial No.	Standard Average Value in (°C)	Measured Average Value in (°C)	Standard Average Value in (%RH)	Measured Average Value in (%RH)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	-10.10	-10.0	31.2	30	
02.	10.15	10.5	49.5	50	± 0.21°C
03.	20.23	20.6	68.4	70	
04.	30.35	29.7	78.6	80	± 1.9%RH
05.	45.42	45.0	91.5	90	

Remarks:

- ❖ These results of calibration refer to only to the particular item submitted for calibration.
- ❖ The above results are valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By
(Calibration Engg./TM)
(D.B SAXENA)
Form No.-32



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End Of Report---