



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)

Cert. No. : CC-2239



CALIBRATION CERTIFICATE

Customer Name And Address	M/S: Ashok Nursing Home. F-3/13-16, Vijay Chowk, Krishan Nagar, Delhi-110051.	Certificate No.	CC223920600023536 F
		Service request No & date	P/01/20/08/2020
		Date of calibration	20/08/2020
		Valid up to	19/08/2021
		Date of issue	21/08/2020

Instrument Details			
Instrument name	Pipette	Sr. No.	NL 485257
Make	Durapet	Location	Path Lab
Range / Size	5 to 50 μ l	Accuracy	-----
Least Count	5 μ l	Visual Inspection	OK
I.D. No.	P-02		

Detail of reference standards & Major equipments used			
Equipment Name	A set of weight box	Digital Weighing Balance	
Make	Weigh India	AND	
Model / SR No.	-----	GH-252	
Certificate No.	WMCL/E/2018-12/2533	CC205320000000016F	
Calibration Validity	26/12/2021	16/02/2021	
Calibration by	WEIGHTRONICS	Kelvin Research Centre	


Environmental Condition	Temperature	23 \pm 3 $^{\circ}$ C	Calibration Reference	IS:1997
	Relative Humidity	50 \pm 20%	Calibration Procedure	CP-33

Calibration Results			
Serial No.	Nominal Capacity in (μ l)	Measured Capacity in (μ l)	Uncertainty At 95% C.L. (coverage factor $k=2$)
01.	10.0	10.02125	$\pm 0.10 \mu$ l
02.	20.0	19.98945	
03.	30.0	29.97505	
04.	40.0	39.96912	
05.	50.0	49.96245	

Remarks:

- ❖ This result of calibration refer to only to the particular item submitted for calibration.
- ❖ The above result are valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By
(Calibration Engg./TM)
(MAHUPAL)
Form No.-32



Approved By
(OM/TM)
(NEERAJ TIAGI)
Page No. 1 of 1

---End Of Report---