



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)

Cert. No. : CC-2239



CALIBRATION CERTIFICATE

Customer Name And Address	M/S: Ashok Nursing Home. F-3/13-16, Vijay Chowk, Krishan Nagar, Delhi-110051.	Certificate No.	CC223920600023528 F
		Service request No & date	P/01/20/08/2020
		Date of calibration	20/08/2020
		Valid up to	19/08/2021
		Date of issue	21/08/2020

Instrument Details			
Instrument name	Micro Pipette	Sr. No.	DY13314
Make / Model No.	Erba	Location	Path Lab
Range / Size	100 to 1000 µl	Accuracy	-----
Least Count	5 µl	Visual Inspection	OK
I.D. No.	MP-02		

Detail of reference standards & Major equipments used			
Equipment Name	A set of weight box	Digital Weighing Balance	
Make	Weigh India	AND	
Model / SR No.	-----	GH-252	
Certificate No.	WMCL/E/2018-12/2533	CC205320000000016F	
Calibration Validity	26/12/2021	16/02/2021	
Calibration by	WEIGHTRONICS	Kelvin Research Centre	

Environmental Condition	Temperature	23± 3 °C	Calibration Reference	IS:1997
	Relative Humidity	50±20%	Calibration Procedure	CP-33

Calibration Results			
Serial No.	Nominal Capacity in (µl)	Measured Capacity in (µl)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	100	100.10	±0.46 µl
02.	300	298.62	
03.	500	499.85	
04.	800	798.75	
05.	1000	999.65	

Remarks:

<ul style="list-style-type: none"> ❖ This result of calibration refer to only to the particular item submitted for calibration. ❖ The above result are valid at the time of and under the stated conditions measurement. ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.
--

Calibrated By
(Calibration Engg./TM)
(MAHIPAL)
Form No.-32



Approved By
(M/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End Of Report---