



RML – Quality Assurance Program (RML – QAP)



Lab Code No. 1770

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 10

ROUND: 2

TOTAL PARTICIPANT: 266

DATE: 30/03/2021

Parameter	Total Responses	Your Result	All Lab Result	%	Remarks
Anti HCV	234	Non-Reactive	Reactive : 0 Non-Reactive : 234	100%	Within Consensus
Typhoid	226	Negative	Positive : 8 Negative : 218	96%	Within Consensus

Chief Coordinator

Dr. Sanjay Mehrotra

Programme Director

Dr. Bandana Mehrotra

Checked By:

****End of Report****

