

## LAB MONTHLY SUMMARY



Lab Name **CURIS DIAGNOSTICS PVT LTD** Lab No **10193**  
 Month **June** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 15/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	940	193.80	10.52	20.38	187.5 mg/dl	-0.31	1.33
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	580	74.92	14.27	10.69	78.7 mg/dl	0.35	0.89
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	583	4.32	13.82	0.60	5.19 mg/dl	1.46	0.05
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	800	3.86	16.46	0.64	4.3 mg/dl	0.69	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	729	5.51	11.11	0.61	5.1 g/dl	-0.67	0.05
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	703	3.28	10.79	0.35	2.9 g/dl	-1.07	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	756	5.04	22.70	1.14	2.7 mg/dl	-2.04	0.08
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	906	109.30	12.80	13.99	94.7 mg/dl	-1.04	0.93
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	774	139.56	14.86	20.74	137.9 mg/dl	-0.08	1.49

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

