



# HYTECH INSTRUMENT

2SF, C-2 Shriram Complex, Nyay Khand-I, Indirapuram, Ghaziabad-201014 (U.P) INDIA  
Ph. No. +91-9899380410, +91-9716668085, +91-7669021291-95

## Calibration Certificate

Format No. : HY/FM/33

Page No. : 01 of 02

<b>Certificate No.</b>	HY/21/1121-2	<b>CALIBRATION CERTIFICATE OF</b>  <b>Hematology Analyzer</b>
<b>Date of Calibration :</b>	10.08.2021	
<b>Next Date for Calibration</b> (As per agreed with the Customer)	10.08.2022	
<b>Certificate Issue Date</b>	11.08.2021	

<b>CUSTOMER ADDRESS</b>		<b>DESCRIPTION OF DEVICE UNDER CALIBRATION</b>	
M/s Curis Diagnostics F-24/25, Sector-3, Rohini, New Delhi -110085 INDIA		<b>Make / Model</b>	Accurek/Acculab
<b>SERVICE REQUEST FORM DETAILS</b>		<b>Instrument Sr. No.</b>	2900Pet02683
		<b>Resolution / Least Count</b>	As Per Range
<b>Service Request / Job No.</b>	1121/1121-2	<b>Range / Size</b>	As Per Instrument
<b>Service Request Dated</b>	09.08.2021	<b>Asset Code / I.D No.</b>	CD/H.A/LAB/01
		<b>Location</b>	LAB
		<b>Instrument Condition</b>	In working

<b>Calibration Procedure</b>	HY/WI/2MD08	<b>Calibration Performed At</b>	At Site	
<b>Reference Standard</b>	NABL-126	<b>Discipline</b>	Medical	
<b>ENVIRONMENTAL CONDITIONS</b>	<b>Temperature</b>	(25 ± 15) °C	<b>Relative Humidity</b>	(60 ± 15) %RH

### DETAILS OF REFERENCE STANDARDS AND MAJOR EQUIPMENTS USED FOR CALIBRATION

Sr. No	Instrument Details	Certificate No.	Calibrated By	Calibration Due Date
1	Process Calibrator	CC20552000004619F	TNC	17.03.2022
2	Digital Multimeter with PRT Sensor	CC28120000001496F	T MPL	08.10.2021
3	Std. Weights	CC205521000004621F	TNC	17.03.2022
4	Digital Weighing Machine	CC205520000004620F	TNC	17.03.2022
5	Electrical Safety Analyzer	CC205520000000755F	TNC	18.01.2022

The Standards used for calibration are traceable to National Standards.

### CALIBRATION RESULT

Maxima Found at	Standard Wavelength	Tolerance Limit
360nm	360.50nm	(upto 400 nm)±1nm
419nm	418.80nm	(400 upto 600 nm)±3nm
453nm	455.90nm	(400 upto 600 nm)±3nm
537nm	537.5nm	(400 upto 600 nm)±3nm
640nm	639.50nm	(600 upto 1000 nm)±4nm

PARAMETER	DUC VALUE (°C)	STANDARD VALUE (°C)	ERROR (°C)	Uncertainty ± (%)
Temperature	20.1	20.002	0.098	0.95
	25.1	25.004	0.096	0.95
	28.2	28.009	0.191	0.95
	30.2	30.011	0.189	0.95
	35.3	35.018	0.282	0.95

Expanded Uncertainty of measurment at approximately 95% Confidence Level with coverage factor k=2

- Note :-**
1. DUC = Device Under Calibration
  2. The Certificate refers only to the particular items submitted for Calibration.
  3. This Certificate shall not be reproduced, except in full, without permission of Hytech Instrument, Ghaziabad.
  4. Result Reported are Valid at the time of and under the stated conditions of measurement.
  5. Calibration Certificate issue for weight & Measure parameters like Mass, Balance, Volumetric equipment, Measuring Scale/Tapes etc. for scientific purpose only and should not be used for Trade / Commerical use.



**CALIBRATED BY**  
(Calib. Engineer)

**APPROVED BY**  
(QM/TM)





# HYTECH INSTRUMENT

2SF, C-2 Shriram Complex, Nyay Khand-I, Indirapuram, Ghaziabad-201014 (U.P) INDIA  
Ph. No. +91-9899380410, +91-9716668085, +91-7669021291-95

## Calibration Certificate

Format No. : HY/FM/33

Page No. : 02 of 02

<b>Certificate No.</b>	<b>HY/21/1121-2</b>	<b>CALIBRATION CERTIFICATE OF</b>  <b>Hematology Analyzer</b>
<b>Date of Calibration :</b>	10.08.2021	
<b>Next Date for Calibration</b>	10.08.2022	
(As per agreed with the Customer)		
<b>Certificate Issue Date</b>	11.08.2021	

### CALIBRATION RESULT

PARAMETER	DUC VALUE ( $\mu$ l)	STANDARD VALUE ( $\mu$ l)	ERROR ( $\mu$ l)	Uncertainty $\pm$ ( $\mu$ l)
Volume	200	200.07	-0.07	4.0
	500	500.13	-0.13	4.0

### 1. Electrical Safety

#### Visual Test


Sr. No	TESTS	REMARKS
1	Power Chords, Cable Check	OK
2	Main Socket Check	OK
3	Equipment Type (B/BF/CF)	B
4	Equipment Class (I/II/III)	I

B-BODY TYPE, BF-BODY FLOAT TYPE, CF-CARDIAC FLOAT TYPE, I-PROPERLY EARTHED, II-DOUBLE INSULATED  
III-WITH EXTRA SAFETY LOW VOLTAGE.

### 2. Electrical Safety Test

Sr. No	PARAMETERS	Measured Value	REMARKS
1	Voltage Between Line & Neutral (V in)	236.4 V	OK
2	Voltage Between Line & Earth (V ie)	243.6 V	OK
3	Voltage between Neutral & Earth (V en)	2.2 V	OK
4	Load Current (I a)	0.5 A	OK
5	Earth Leakage Current (I L)	7.8 $\mu$ A	OK
6	Enclosure Current	0.6 $\mu$ A	OK
7	Patient Leakage Current (P L)	0.7 $\mu$ A	OK

Expanded Uncertainty of measurment at approximately 95% Confidence Level with coverage factor  $k=2$

<p><b>Note :-</b></p> <ol style="list-style-type: none"> <li>DUC = Device Under Calibration</li> <li>The Certificate refers only to the particular items submitted for Calibration.</li> <li>This Certificate shall not be reproduced, except in full, without permission of Hytech Instrument, Ghaziabad.</li> <li>Result Reported are Valid at the time of and under the stated conditions of measurement.</li> <li>Calibration Certificate issue for weight &amp; Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scale/Tapes etc. for scientific purpose only and should not be used for Trade / Commerical use.</li> </ol>		<p><b>CALIBRATED BY</b> (Calib. Engineer)</p> <p><b>APPROVED BY</b> (QM / TM)</p>
--	---	---





# HYTECH INSTRUMENT

2SF, C-2 Shriram Complex, Nyay Khand-I, Indirapuram, Ghaziabad-201014 (U.P) INDIA  
Ph. No. +91-9899380410, +91-9716668085, +91-7669021291-95

## Calibration Certificate

Format No. : HY/FM/33

Page No. : 01 of 02

<b>Certificate No.</b>	<b>HY/21/1121-1</b>	<b>CALIBRATION CERTIFICATE OF</b>  <b>Bio-Chemistry Analyzer</b>
<b>Date of Calibration :</b>	10.08.2021	
<b>Next Date for Calibration</b>	10.08.2022	
(As per agreed with the Customer)		
<b>Certificate Issue Date</b>	11.08.2021	

<b>CUSTOMER ADDRESS</b>		<b>DESCRIPTION OF DEVICE UNDER CALIBRATION</b>	
M/s Curis Diagnostics F-24/25, Sector-3, Rohini, New Delhi -110085 INDIA		<b>Make / Model</b>	Microlab In shine /RX-50V
<b>SERVICE REQUEST FORM DETAILS</b>		<b>Instrument Sr. No.</b>	1711963
		<b>Resolution / Least Count</b>	As Per Range
<b>Service Request / Job No.</b>	1121/1121-1	<b>Range / Size</b>	As Per Instrument
<b>Service Request Dated</b>	09.08.2021	<b>Asset Code / I.D No.</b>	CD/BIO-CHEM/LAB/01
		<b>Location</b>	LAB
		<b>Instrument Condition</b>	In working

<b>Calibration Procedure</b>	HY/WI/2MD08	<b>Calibration Performed At</b>	At Site	
<b>Reference Standard</b>	NABL-126	<b>Discipline</b>	Medical	
<b>ENVIRONMENTAL CONDITIONS</b>	<b>Temperature</b>	(25 ± 15) °C	<b>Relative Humidity</b>	(60 ± 15) %RH

### DETAILS OF REFERENCE STANDARDS AND MAJOR EQUIPMENTS USED FOR CALIBRATION

Sr. No	Instrument Details	Certificate No.	Calibrated By	Calibration Due Date
1	Process Calibrator	CC205520000004619F	TNC	17.03.2022
2	Digital Multimeter with PRT Sensor	CC281200000001496F	T MPL	08.10.2021
3	Std. Weights	CC205521000004621F	TNC	17.03.2022
4	Digital Weighing Machine	CC205520000004620F	TNC	17.03.2022
5	Electrical Safety Analyzer	CC205520000000755F	TNC	18.01.2022

The Standards used for calibration are traceable to National Standards.

### CALIBRATION RESULT

Maxima Found at	Standard Wavelength	Tolerance Limit
360nm	360.50nm	(upto 400 nm)±1nm
419nm	418.80nm	(400 upto 600 nm)±3nm
453nm	453.90nm	(400 upto 600 nm)±3nm
537nm	536.40nm	(400 upto 600 nm)±3nm
640nm	638.30nm	(600 upto 1000 nm)±4nm

PARAMETER	DUC VALUE (°C)	STANDARD VALUE (°C)	ERROR (°C)	Uncertainty ± (%)
Temperature	20.1	20.002	0.098	0.95
	25.1	25.006	0.094	0.95
	28.2	28.009	0.191	0.95
	30.2	30.012	0.188	0.95
	35.3	35.018	0.282	0.95

Expanded Uncertainty of measurement at approximately 95% Confidence Level with coverage factor  $k=2$

<b>Note :-</b>	1. DUC = Device Under Calibration		<b>CALIBRATED BY</b> (Calib. Engineer)
	2. The Certificate refers only to the particular items submitted for Calibration.		<b>APPROVED BY</b> (QM / TM)
	3. This Certificate shall not be reproduced, except in full, without permission of CEO Hytech Instrument, Ghaziabad.		
	4. Result Reported are Valid at the time of and under the stated conditions of measurement.		
	5. Calibration Certificate issue for weight & Measure parameters i.e. Mass, Force, Volumetric equipment, Measuring Scale/Tapes etc. for scientific purpose only and should not be used for Trade / Commercial use.		





# HYTECH INSTRUMENT

2SF, C-2 Shriram Complex, Nyay Khand-I, Indirapuram, Ghaziabad-201014 (U.P) INDIA  
Ph. No. +91-9899380410, +91-9716668085, +91-7669021291-95

## Calibration Certificate

Format No. : HY/FM/33

Page No. : 02 of 02

<b>Certificate No.</b>	<b>HY/21/1121-1</b>	<b>CALIBRATION CERTIFICATE OF</b>  <b>Bio-Chemistry Analyzer</b>
<b>Date of Calibration :</b>	10.08.2021	
<b>Next Date for Calibration</b> (As per agreed with the Customer)	10.08.2022	
<b>Certificate Issue Date</b>	11.08.2021	

### CALIBRATION RESULT

PARAMETER	DUC VALUE ( $\mu$ l)	STANDARD VALUE ( $\mu$ l)	ERROR ( $\mu$ l)	Uncertainty $\pm$ ( $\mu$ l)
Volume	200	200.11	-0.11	4.0
	500	500.23	-0.23	4.0

### 1. Electrical Safety

#### Visual Test

Sr. No	TESTS	REMARKS
1	Power Chords, Cable Check	OK
2	Main Socket Check	OK
3	Equipment Type (B/BF/CF)	B
4	Equipment Class (I/II/III)	I

B-BODY TYPE, BF-BODY FLOAT TYPE, CF-CARIDAC FLOAT TYPE, I-PROPERLY EARTHED, II-DOUBLE INSULATED  
III-WITH EXTRA SAFETY LOW VOLTAGE.

### 2. Electrical Safety Test

Sr. No	PARAMETERS	Measured Value	REMARKS
1	Voltage Between Line & Neutral (V in)	236.4 V	OK
2	Voltage Between Line & Earth (V ie)	243.6 V	OK
3	Voltage between Neutral & Earth (V en)	2.2 V	OK
4	Load Current (I a)	0.5 A	OK
5	Earth Leakage Current (I L)	7.8 $\mu$ A	OK
6	Enclosure Current	0.6 $\mu$ A	OK
7	Patient Leakage Current (P L)	0.7 $\mu$ A	OK

Expanded Uncertainty of measurement at approximately 95% Confidence Level with coverage factor  $k=2$

#### Note :-

- DUC = Device Under Calibration
- The Certificate refers only to the particular items submitted for calibration.
- This Certificate shall not be reproduced, except in full, without permission of CEO Hytech Instrument, Ghaziabad.
- Result Reported are Valid at the time of and under the stated conditions of measurement.
- Calibration Certificate issue for weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scale/Tapes etc. for scientific purpose only and should not be used for Trade / Commerical use.

**CALIBRATED BY**  
(Calib.,Engineer)

**APPROVED BY**  
(QM / TM)

