



HYTECH INSTRUMENT

25F, C-2 Shriram Complex, Nyay Khand-I, Indrapuram, Ghaziabad-201014 (U.P) INDIA
Ph. No. +91-9899380410, +91-9716668085, +91-7669021291-95



AII ISO 9001-2015
Certified Company

Calibration Certificate

Format No. : IY/FM/33

Page No. : 01 of 01

Certificate No.	HY/21/1121-9	CALIBRATION CERTIFICATE OF Refrigerator (Temperature Indicator)
Date of Calibration :	10.08.2021	
Next Date for Calibration	10.08.2022	
<i>(As per agreed with the Customer)</i>		
Certificate Issue Date	11.08.2021	

CUSTOMER ADDRESS		DESCRIPTION OF DEVICE UNDER CALIBRATION	
M/s Curis Diagnostics F-24/25, Sector-3, Rohini, New Delhi -110085 INDIA		Make / Model	BPL
SERVICE REQUEST FORM DETAILS		Instrument Sr. No.	NA
		Resolution / Least Count	0.1 °C
Service Request / Job No.	1121/1121-9	Range / Size	0 to 10 °C
Service Request Dated	09.08.2021	Asset Code / I.D No.	---
		Location	LAB
		Instrument Condition	In Working

Calibration Procedure	HY/WI/2T02	Calibration Performed At	At Site	
Reference Standard	NABL-126	Discipline	Thermal	
ENVIRONMENTAL CONDITIONS	Temperature	(25 ± 15) °C	Relative Humidity	(60 ± 15) %RH

DETAILS OF REFERENCE STANDARDS AND MAJOR EQUIPMENTS USED FOR CALIBRATION

Sr. No	Instrument Details	Certificate No.	Calibrated By	Calibration Due Date
1	Digital Multimeter with PRT Sensor	CC282120000001496F	TMPL	08.10.2021

The Standards used for calibration are traceable to National Standards.

CALIBRATION RESULT

Temperature in (°C)				
PARAMETER	DUC VALUE	STANDARD VALUE	ERROR	Uncertainty ± (%)
	2.0	2.001	-0.001	0.95
	4.0	4.002	-0.002	0.95
	5.0	5.004	-0.004	0.95
	6.0	6.007	-0.007	0.95
	8.0	8.009	-0.009	0.95

Expanded Uncertainty of measurement at approximately 95% Confidence Level with coverage factor $k=2$

- Note :-**
- DUC = Device Under Calibration
 - The Certificate refers only to the particular items submitted for Calibration.
 - This Certificate shall not be reproduced, except in full, without the permission of CEO Hytech Instrument, Ghaziabad.
 - Result Reported are Valid at the time of and under the stated conditions of measurement.
 - Calibration Certificate issue for weight & Measure parameters is not to be used for Trade / Commercial use.

CALIBRATED BY
(Calib. Engineer)

APPROVED BY
(QM / TM)

