

Ortho Clinical Diagnostics

403, Leela Business Park,
Andheri Kurla Road, Andheri
East, Mumbai - 400059
T : +91 22 67879300
F : +91 22 67879333

CALIBRATION CERTIFICATE

CERTIFICATE NO : OCD/PRIMA/02
CALIBRATED FOR : PRIMA DIAGNOSTICS
LOCATION : BANGLORE

CALIBRATION DATE : 13/04/2021
CALIBRATION DUE : 12/10/2021
CALIBRATED AT : AT SITE

SPECIFICATION OF INSTRUMENT :

INSTRUMENT	MODEL	SERIAL NO
VITROS CHEMISTRY SYSTEM	VITROS 250	27001264
	J NUMBER	J20705

Specification of Software:

SOFTWARE NAME	SOFTWARE VERSION	UPDATED ON
DOS Rmx os	V9.6	March 2019

The Reference of Calibration: The performance/Adjustments of various subsystems has been calibrated/tested by trained site engineer with pre-designed /calibrated tools provided for the particular subsystems by the company.

PM Procedure Checklists:

Subsystem Name

Adjusted/Verified

1. Sample Supply
* Hook Drive Belt
2. Slide Supply 1, 2 & slide transport
Mechanical Adjustment:
 - * Dispense Blade Centering Adjustment
 - * Dispense Blade Drive Belt Tension
 - * Plunger Centering AlignmentSoftware Adjustment:
 - * Dispense To Metering Position
3. Sample Metering
Mechanical Adjustment:
 - * Truck Sensor
 - * Truck Guide Button
 - * Metering Truck

Ortho Clinical Diagnostics

403, Leela Business Park,
Andheri Kurla Road, Andheri
East, Mumbai – 400059
T : +91 22 67879300
F : +91 22 67879333

Software Adjustment:

- * Tip Pickup X,Y and Height
- * Sample Tray at Sample Position –Auto Sample Tray
- * Proboscis Centering and Sample Cup Depth
- * Mixing Cup X,Y and its Depth
- * Diluents Tray
- * Rate / Colorimetric Metering Position
- * Potentiometric Metering Position
- * Proboscis Compression
- * Wetness Detector



6. Slide Incubator

Software Adjustment:

- * Incubator Disk Sizing
- * Incubator Disk Stopping
- * PM Slide Insertion Depth
- * Slide Read Position (RT/CM Depth & Read Sync)
- * Potentiometric Eject Blade Depth



7. Reflectometer

- * IRIS Lamp Output Voltage
- * Correction Factor




9. Master Computer

- * Touch Screen Calibration
- * System Full Backup



CALIBRATED BY


Yogeesh K
SR.Service engineer
Orthoclinical diagnostics
Banglore

NOTE:

1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in the certificate is valid at the time of and under the stated condition of the measurement.