

Date: 19/07/2021

CALIBRATION CERTIFICATE

This is to state that the hematology cell counter model **YUMIZEN H500** bearing Serial no: **710YOXH01128** at **Prima Diagnostic, Whitefield, Karnataka**, was calibrated with Calibrator on **19/07/2021**.

The Calibration Coefficient is automatically generated after calibration. Subsequently controls were processed and found them in respective range.

Calibrator Used: **ABX MINOCAL**

Lot No.: **CX 460**

Expiry Date: **05/08/2021**

Sr. No	Tested Parameter	Remarks
1	Startup/Blank Cycle	Passed
2	Repeatability/ Precision	Passed
3	Calibration	Passed
4	Quality Control	Passed

Next calibration cycle is due on **18/07/2022**.

Yeshwanth Padashetty



ENGINEER CUSTOMER SUPPORT
Horiba Medical.

