

Date: 20/07/2021

## CALIBRATION CERTIFICATE

This is to state that the hematology cell counter model **YUMIZEN H500** bearing Serial no: **803YOXH01348** at **Prima Diagnostic, Yelahanka, Bangalore, Karnataka**, was calibrated with Calibrator on **20/07/2021**.

The Calibration Coefficient is automatically generated after calibration. Subsequently controls were processed and found them in respective range.

Calibrator Used: **ABX MINOCAL**

Lot No.: **CX 460**

Expiry Date: **05/08/2021**

Sr. No	Tested Parameter	Remarks
1	Startup/Blank Cycle	Passed
2	Repeatability/ Precision	Passed
3	Calibration	Passed
4	Quality Control	Passed

Next calibration cycle is due on **19/07/2022**.

Yeshwanth Padashetty



ENGINEER CUSTOMER SUPPORT  
**Horiba Medical.**

