

Ortho Clinical Diagnostics

403, Leela Business Park, Andheri Kurla Road,
Andheri East, Mumbai – 400059
T : +91 22 6787 9300
F : +91 22 6787 9333

CALIBRATION CERTIFICATE

CERTIFICATE NO : OCD/PRIMA /002
CALIBRATED FOR : PRIMA DIAGNOSTICS
LOCATION : BANGALORE

CALIBRATION DATE : 07/05/2021
CALIBRATION DUE. : 06/11/2021
CALIBRATED AT : AT SITE

SPECIFICATION OF INSTRUMENT :

INSTRUMENT	MODEL	SERIAL NO
VITROS CHEMISTRY SYSTEM	VITROS 250	27001765
	J NUMBER	J27001765

Specification of Software:

SOFTWARE NAME	SOFTWARE VERSION	UPDATED ON
DOS Rmx os	V9.7	Nov 2019

The Reference of Calibration: The performance/Adjustments of various subsystems has been calibrated/tested by trained site engineer with pre-designed /calibrated tools provided for the particular subsystems by the company.

PM Procedure Checklists:

Subsystem Name

Adjusted/Verified

1. Sample Supply
* Hook Drive Belt




2. Slide Supply 1, 2 & slide transport
Mechanical Adjustment:

- * Dispense Blade Centering Adjustment
- * Dispense Blade Drive Belt Tension
- * Plunger Centering Alignment

Software Adjustment:

- * Dispense To Metering Position




3. Sample Metering

Mechanical Adjustment:

- * Truck Sensor
- * Truck Guide Button
- * Metering Truck

Software Adjustment:



Ortho Clinical Diagnostics India Private Limited

Registered Office: 403, Leela Business Park, Andheri Kurla Road, Andheri East, Mumbai - 400059 | T: +91 22 6787 9300 | F: +91 22 6787 9333 | CIN:
U51397MH2015FTC262650

Ortho Clinical Diagnostics


403, Leela Business Park, Andheri Kurla Road,
Andheri East, Mumbai – 400059
T : +91 22 6787 9300
F : +91 22 6787 9333

- * Tip Pickup X,Y and Height
- * Sample Tray at Sample Position –Auto Sample Tray
- * Proboscis Centering and Sample Cup Depth
- * Mixing Cup X,Y and its Depth
- * Diluents Tray
- * Rate / Colorimetric Metering Position
- * Potentiometric Metering Position
- * Proboscis Compression
- * Wetness Detector

6. Slide Incubator

Software Adjustment:

- * Incubator Disk Sizing
- * Incubator Disk Stopping
- * PM Slide Insertion Depth
- * Slide Read Position (RT/CM Depth & Read Sync)
- * Potentiometric Eject Blade Depth




7. Reflectometer

- * IRIS Lamp Output Voltage
- * Correction Factor



9. Master Computer

- * Touch Screen Calibration
- * System Full Backup



CALIBRATED BY



Mr.Sayandh Parayil(Service Engineer)

NOTE:

1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in the certificate is valid at the time of and under the stated condition of the measurement.

Ortho Clinical Diagnostics India Private Limited

Registered Office: 403, Leela Business Park, Andheri Kurla Road, Andheri East, Mumbai - 400059 | T: +91 22 6787 9300 | F: +91 22 6787 9333 | CIN: US1397MH2015FTC262650