

2SF, C-2 Shriram Complex, Nyay Khand-I, Indirapuram, Ghaziabad-201014 (U.P) INDIA Ph. No. +91-9899380410, +91-9716668085, +91-7669021291-95

Calibration Cartificate

Page No.: 01 of 01

Date of Calibration:	HY/21/1132-1 23.08.2021 23.08.2022	CALIBRATION CERTIFICATE OF
Next Date for Calibration (As per agreed with the Customer) Certificate Issue Date	24.08.2021	Centrifuge Machine

Certificate Issue Date	24.00.2021		INDER CALIBRATION
		DESCRIPTION OF DEVICE U	JNDER CREIBIATTION
CUSTOMER ADDRESS		Make / Model	Remi / 8-C
M/s Curis Diagnostics		Instrument Sr. No.	Remi-1F-032
F-24/25, Sector-3, Rohini,		Resolution / Least Count	As Per Range
New Delhi -110085 INDIA		Range / Size	NA
		Asset Code / I.D No.	CENT-01
SERVICE REQUEST FORM	DETAILS	Location	LAB
Service Request / Job No.	1132/1132-1	Instrument Condition	In Working
Service Request Dated	22.08.2021		At Site
	HY/WI/2M14	Calibration Performed At	
Calibration Procedure	H1/W1/2011	Dissipling	Medical

Service Request Dated	22.08.2021			At Site
	HY/WI/2M14	Calibration Per	lormed At	Medical
Calibration Procedure	NABL-126	Discipline		153.04011
Reference Standard		(25 ± 15) °C	Relative Humidity	(60 ± 13) 70KH
ENVIRONMENTAL CONDITIONS	remperature		THE HEED FOR CALL	IRRATION

DETAILS OF REFERENCE STANDARDS AND MAJOR EQUIPMENTS USED FOR CALIBRATION

	DETAILS OF REFERENCE ST.	ANDARDS AND MAJOR EQUIPME	N 15 USED FOR CA	Calibration Due Date
		Certificate No.	Calibrated By	Calibration Due Date
Sr. N	o Instrument Details		TNC	18.01.2022
1	Digital Tachometer	CC20552000000753F		

The Standards used for calibration are traceable to National Standards.

CALIBRATION RESULT

- AMERICA	DUC VALUE	STANDARD VALUE (rpm)	ERROR (rpm)	Uncertainty ± (%)
PARAMETER	(rpm)		-1.30	0.70
Speed	800	801.3	-1.30	0
Speed		1004	-4.00	0.70
	1000	100.	25/5	0.70
	1500	1505	-5.00	0.70
		2522	-3.00	0.70
	2500	2503	2503	
	2000	3006	-6.00	0.70
	3000			

Janue	d Uncertainty of measurment at approximately 95% Confidence Level with coverage factor $k=2$	CALIBRATED B
e :-	1. DUC = Device Under Calibration	
	2. The Certificate refers only to the particular items submitted for The St. Co. Co. Hytech Instrument, Ghaziabad.	(Calib. Engineer
	a The Cartificate shall not be reproduced, except in full, with the cartificate shall not be reproduced, except in full, with the cartificate shall not be reproduced, except in full, with the cartificate shall not be reproduced, except in full, with the cartificate shall not be reproduced.	
	to the time of and under the supply	APPROVED BY
	4. Result Reported are Valid at the time of an experiment of the Schlafter Volumetric equipment, 5. Calibration Certificate issue for weight & Measure parameter the Schlafter Volumetric equipment, Measuring Scale/Tapes etc. for scientific purpose only and Scale/Tapes etc. for scientific purpose only and Scale/Tapes etc.	(QM/TM)





