

2SF, C-2 Shriram Complex, Nyay Khand-I, Indirapuram, Ghaziabad-201014 (U.P) INDIA Ph. No. +91-9899380410, +91-9716668085, +91-7669021291-95

Calibration Certificate

Page No.: 01 of 01

Date of Calibration :	HY/21/1132-1 23.08.2021 23.08.2022	CALIBRATION CERTIFICATE OF
Next Date for Calibration (As per agreed with the Customer) Certificate Issue Date		Centrifuge Machine

Cartificate Issue Date	24.00.2021			
Certificate Issue Date 24.06.2021		DESCRIPTION OF DEVICE UNDER CALIBRATION		
CUSTOMER ADDRESS		Make / Model	Remi / 8-C	
M/s Curis Diagnostics F-24/25, Sector-3, Rohini,		Instrument Sr. No. Resolution / Least Count	Remi-1F-032	
			As Per Range	
New Delhi -110085 INDIA		Range / Size	NA	
		Asset Code / I.D No.	CENT-01	
SERVICE REQUEST FORM DETAILS		Location	LAB	
Service Request / Job No.	1132/1132-1	Instrument Condition	In Working	
Service Request Dated	22.08.2021	Instrument Condition		
Der in it		C. 13 - Non Performed At	At Site	

Service Request Dated	2.00,011	Calibration Performed A		At Site
Calibration Procedure	HY/WI/2M14			Medical
Reference Standard	NABL-126	Discipline Relativ	e Humidity	(60 ± 15) %RH
ENVIRONMENTAL CONDITIONS	Temperature	(25 ± 15) °C Relativ		

DETAILS OF REFERENCE STANDARDS AND MAJOR EQUIPMENTS USED FOR CALIBRATION

	DETAILS OF REFERENCE STA	ANDARDS AND MAJOR EQUIT M		Calibration Due Date
	Instrument Details	Certificate No.	Calibrated By	
Sr. No		CC20552000000753F	TNC	18.01.2022
1	Digital Tachometer		National Standards.	

The Standards used for calibration are traceable to National Standards.

CALIBRATION RESULT

		STANDARD VALUE	ERROR	Uncertainty
PARAMETER	DUC VALUE (rpm)	(rpm)	(rpm)	± (%)
Speed	800	801.3	-1.30	0.70
	1000	1004	-4.00	0.70
	1500	1505	-5.00	0.70
	2500	2503	-3.00	0.70
	3000	3006	-6.00	0.70

ote :-	ed Uncertainty of measurment at approximately 95% Confidence Level with coverage factor k = 2 1. DUC = Device Under Calibration 2. The Certificate refers only to the particular items submitted for the Section of Hytech Instrument, Ghaziabad.	(Calib. Engineer
	3. This Certificate shall not be reproduced, except in full, without transport of SEO Hytech Instrument, Chaziabad. 4. Result Reported are Valid at the time of and under the states to the Company of the prement. 5. Calibration Certificate issue for weight & Measure parameter to the Company of the prement of of the Premen	APPROVED BY



