



CALIBRATION CERTIFICATE

Customer Details

Customer Name : **Pace Hospitals**

Address: Number C1775, Metro Pillar, 18, Hitech City Rd, Hyderabad, Telangana 500081

Contact Name : Mr. Quddus

Contact No.: 096527 27383

Instrument Details

Instrument Type: miniVIDAS

Serial No.: **IVD5209998**

S/w Version: 5.6.1

Calibration Details

Calibration Date: 21-07-2021

Calibration expired on: 20-07-2022

Calibration Report:

1. All the alignments are in the specified range	Yes
2. All the voltages are in the specified range	Yes
3. The Tray Temperatures are in the specified range (37.0°C +/- 1.7°C)	Yes
4. The SPR Temperatures are in the specified range (37.0°C +/- 1.5°C)	Yes
5. Calibration of optics (Fluorescence) value is within range with calibrator strip value (3168+/-50) RFUs	Yes

Used Equipments : Fluke Multimeter (SI No : 12790758)

Authorised Signatory

Engineer's Name: Ravi POPURI

Signature: