



CALIBRATION CERTIFICATE

1.CUSTOMER :- G1 PATHOLOGY LAB SERVICES Sr No. 10/4,VadgaonKhurd, ,Singhad Road,Lokamt Office, 41, Nanded City Rd, Pune,411041	Page No. :- 1 of 1 Certificate No. :- GTS/210629/01-004 Date of Received :- 29.06.2021 Date of Calibration :- 29.06.2021 Next Calibration Due On :- 28.06.2022 Issue Date :- 30.06.2021 Calibration method No. :- TH-WI-03 ULR No :- CC295721000004995F
Ambient Temp. (°C) :- 25.6 Relative Humidity (%RH) :- 51 Location of calibration :- In Lab Condition of Item :- Ok	

2. Description of Item			
Name :- Glass Thermometer	Range :- -10 To 110 °C	Id No :- G1PLS/ACC/GT-01	Least Count :- 1 °C
Make :- Zeal	Location :- Accession	Type :- Glass	

3.Details of Equipment used for calibration				
Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2106/002/004	Nishitronics Instrumentation	GTS/PTH/01	31.05.2022

*Thermal Calibration				
4.Calibration Results :-				
Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
-10	-10.18	-10	0.18	0.80
50	50.34	50	-0.34	0.80
110	110.42	110	-0.42	0.80

- Note:**
- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95.45% for normal distribution
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
 - 6) Type of immersion : Partial

Calibrated By
D.B
Calibration Engineer
D Ballal



Approved By
SB
Technical Manager
Swapnil Bhagwat



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER :- G1 PATHOLOGY LAB SERVICES Sr No. 10/4,VadgaonKhurd, ,Singhad Road,Lokamt Office, 41, Nanded City Rd, Pune,411041	Page No. :- 1 of 1 Certificate No. :- GTS/210629/01-005 Date of Received :- 29.06.2021 Date of Calibration :- 29.06.2021 Next Calibration Due On :- 28.06.2022 Issue Date :- 30.06.2021 Calibration method No. :- TH-WI-03 ULR No :- CC295721000004996F
Ambient Temp. (°C) :- 27.0 Relative Humidity (%RH) :- 51 Location of calibration :- In Lab Condition of Item :- Ok	

Description of Item

Name :- Thermometer	Range :- -50 To 85 °C
Id No :- G1PLS/ TH-01	Least Count :- 0.1 °C
Make :- -	Location :- Lab
Type :- Digital	

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2106/002/004	Nishitronics Instrumentation	GTS/PTH/01	31.05.2022

*Thermal Calibration

4.Calibration Results :-

Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	2.10	2	-0.10	0.80
5	5.24	5	-0.24	0.80
8	8.31	8	-0.31	0.80

- Note:**
- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
 - 6) Type of immersion : Partial

Calibrated By

D.B
Calibration Engineer
D Ballal



Approved By

Sw
Technical Manager
Swapnil Bhagwat

RF-55/01

End of Certificate



CALIBRATION CERTIFICATE

1.CUSTOMER :- G1 PATHOLOGY LAB SERVICES Sr No. 10/4,VadgaonKhurd, ,Singhad Road,Lokamt Office, 41, Nanded City Rd, Pune,411041	Page No. :- 1 of 1 Certificate No. :- GTS/210629/01-006 Date of Received :- 29.06.2021 Date of Calibration :- 29.06.2021 Next Calibration Due On :- 28.06.2022 Issue Date :- 30.06.2021 Calibration method No. :- TH-WI-03 ULR No :- CC295721000004997F
Ambient Temp. (°C) :- 25.5 Relative Humidity (%RH) :- 49 Location of calibration :- In Lab Condition of Item :- Ok	

Description of Item			
Name :- Thermometer	Range :- -50 To 85 °C	Id No :- G1PLS/ TH-02	Least Count :- 0.1 °C
Make :- -	Location :- Lab	Type :- Digital	

3.Details of Equipment used for calibration				
Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2106/002/004	Nishitronics Instrumentation	GTS/PTH/01	31.05.2022

*Thermal Calibration				
4.Calibration Results :-				
Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	2.22	2	-0.22	0.80
5	5.29	5	-0.29	0.80
8	8.45	8	-0.45	0.80

- Note:**
- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
 - 6) Type of immersion : Partial

Calibrated By Calibration Engineer D Ballal		Approved By Technical Manager Swapnil Bhagwat
RF-55/01	End of Certificate	



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER :- G1 PATHOLOGY LAB SERVICES Sr No. 10/4,VadgaonKhurd, ,Singhad Road,Lokamt Office, 41, Nanded City Rd, Pune,411041	Page No. :- 1 of 1 Certificate No. :- GTS/210629/01-007 Date of Received :- 29.06.2021 Date of Calibration :- 29.06.2021 Next Calibration Due On :- 28.06.2022 Issue Date :- 30.06.2021 Calibration method No. :- TH-WI-03 ULR No :- CC295721000004998F
Ambient Temp. (°C) :- 26.8 Relative Humidity (%RH) :- 50 Location of calibration :- In Lab Condition of Item :- Ok	

Description of Item

Name :- Thermometer	Range :- -50 To 85 °C
Id No :- G1PLS/ TH-03	Least Count :- 0.1 °C
Make :- -	Location :- Lab
Type :- Digital	

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2106/002/004	Nishitronics Instrumentation	GTS/PTH/01	31.05.2022

*Thermal Calibration

4.Calibration Results :-

Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	2.25	2	-0.25	0.80
5	5.40	5	-0.40	0.80
8	8.52	8	-0.52	0.80

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
- 6) Type of immersion : Partial

Calibrated By

D.B

Calibration Engineer

D Ballal

RF-55/01



Approved By

Sw

Technical Manager

Swapnil Bhagwat

End of Certificate



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728

CALIBRATION CERTIFICATE

1.CUSTOMER :-
G1 PATHOLOGY LAB SERVICES
Sr No. 10/4,VadgaonKhurd,
,Singhad Road,Lokamt Office,
41, Nanded City Rd, Pune,411041

Ambient Temp. (°C) :- 25.6
Relative Humidity (%RH) :- 49
Location of calibration :- In Lab
Condition of Item :- Ok

Page No. :- 1 of 1
Certificate No. :- GTS/210629/01-008
Date of Received :- 29.06.2021
Date of Calibration :- 29.06.2021
Next Calibration Due On :- 28.06.2022
Issue Date :- 30.06.2021
Calibration method No. :- TH-WI-02

2. Description of Item

Name :- Thermohygrometer Range :- -50° to70° C 10% to99%RH
Id No :- G1PLS/ HYG-01 Least Count :- 0.1 ° C / 1% RH
Make :- HTC Location :- Lab
Type :- Digital Sr No :- --

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Thermo-Hygrometer	GTS/CAL-2105/26-001	Technocal	GTS/HYG-01	25.05.2022

*Thermal Calibration

4.Calibration Results :-

Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Unc. in ± °C
20	20.02	20.2	0.18	0.80
30	30.03	30.3	0.27	0.80
50	50.10	50.5	0.40	0.80
%RH	%RH	%RH	%RH	in ± %RH
30	30.15	30	-0.15	2.80
50	50.24	51	0.76	2.80
80	80.33	81	0.67	2.80

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By

D.B
Calibration Engineer
D Ballal



Approved By

Sw
Technical Manager
Swapnil Bhagwat

RF-55/01

End of Certificate



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728

CALIBRATION CERTIFICATE

1.CUSTOMER :- G1 PATHOLOGY LAB SERVICES Sr No. 10/4,VadgaonKhurd, ,Singhad Road,Lokamt Office, 41, Nanded City Rd, Pune,411041	Page No. :- 1 of 1 Certificate No. :- GTS/210629/01-009 Date of Received :- 29.06.2021 Date of Calibration :- 29.06.2021 Next Calibration Due On :- 28.06.2022 Issue Date :- 30.06.2021 Calibration method No. :- TH-WI-02
Ambient Temp. (°C) :- 25.2 Relative Humidity (%RH) :- 51 Location of calibration :- In Lab Condition of Item :- Ok	

2. Description of Item	
Name :- Thermohygrometer I No :- G1PLS/ HYG-02 Make :- HTC Type :- Digital	Range :- -50° to70° 10% to99%RH Least Count :- 0.1° C / 1% RH Location :- Lab Sr No :- --

3.Details of Equipment used for calibration				
Name	Certificate No.	Certified By	ID No.	Calibration Validity
Thermo-Hygrometer	GTS/CAL-2105/26-001	Technocal	GTS/HYG-01	25.05.2022

*Thermal Calibration				
4.Calibration Results :-				
Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Unc. in ± °C
20	20.04	20.2	0.16	0.80
30	30.04	30.4	0.36	0.80
50	50.12	50.5	0.38	0.80
%RH	%RH	%RH	%RH	in ± %RH
30	30.15	30	-0.15	2.80
50	50.24	52	1.76	2.80
80	80.42	82	1.58	2.80

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By

D.B

Calibration Engineer
D Ballal



Approved By

SMB

Technical Manager
Swapnil Bhagwat

RF-55/01

End of Certificate