

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Bilirubin Direct	Erba Chem 5/Plus	0.19	mg/dl	2.59
! Calcium	Erba Chem 5/Plus	6.25	mg/dl	-2.32
! Uric Acid	Erba Chem 5/Plus	4.09	mg/dl	-2.01
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**SADBHAVANA PATHOLOGY LAB (446)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 210207 and Sample No 01**


**Report Date : 19/07/2021**

**Instrument : Erba Chem 5/Plus**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Bilirubin Total	mg/dl	0.28	--	1.19	0.74
! Bilirubin Direct	mg/dl	0.19	--	2.59	2.48
! Calcium	mg/dl	6.25	8.50	-2.32	-1.88
! Uric Acid	mg/dl	4.09	--	-2.01	-1.25
@ Triglycerides	mg/dl	117.90	--	-1.91	-2.46
@ ALT	U/L	20.33	--	-0.08	0.79

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	6
Not Evaluated Parameters	0
Evaluated Parameters	6
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416.  
Commercial Building - 1A,  
Kohinoor Mall, Kiroli Road, Kurla (W),  
Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Bilirubin Total	Erba Chem 5/Plus	0.34	mg/dl	2.16
X Bilirubin Direct	Erba Chem 5/Plus	0.23	mg/dl	3.51
! Calcium	Erba Chem 5/Plus	6.12	mg/dl	-2.86
! Triglycerides	Erba Chem 5/Plus	111.10	mg/dl	-2.18

Legend @ : z score ≤ 2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score ≥ 3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**SADBHAVANA PATHOLOGY LAB (446)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 210207 and Sample No 02**


**Report Date : 19/07/2021**

**Instrument : Erba Chem 5/Plus**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
! Bilirubin Total	mg/dl	0.34	--	2.16	0.74
X Bilirubin Direct	mg/dl	0.23	--	3.51	2.48
! Calcium	mg/dl	6.12	8.70	-2.86	-1.88
@ Uric Acid	mg/dl	3.83	--	-1.73	-1.25
! Triglycerides	mg/dl	111.10	--	-2.18	-2.46
@ ALT	U/L	16.80	--	-0.92	0.79

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	6
Not Evaluated Parameters	0
Evaluated Parameters	6
Outlier Parameters (X)	1
EQAS Score Biochemistry	83.33 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416.  
Commercial Building - 1A,  
Kohinoor Mall, Kiroli Road, Kurla (W),  
Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Glucose	Erba Chem 5/Plus	52.03	mg/dl	-2.32
X Urea	Erba Chem 5/Plus	32.24	mg/dl	3.37
! Creatinine	Erba Chem 5/Plus	1.33	mg/dl	2.49
! Cholesterol	Erba Chem 5/Plus	125.20	mg/dl	-2.90
Legend @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**SADBHAVANA PATHOLOGY LAB (446)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 210207 and Sample No 03**


**Report Date : 19/07/2021**

**Instrument : Erba Chem 5/Plus**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
! Glucose	mg/dl	52.03	--	-2.32	-0.68
X Urea	mg/dl	32.24	--	3.37	2.87
! Creatinine	mg/dl	1.33	--	2.49	1.43
! Cholesterol	mg/dl	125.20	--	-2.90	-1.81

Legend @ : z score ≤ 2.0 - Acceptable  
 ! : 2.0 < z score < 3.0 - Warning  
 X : z score ≥ 3 - Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	4
Not Evaluated Parameters	0
Evaluated Parameters	4
Outlier Parameters (X)	1
EQAS Score Biochemistry	75.00 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416,  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Urea	Erba Chem 5/Plus	29.61	mg/dl	3.89
! Cholesterol	Erba Chem 5/Plus	122.30	mg/dl	-2.41
! Glucose	Erba Chem 5/Plus	57.73	mg/dl	-2.21
Legend @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**SADBHAVANA PATHOLOGY LAB (446)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 210207 and Sample No 04**


**Report Date : 19/07/2021**

**Instrument : Erba Chem 5/Plus**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
X Urea	mg/dl	29.61	--	3.89	2.87
! Cholesterol	mg/dl	122.30	--	-2.41	-1.81
! Glucose	mg/dl	57.73	--	-2.21	-0.68
@ Creatinine	mg/dl	1.15	--	1.56	1.43

Legend @ : z score ≤ 2.0 - Acceptable  
 ! : 2.0 < z score < 3.0 - Warning  
 X : z score ≥ 3 - Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	4
Not Evaluated Parameters	0
Evaluated Parameters	4
Outlier Parameters (X)	1
EQAS Score Biochemistry	75.00 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416,  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070

End of Report