

LAB MONTHLY SUMMARY



Lab Name **SANJEEVAN MEDICAL RESEARCH CENTRE Pvt Ltd** Lab No **5284**
 Month **July** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 14/07/2021

Date of Report Published : 02/08/2021

Sl. No	Parameter	Method	Any Analyser (Automation / Semi Automation)	900	147.82	9.98	14.75	128.3 mg/dl	-1.32	0.94
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	900	147.82	9.98	14.75	128.3 mg/dl	-1.32	0.94
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	567	32.30	14.53	4.69	26.1 mg/dl	-1.32	0.39
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	601	2.70	13.88	0.38	2.8 mg/dl	0.27	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	775	2.96	15.26	0.45	2.8 mg/dl	-0.35	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	742	5.53	9.90	0.55	5.1 g/dl	-0.78	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	711	3.31	9.70	0.32	3.7 g/dl	1.22	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	834	5.17	19.21	0.99	3.7 mg/dl	-1.48	0.07
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	984	115.76	13.00	15.04	95.8 mg/dl	-1.33	0.96
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	818	177.61	12.70	22.56	156.4 mg/dl	-0.94	1.58

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

[Handwritten Signature]

[Handwritten Signature]

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****