

## Calibration Certificate

Following voltages have been checked & adjusted on Blood Cell Counter SWELAB ALFA (SI.No. 23115) at **Sanjeevan Hospital (A Unit of Sanjeevan Medical Research Centre Pvt Ltd.), 4869/24, Ansari Road, Darya Ganj, New Delhi-110002.**

Power Supply: Voltages 24 V are OK.

Checked the Gain and Offset of Analog Board and found OK.

All instrument thru' the Service Program and found OK.

Checked the Pumps & Valves and found OK.

Parameters background check is also under acceptable range.

Calibrated the instrument using Boule Calibrator Lot No-22101-34. (Exp 2021-04-27)

### Calibration Report

Parameter	Target	Range	Actual Result
RBC( $10^{10}/L$ )	4.22	0.10	4.16
MCV(fL)	82.0	2.0	81.5
Plt( $10^9/L$ )	231	15.0	240
MPV(fL)	10.6	1.5	10.8
WBC( $10^9/L$ )	8.3	0.3	8.3
Hgb(g/dl)	12.2	0.2	12.1

Next Calibration will be performed on \_\_\_\_ 13/04/2022 \_\_\_\_

**Best Regards**  
**For, AAMX BIOMEDICALS**

**Date: 14/04/2021**



**Man Singh**  
Head-Customer Support  
Cell+91 9310321234