CERTIFICATE OF CALIBRATION

Ref. No.: Applicant : MSM005 MEDICITY SPECIALITY MEDICAL LABORATORY

Device Calibrated : Make/Model: Sl.No: FRIDGE THERMOMETER ../..

Condition at receipt : Functional

Calibrated at : Onsite Date of Calibration : 30/08/2021 Calibration due date : 29/08/2022

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue: 31/08/2021

Approved Signatory

no then

Cyrix Healthcare Pvt. Ltd., 30/641 B Petta Junction, Poonithura, Kochi-38. Ph: +91 484 6544033, Mobile: 97455 55073 Email: sales@cyrixhealthcare.com service@cyrixhealthcare.com



www.cyrixhealthcare.com



AERB Approved QA Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MEDICITY SPECIALITY MEDICAL LABORATORY THALIPARAMBA, KERALA 670141	Certificate No.	MSM005
		Date of Calibration	30/08/2021
		Valid Upto	29/08/2022
		Date of Issue	31/08/2021

Instruments Details			
Instrument Name FRIDGE THERMOME		S No.	NA
Make/Model No.	NA	Location	LAB
RANGE	-50°C - +85°C	Visual Inspection	ОК

Details of Reference Standards & Major Equipments Used			
Equipment Name	Equipment Name THERMO HYGROMETER INFRARED THERMOMETER		
Make	HTC-2	FLUKE	
Model/SR No.			
Calibration Validity	22/09/2021	10/01/2022	
Calibration By	NENO TECHNICAL SERVICES	NENO TECHNICAL SERVICES	

Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	-10	-10	±2	PASS
2	-5	-5	±2	PASS
3	5	5	±2	PASS

REMARKS:-

* This result of calibration refers to only to the particular items submitted for calibration.

The above result is valid at the time of and under the stated conditions measurement.

Cali bration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

(SE)





Cyrix Healthcare Pvt. Ltd., 30/641 B Petta Junction, Poonithura, Kochi-38. Ph: +91 484 6544033, Mobile: 97455 55073 Email: sales@cyrixhealthcare.com service@cyrixhealthcare.com



www.cyrixhealthcare.com





