

CERTIFICATE OF CALIBRATION

Ref. No.: MSM004
Applicant : MEDICITY SPECIALITY MEDICAL LABORATORY
Device Calibrated : FRIDGE THERMOMETER
Make/Model: ../..
Sl.No: ...

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 30/08/2021
Calibration due date : 29/08/2022

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 31/08/2021

Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: +91 484 6544033, Mobile: 97455 55073
Email: sales@cyrixhealthcare.com
service@cyrixhealthcare.com

CYRIX
HEALTH CARE PVT LTD

www.cyrixhealthcare.com



ISO 9001:2008 ISO 13485:2012
Certified Company

AERB Approved
QA Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MEDICITY SPECIALITY MEDICAL LABORATORY THALIPARAMBA, KERALA 670141	Certificate No.	MSM004
		Date of Calibration	30/08/2021
		Valid Upto	29/08/2022
		Date of Issue	31/08/2021

Instruments Details			
Instrument Name	FRIDGE THERMOMETER	S No.	NA
Make/Model No.	NA	Location	LAB
RANGE	-50°C - +85°C	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	THERMO HYGROMETER	INFRARED THERMOMETER	
Make	HTC-2	FLUKE	
Model/SR No.	
Calibration Validity	22/09/2021	10/01/2022	
Calibration By	NENO TECHNICAL SERVICES	NENO TECHNICAL SERVICES	


Environmental Condition			
Temperature	28°C	Relative Humidity	58%

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	-10	-10	±2	PASS
2	-5	-5	±2	PASS
3	5	5	±2	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-39.
Ph: +91 484 6544033, Mobile: 97455 55073
Email: sales@cyrixhealthcare.com
service@cyrixhealthcare.com

CYRIX
HEALTH CARE PVT LTD

www.cyrixhealthcare.com



SAMSUNG

DIGITAL INVERTER
10 YEAR WARRANTY

Energy Efficiency Label with star rating and text.

Smart System
Smart Cooling System
Smart Power Save Mode
Smart Fresh Guard
Smart Family Care

Temperature gauge on the freezer door.

Temperature gauge on the refrigerator door.

K.V
ology

F 0.5°C

ON/OFF °C/°F

