



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3067

Distribution No.: 152-H

Month/Year: March/2021

Instrument ID: 1/152H

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 17-05-2021[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	5.6	5.6	11.2	10.5	0.0660	0.73	0	0.1	0.0130	-0.84
RBC x10 <sup>6</sup> /µl	1	3.79	3.67	7.46	6.68	0.0120	5.13	0.12	0.03	0.0040	2.86
Hb g/dl	1	9.6	9.5	19.1	20.3	0.0350	-2.70	0.1	0.1	0.0120	0.00
HCT%	1	38.6	36.8	75.4	65.35	0.2100	3.25	1.8	0.3	0.0350	4.76
MCV-fl	1	100.7	100.2	200.9	196.35	0.4970	0.59	0.5	0.3	0.0420	0.45
MCH-Pg	1	26.2	25.1	51.3	60.8	0.1170	-6.18	1.1	0.2	0.0310	3.04
MCHC-g/dl	1	26.1	24.9	51	61.8	0.2170	-3.43	1.2	0.25	0.0350	3.02
Plt. x10 <sup>3</sup> /µl	1	176	175	351	292.5	2.10	1.79	1	5.5	0.61	-0.84
Retic %	2										

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs = , Poly= L=, E=, Mono/Promono= , B1= P.M.=, Mye=, Meta=, Other=	Poly: 30 - 55, Myelo: 10 - 40, Meta: 5 - 15, Promyelo/Eos/Blast: 1 - 10, nRBC/Baso/Lympho/Mono: 0 - 5		
RBC Morphology	3		Predominantly: Normocytic/Normochromic; Moderate: Microcytosis, Hypochromia, Anisocytosis; Mild: Macrocytosis, Poikilocytosis		
Diagnosis	3		Chronic Myeloid Leukemia (CML)		

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /µl	1	92	104	90.38	83.65	3.85	5.77	5.77	9.62
RBC x10 <sup>6</sup> /µl	1	92	104	82.69	90.38	9.62	6.73	7.69	2.88
Hb g/dl	1	92	104	78.85	94.23	9.62	3.85	11.54	1.92
HCT%	1	92	104	91.35	89.42	6.73	5.77	1.92	4.81
MCV-fl	1	92	104	93.27	92.31	3.85	1.92	2.88	5.77
MCH-Pg	1	92	104	81.73	91.35	10.58	4.81	7.69	3.85
MCHC-g/dl	1	92	104	91.35	92.31	3.85	3.85	4.81	3.85
Plt. x10 <sup>3</sup> /µl	1	92	104	94.23	94.23	4.81	0.96	0.96	4.81
ReticCount%	2	92	91	96.7	93.41	2.2	1.1	1.1	5.49
PS Assessment	3	92	95	Acceptable:94.5,Warning Signal:3.3,Unacceptable :2.2					

**Comments:**

1). Among Lab (EQA) : CBC result for RBC, HCT, MCH & MCHC unacceptable, please check calibration/human error, Retic Results & PS not reported, Remaining results acceptable.

2). Within Lab (IQA) : Difference in the CBC measurement values for HCT, MCH & MCHC unacceptable, please check precision/human error. Remaining precision acceptable.

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).


**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----