



CALIBRATION CERTIFICATE

1.CUSTOMER :- G1 PATHOLOGY LAB SERVICES Sr No. 10/4,VadgaonKhurd, ,Singhad Road,Lokamt Office, 41, Nanded City Rd, Pune,411041	Page No. :- 1 of 1 Certificate No. :- GTS/210629/01-004 Date of Received :- 29.06.2021 Date of Calibration :- 29.06.2021 Next Calibration Due On :- 28.06.2022 Issue Date :- 30.06.2021 Calibration method No. :- TH-WI-03 ULR No :- CC295721000004995F
Ambient Temp. (°C) :- 25.6 Relative Humidity (%RH) :- 51 Location of calibration :- In Lab Condition of Item :- Ok	

Description of Item

Name :- Glass Thermometer	Range :- -10 To 110 °C
Id No :- G1PLS/ACC/GT-01	Least Count :- 1 °C
Make :- Zeal	Location :- Accession
Type :- Glass	

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2106/002/004	Nishitronics Instrumentation	GTS/PTH/01	31.05.2022

*Thermal Calibration

4.Calibration Results :-

Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
-10	-10.18	-10	0.18	0.80
50	50.34	50	-0.34	0.80
110	110.42	110	-0.42	0.80

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
- 6) Type of immersion : Partial

Calibrated By

D.B
Calibration Engineer
D Ballal



Approved By

SwB
Technical Manager
Swapnil Bhagwat

RF-55/01

End of Certificate



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER :- G1 PATHOLOGY LAB SERVICES Sr No. 10/4,VadgaonKhurd, ,Singhad Road,Lokamt Office, 41, Nanded City Rd, Pune,411041	Page No. :- 1 of 1 Certificate No. :- GTS/210629/01-005 Date of Received :- 29.06.2021 Date of Calibration :- 29.06.2021 Next Calibration Due On :- 28.06.2022 Issue Date :- 30.06.2021 Calibration method No. :- TH-WI-03 ULR No :- CC295721000004996F
Ambient Temp. (°C) :- 27.0 Relative Humidity (%RH) :- 51 Location of calibration :- In Lab Condition of Item :- Ok	

Description of Item

Name :- Thermometer	Range :- -50 To 85 °C
Id No :- G1PLS/ TH-01	Least Count :- 0.1 °C
Make :- -	Location :- Lab
Type :- Digital	

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2106/002/004	Nishitronics Instrumentation	GTS/PTH/01	31.05.2022

*Thermal Calibration

4.Calibration Results :-

Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	2.10	2	-0.10	0.80
5	5.24	5	-0.24	0.80
8	8.31	8	-0.31	0.80

- Note:**
- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
 - 6) Type of immersion : Partial

Calibrated By

D.B
Calibration Engineer
D Ballal



Approved By

Sw
Technical Manager
Swapnil Bhagwat

RF-55/01

End of Certificate



CALIBRATION CERTIFICATE

1.CUSTOMER :- G1 PATHOLOGY LAB SERVICES Sr No. 10/4,VadgaonKhurd, ,Singhad Road,Lokamt Office, 41, Nanded City Rd, Pune,411041	Page No. :- 1 of 1 Certificate No. :- GTS/210629/01-006 Date of Received :- 29.06.2021 Date of Calibration :- 29.06.2021 Next Calibration Due On :- 28.06.2022 Issue Date :- 30.06.2021 Calibration method No. :- TH-WI-03 ULR No :- CC295721000004997F
Ambient Temp. (°C) :- 25.5 Relative Humidity (%RH) :- 49 Location of calibration :- In Lab Condition of Item :- Ok	

Description of Item

Name :- Thermometer	Range :- -50 To 85 °C
Id No :- G1PLS/ TH-02	Least Count :- 0.1 °C
Make :-	Location :- Lab
Type :- Digital	

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2106/002/004	Nishitronics Instrumentation	GTS/PTH/01	31.05.2022

*Thermal Calibration

4.Calibration Results :-

Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	2.22	2	-0.22	0.80
5	5.29	5	-0.29	0.80
8	8.45	8	-0.45	0.80

- Note:**
- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
 - 6) Type of immersion : Partial

Calibrated By

D.B

Calibration Engineer

D Ballal



Approved By

Sw

Technical Manager

Swapnil Bhagwat

RF-55/01

End of Certificate



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER :- G1 PATHOLOGY LAB SERVICES Sr No. 10/4,VadgaonKhurd, ,Singhad Road,Lokamt Office, 41, Nanded City Rd, Pune,411041	Page No. :- 1 of 1 Certificate No. :- GTS/210629/01-007 Date of Received :- 29.06.2021 Date of Calibration :- 29.06.2021 Next Calibration Due On :- 28.06.2022 Issue Date :- 30.06.2021 Calibration method No. :- TH-WI-03 ULR No :- CC295721000004998F
Ambient Temp. (°C) :- 26.8 Relative Humidity (%RH) :- 50 Location of calibration :- In Lab Condition of Item :- Ok	

Description of Item

Name :- Thermometer	Range :- -50 To 85 °C
Id No :- G1PLS/ TH-03	Least Count :- 0.1 °C
Make :- -	Location :- Lab
Type :- Digital	

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2106/002/004	Nishitronics Instrumentation	GTS/PTH/01	31.05.2022

*Thermal Calibration

4.Calibration Results :-

Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	2.25	2	-0.25	0.80
5	5.40	5	-0.40	0.80
8	8.52	8	-0.52	0.80

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
- 6) Type of immersion : Partial

Calibrated By

D.B
Calibration Engineer
D Ballal

RF-55/01



Approved By

SMB
Technical Manager
Swapnil Bhagwat

End of Certificate