

G1 PATHOLOGY LAB SERVICES (559)
Outlier And Analyte Summary Report

Outlier Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! AST	Arkray Autochem Xact Pro	30.10	U/L	2.32
! Total Protein	Arkray Autochem Xact Pro	5.68	g/dL	-2.01
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

G1 PATHOLOGY LAB SERVICES (559)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Instrument : Arkray Autochem Xact Pro


Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
! AST	U/L	30.10	--	2.32	2.03
@ Amylase	U/L	52.88	--	-1.91	-1.91
@ Bilirubin Total	mg/dl	0.53	--	-0.03	0.10
@ Bilirubin Direct	mg/dl	0.26	--	0.71	0.41
! Total Protein	g/dL	5.68	--	-2.01	-1.88
@ Albumin	g/dL	3.95	--	1.32	0.46
@ Uric Acid	mg/dl	4.59	--	1.43	1.18
@ Calcium	mg/dl	7.15	--	-1.06	-1.27
@ Triglycerides	mg/dl	164.00	--	0.84	0.26
@ ALT	U/L	22.97	--	-0.32	-0.39
@ ALP	U/L	78.97	--	0.54	0.50

Instrument : Cornley-Acculyte-3P Electrolyte Analyser

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	100.46	107.00	-1.75	-1.66
@ Sodium	mmol/L	186.46	--	1.04	1.71
@ Potassium	mmol/L	6.43	6.32	0.53	0.63

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	14
Not Evaluated Parameters	0
Evaluated Parameters	14
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070

End of Report

**G1 PATHOLOGY LAB SERVICES (559)
Outlier And Analyte Summary Report**

Outlier Details For Cycle No 210207 and Sample No 02

Report Date : 19/07/2021

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Amylase	Arkray Autochem Xact Pro	2.01	IU/L	*
! Sodium	Cornley-Acculyte-3P Electrolyte Analyser	184.71	mEq/L	2.37

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

G1 PATHOLOGY LAB SERVICES (559)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 02

Report Date : 19/07/2021

Instrument : Arkray Autochem Xact Pro

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ ALT	U/L	14.83	--	-0.46	-0.39
X Amylase	U/L	2.01	--	*	-1.91
@ Bilirubin Total	mg/dl	0.55	--	0.23	0.10
@ Bilirubin Direct	mg/dl	0.16	--	0.10	0.41
@ Total Protein	g/dL	5.60	--	-1.75	-1.88
@ Albumin	g/dL	3.56	--	-0.41	0.46
@ Uric Acid	mg/dl	4.18	--	0.92	1.18
@ Calcium	mg/dl	7.21	--	-1.48	-1.27
@ Triglycerides	mg/dl	145.94	--	-0.33	0.26
@ AST	U/L	26.10	--	1.73	2.03
@ ALP	U/L	78.97	--	0.45	0.50

Instrument : Cornley-Acculyte-3P Electrolyte Analyser

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	99.29	106.00	-1.56	-1.66
! Sodium	mmol/L	184.71	--	2.37	1.71
@ Potassium	mmol/L	6.31	6.16	0.73	0.63

Legend @ : z score ≤ 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score ≥ 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	14
Not Evaluated Parameters	0
Evaluated Parameters	14
Outlier Parameters (X)	1
EQAS Score Biochemistry	92.86 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Creatinine	Arkray Autochem Xact Pro	1.09	mg/dl	2.94
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

G1 PATHOLOGY LAB SERVICES (559)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 03

Report Date : 19/07/2021

Instrument : Arkray Autochem Xact Pro

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
! Creatinine	mg/dl	1.09	--	2.94	2.54
@ Glucose	mg/dl	58.00	--	-0.74	-0.98
@ Urea	mg/dl	22.70	--	-0.23	0.01
@ Cholesterol	mg/dl	158.45	--	0.15	0.33
@ HDL	mg/dl	35.21	--	-0.32	-0.32

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	5
Not Evaluated Parameters	0
Evaluated Parameters	5
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Creatinine	Arkray Autochem Xact Pro	0.91	mg/dl	2.13
<p>Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated 🕒 : Delayed Result Entry * : Not considered for evaluation.</p>				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

G1 PATHOLOGY LAB SERVICES (559)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 04


Report Date : 19/07/2021

Instrument : Arkray Autochem Xact Pro

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	61.00	--	-1.22	-0.98
@ Urea	mg/dl	21.98	--	0.25	0.01
! Creatinine	mg/dl	0.91	--	2.13	2.54
@ Cholesterol	mg/dl	157.27	--	0.50	0.33

Legend @ : z score ≤ 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score ≥ 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	4
Not Evaluated Parameters	0
Evaluated Parameters	4
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070

End of Report



G1 PATHOLOGY LAB SERVICES

G1PL - 07

Proficiency Testing outliers and corrective action records

PAGE 1 OF 1

Sr. No	Date	Test Name	Lab Reported Result	Peer / Expected Value	Details of Root Cause analysis and Corrective action taken	Done by	Sign
1.	06/01/21	Amylase	2.01	Not considered	QC data reviewed, no outlier found,	G. Khadake	
2.		Sample-2 Cycle-210207	JUL	for evaluation.	Sample 1 result within acceptable limit,		
3.					2. Sample issue as		
4.					peer is not provided by the PT provider-		
5.					No patient sample		
6.					affected. Review the		
7.					QC for correctness daily.		

Issue No. 1	Issue Date : 25-06-2021	Amend No. 00	Amend Date. 00	Page 1 of 1
 Prepared By : Mr. Suraj Rathod		 Reviewed & Approved By Mr. Gundappa Khadake		

CONTROLLED COPY