

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**G1 PATHOLOGY LAB SERVICES (559)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 210207 and Sample No 01**

**Report Date : 19/07/2021**

**Instrument : -**


Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Negative	Negative

**Instrument : Clinitex status +**

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	6,6.5
@ Specific Gravity	-	1.020	1.010,1.015,1.020
@ Glucose (urine)	-	2+	+1,+2,+3
@ Protein (Urine)	-	1+	+1,+2
@ Ketone (urine)	-	Negative	Negative
@ Blood (urine)	-	Negative	Negative
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
@ Nitrite (urine)	-	Present	Present

Legend @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	10
Not Evaluated Parameters	0
Evaluated Parameters	10
Outlier Parameters (X)	0
<b>EQAS Score Clinical Path - Urine Routine Chemistry</b>	<b>100.00 %</b>



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416.  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**G1 PATHOLOGY LAB SERVICES (559)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 210207 and Sample No 01**

**Report Date : 19/07/2021**


**Instrument : -**

Analyte	Standard Unit	Result Value	Accepted Value
@ Identify	-	Cyst of Giardia	Cyst of Giardia

Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry	Total Parameters	1
	Not Evaluated Parameters	0
	Evaluated Parameters	1
	Outlier Parameters (X)	0
	EQAS Score Clinical Path - Stool Routine Microscopy	100.00 %

Image



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416.  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

End of Report

**Analyte Methodwise Qualitative Result Report**

Lab No.: 559  
 Lab Name : G1 PATHOLOGY LAB SERVICES  
 Report Date : 19/07/2021


Cycle No : 210207  
 Sample No : 1  
 Sample Date : 01-07-2021

Clinical Path - Urine Sediment Analysis Image,Other findings, -

	N
<span style="color: purple;">■</span> All Methods	134
<span style="color: teal;">■</span> -	134
<span style="color: red;">■</span> -	134

<b>Your Result</b>	Calcium oxalate crystal
<b>Result Expected</b>	RBC

**Note** : # is an indication that this parameter has not been evaluated



PC-1022

**Analyte Methodwise Qualitative Result Report**


**Lab No.:** 559  
**Lab Name :** G1 PATHOLOGY LAB SERVICES  
**Report Date :** 19/07/2021

**Cycle No :** 210207  
**Sample No :** 1  
**Sample Date :** 01-07-2021

**Reported Results Legends :-**

RTC : Renal Tubular Epithelial Cells  
TEC : Transitional Epithelial Cells  
CaOx : Calcium oxalate crystal  
SEC : Squamous Epithelial Cells  
PC : Pus cells present  
RBC : Red Blood Cells  
None : None

**Comments :**



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416.  
Commercial Building - 1A,  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070



PC-1022

This is a computer generated report hence signature is not required



G1 PATHALOGY LAB SERVICES

G1PL - 07

Proficiency Testing outliers and corrective action records

PAGE 1 OF 1

Sr. No	Date	Test Name	Lab Reported Result	Peer / Expected Value	Details of Root Cause analysis and Corrective action taken	Done by	Sign
1.	11/0/21	Urine Sediment	Calcium	RBC	Data reviewed	G. Khadake	
2.		Analysis.	Oxalate crystal		It was found that		
3.					while uploading the		
4.					PT results, wrong		
5.					drop down was		
6.					selected, Instead of		
7.					RBC, Calcium oxalate		
					was selected.		
					CA: Review the data		
					entry by second		
					person.		

Issue No. 1	Issue Date : 25-06-2021	Amend No. 00	Amend Date. 00	Page 1 of 1
 Prepared By : Mr. Suraj Rathod		 Reviewed & Approved By Mr. Gundappa Khadake		

CONTROLLED COPY