

**CERTIFICATE OF CALIBRATION**

Instrument : **Mispa Ace Fully Automated Clinical Chemistry Analyser**
 Customer Name : **Meditech Diagnostic Centre, West Mambalam, Chennai**
 Date of Installation : **09/July/2020** S/N:WK-95106413

Alignment Check

Motion Check	Checked	Remarks
Sample Probe	☑	OK
Reagent Probe	☑	OK
Mixing Bar	☑	OK
Reaction Disk	☑	OK
Reagent Disk	☑	OK
Sample Disk	☑	OK
Lamp	☑	OK

System	Checked	Remark
Main unit	☑	OK
Temp. unit	☑	OK
Mixing unit	☑	OK
Reaction unit	☑	OK
Reagent unit	☑	OK
Sample unit	☑	OK

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattanamalam (PO), Dist. Ernakulam, Kerala - 683 562, India.
 Tel. + 91 494 286 7000 | Email: agappemh@agappe.in

MUMBAI (REGISTERED OFFICE)
 402 & 402, 4th Floor, Jai Singh Business Centre, 119,
 Sahar Road, Park Road, Khandivali (East), Mumbai - 400 099, India.
 Tel. +91 22 4300 8000 | Email: mumbai@agappe.in

EQUIPMENT PLANT
 X/705 CS, Block No. 37, KMFBA Sree Industrial Park,
 Noida, Cochin, Kerala, India - 686 728, Tel. +91 494 276 7477.

DELHI OFFICE
 DDA 140, 5th Floor, DLF Tower, Shivalik Marg,
 New Delhi - 110 015, India.
 Tel. +91 11 4558 8416 | Email: delhi@agappe.in

KOLKATA OFFICE
 406, Marla Maina, Plot No 03, Block DM, Sector V,
 Salt Lake City, Kolkata - 700 095.
 Tel. +91 33 4002 9433 | Email: kolkata@agappe.in

BANGALORE OFFICE
 5-6, 6th Floor, Red Cross Bldg, No 26 Race Course Road,
 Bangalore - 560 002, Tel. +91 80 2278 6288
 Email: bangalore@agappe.in



OPTICAL SECTIONS

Filter	Dark Current	Background
340	97	46099
405	100	45920
450	95	47034
510	92	49002
546	98	51085
578	94	51256
630	92	52013
670	91	55121
Reference		

Power Supply : 5V 12V 24V

The equipment is calibrated for the different parameters by using Multi

Calibrator Lot. No: 30020180 Exp: 07-2021

This is to certify that the above mentioned product has been calibrated according to the standard operating procedure given by the MINDRAY Medical Systems.

Date of Calibration : **07/July/2021**

Next Due on : **06/Jan/2022**

For Agappe Diagnostics Ltd,

M.Tex
Authorised signatory
Raja M
 Engineer - Customer Service

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CN : U24239MH1998PLC115413

CORPORATE OFFICE / REGIONAL PLANT
 Agappe Hills, Tattelman (PO), Dist. Ernakulam, Kerala - 686 942, India.
 Tel: +91 484 234 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)
 402 & 403, 4th Floor, Jangalwadi Business Centre, 119,
 Jangal Road, Jangalwadi, Andheri (E) Dist. Mumbai - 400 046, India.
 Tel: +91 22 4326 5100 | Email: Mumbai@agappe.in

EQUIPMENT PLANT
 X-258 CH, Block No. 32, KMP28 Small Industrial Park,
 Mettal, Coimbatore, Dist. Coimbatore - 686 001, Tel: +91 484 234 7477.

BELGAUM OFFICE
 222/242, 5th Floor, D-7 Tower, Shivaji Marg,
 New Belgaum - 591 025, India.
 Tel: +91 834 254 8600 | Email: Belgaum@agappe.in

KOLKATA OFFICE
 ACH, Westin Mall, Plot No-10, Block-CH, Sector 7,
 Salt Lake City, Kolkata - 700 091,
 Tel: +91 33 4001 0471 | Email: Kolkata@agappe.in

BANGALORE OFFICE
 5-4, Dth Floor, Steel Cross Station, No.24, Rajarajeshwari Road,
 Bangalore - 560 091, Tel: +91 80 2279 6100
 Email: Bangalore@agappe.in