

Ortho Clinical Diagnostics

403, Leela Business Park,
Andheri Kurla Road, Andheri
East, Mumbai - 400059
T : +91 22 67879300
F : +91 22 67879333

CALIBRATION CERTIFICATE

CERTIFICATE NO : OCD/PRIMA/02
CALIBRATED FOR : PRIMA DIAGNOSTICS
LOCATION : BANGLORE

CALIBRATION DATE : 13/04/2021
CALIBRATION DUE : 12/10/2021
CALIBRATED AT : AT SITE

SPECIFICATION OF INSTRUMENT :

INSTRUMENT	MODEL	SERIAL NO
VITROS CHEMISTRY SYSTEM	VITROS 250	27001264
	J NUMBER	J20705

Specification of Software:

SOFTWARE NAME	SOFTWARE VERSION	UPDATED ON
DOS Rmx os	V9.6	March 2019

The Reference of Calibration: The performance/Adjustments of various subsystems has been calibrated/tested by trained site engineer with pre-designed /calibrated tools provided for the particular subsystems by the company.

PM Procedure Checklists:

Subsystem Name

Adjusted/Verified

1. Sample Supply
* Hook Drive Belt

2. Slide Supply 1, 2 & slide transport
Mechanical Adjustment:
* Dispense Blade Centering Adjustment
* Dispense Blade Drive Belt Tension
* Plunger Centering Alignment
Software Adjustment:
* Dispense To Metering Position

3. Sample Metering
Mechanical Adjustment:
* Truck Sensor

* Truck Guide Button
* Metering Truck

Ortho Clinical Diagnostics

403, Leela Business Park,
Andheri Kurla Road, Andheri
East, Mumbai – 400059
T : +91 22 67879300
F : +91 22 67879333

Software Adjustment:

- * Tip Pickup X,Y and Height
- * Sample Tray at Sample Position –Auto Sample Tray
- * Proboscis Centering and Sample Cup Depth
- * Mixing Cup X,Y and its Depth
- * Diluents Tray
- * Rate / Colorimetric Metering Position
- * Potentiometric Metering Position
- * Proboscis Compression
- * Wetness Detector



6. Slide Incubator

Software Adjustment:

- * Incubator Disk Sizing
- * Incubator Disk Stopping
- * PM Slide Insertion Depth
- * Slide Read Position (RT/CM Depth & Read Sync)
- * Potentiometric Eject Blade Depth



7. Reflectometer

- * IRIS Lamp Output Voltage
- * Correction Factor




9. Master Computer

- * Touch Screen Calibration
- * System Full Backup



CALIBRATED BY


Yogeesh K
SR.Service engineer
Orthoclinical diagnostics
Banglore

NOTE:

1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in the certificate is valid at the time of and under the stated condition of the measurement.

Calibration Report

Customer Name :- Prima Diagnostics, Bangalore

**Address:-Siddapura Village, Varthur, White Field Main Road,
Bangalore 560 066.**

Instrument Model:- ST-200 CL Electrolyte Analyzer /Sr. No.:-ST-4537

Calibration Date:- 19/08/2021

Calibration Due Date:- 18/08/2022

As per the Standard Operating procedure of ST-200 CL Electrolyte Analyzer, the instrument has been calibrated at the factory.

As per method mentioned into manuals/sop, the Known Value Sample/Control run on the analyzer for five times for the Electrolyte parameters. All obtained values are in the range and CV values were checked. CV values were within a normal range as mentioned in performance certificate. This indicated that an instrument is performing accurately and precisely.

Hence, we certified that ST-200 CL Electrolyte Analyzer is calibrated.



Kiran Date
(General Manager-Global Service)

Sensa Core Medical Instrumentation Pvt. Ltd.



Instrument Name : ST 200 CL Na/K/Cl Electrolyte Analyzer

Instrument Number : ST-4537 **"K" Electrode :** 21FF26
"Na" Electrode : 21FC25 **"Cl" Electrode :** 21FC30

Calibration Slope :

S.No	Na+	K+	Cl-
1	64.33	66.88	54.97
2	64.98	66.53	55.34
3	64.65	66.65	55.75
Mean	64.65	66.69	55.35
Ranges	35 - 100 mV/decade	35 - 100 mV/decade	30 - 100 mV/decade
Status	OK	OK	OK

Quality Control (I) : Lot:UQC21F-L1

S.No	Na+	K+	Cl-
1	126.01	2.14	75.50
2	126.02	2.15	75.51
3	126.50	2.16	75.53
Mean	126.18	2.15	75.51
Ranges	112 - 132 mmol/L	1.48 - 2.48 mmol/L	74 - 92 mmol/L
Status	OK	OK	OK

Quality Control (II) : Lot:UQC21F-L2

S.No	Na+	K+	Cl-
1	141.01	3.87	96.40
2	140.90	3.88	96.41
3	141.30	3.89	97.45
Mean	141.07	3.88	96.75
Ranges	132 - 149 mmol/L	3.49 - 4.49 mmol/L	95 - 110 mmol/L
Status	OK	OK	OK

Quality Control (III) : Lot:UQC21F-L3

S.No	Na+	K+	Cl-
1	159.40	5.75	120.20
2	159.41	5.76	121.20
3	159.45	5.78	122.10
Mean	159.42	5.76	121.17
Ranges	150 - 169 mmol/L	5.45 - 6.45 mmol/L	114 - 128 mmol/L
Status	OK	OK	

BIO RAD Level (I) : Lot: 26431

S.No	Na+	K+	Cl-
1	150.50	4.11	95.70
2	150.80	4.13	94.90
3	150.65	4.11	96.65
Mean	150.65	4.12	95.75
Ranges	140 - 161 mmol/L	3.79 -4.43 mmol/L	91 - 113 mmol/L
Status	OK	OK	OK

BIO RAD Level (II) : Lot: 26432

S.No	Na+	K+	Cl-
1	134.90	6.22	83.50
2	134.90	6.24	83.50
3	134.65	6.23	84.65
Mean	134.82	6.23	83.88
Ranges	123 - 137 mmol/L	5.78 - 6.80 mmol/L	77 - 94 mmol/L
Status	OK	OK	OK

Urine Quality Control (I) : Lot:UQC21F-UL-1

S.No	Na+	K+	Cl-
1	383.26	67.35	359.12
2	382.65	68.27	356.77
3	379.46	68.05	358.61
Mean	381.79	67.89	358.17
Ranges	280 – 450 mmol/L	55 -85 mmol/L	300-450 mmol/L
Status	OK	OK	OK

Urine Quality Control (II) : Lot:UQC21F-UL-2

S.No	Na+	K+	Cl-
1	671.88	358.37	819.22
2	674.57	360.19	819.42
3	672.95	359.65	820.65
Mean	673.13	359.40	819.76
Ranges	580 -750 mmol/L	280 - 500 mmol/L	700 - 900 mmol/L
Status	OK	OK	OK

100. 000000000000000000

100. 000000000000000000
100. 000000000000000000
100. 000000000000000000

100. 000000000000000000
100. 000000000000000000
100. 000000000000000000

Date	Time	Int. No.	Temp.	Instrument	Response	Figure
02/20/21	18:51	0007-2054	29.83	0.23370	1.05473	1.05473
03/26/21	11:38	0008-4350	28.91	0.23310	1.07870	1.07870
04/14/21	13:53	0008-4350	27.88	0.25553	1.08920	1.08920
05/26/21	18:39	0009-7440	27.88	0.23107	1.06371	1.06371
06/23/20	16:08	0002-4238	28.23	0.23125	1.03352	1.03352
12/22/20	16:52	0001-4238	28.27	0.23370	0.99998	1.50780

100. 000000000000000000
100. 000000000000000000

100. 000000000000000000
100. 000000000000000000

100. 000000000000000000

THE UNITED STATES DISTRICT COURT

WESTERN DISTRICT OF TEXAS

Case No. 03-10001
In re: [Name]
Debtor

Chapter 11
Case No. 03-10001
Date Filed: 03/10/03

Page 11 of 11

Date	Time	Lot Num	Ycomp	1	2	3	4
05/09/01	10:03	0131-4257	28.05	0.11442	0.93903	1.76903	
05/25/01	11:51	0131-4257	27.55	0.12985	0.94740	1.76140	
09/01/01	13:03	0131-7447	28.55	0.13551	0.99865	1.67016	
12/22/00	16:08	0129-3218	28.23	0.13416	0.95539	1.71661	
12/31/00	16:19	0129-3218	27.92	0.12590	0.97103	1.68237	
02/24/01	18:50	0131-6102	29.63	0.13903	0.90351	1.67997	
03/26/01	17:25	0131-6102	28.74	0.14734	0.90863	1.74445	

AREA REPORT

>Touch - target to select date for display.

Id: 0031 WEST Lat: 44.12
 E: 0031 EAST Lon: 101.12

DEC EARLY LOW DISCOUNT

LAB CORP
 4501 Hwy 110, Suite 110
 Little Rock, AR 72205
 Telephone: 501-223-1100

FORM 1041
 400 - 101
 10-1-1975
 10-1-1975
 10-1-1975

UNITARY TRUSTS - Estate of Low Discount
 Description Date

Date	Time	Lot No.	Time	Rate	Low Discount	Relinquish	Rate
01/24/21	18:51	1531-7452	29.81	0.00302	0.00910	0.06251	
04/14/21	18:28	1531-7452	20.05	0.00303	0.00925	0.06373	
04/09/21	10:01	1536-4121	27.80	0.00300	0.00909	0.06261	
09/01/21	1:45	1536-7451	28.50	0.00304	0.00923	0.06308	
12/22/20	16:09	1531-6421	28.53	0.00293	0.00912	0.06158	
12/31/20	10:07	1531-6421	27.22	0.00277	0.00827	0.05932	
12/31/20	11:40	1531-6421	28.09	0.00302	0.00937	0.06005	
03/24/21	11:13	1531-6621	28.91	0.00311	0.00950	0.06421	

CREA GROUP

March 2 1975 - to select date for details.

15-0430 NEXT 104 104 104 104

THE TV/R NR DIS-ABLED

QUALITY DIRECT - California District
B-Supervisors

Page 1 of 1
12/31/20
12/31/20

Date	Time	Loc. Area	Time	Page 01 of 02			
				1	2	3	4
05/05/21	10:00	0549-4987	28.08	0.23005	0.68179	1.38091	
06/02/20	13:56	0545-2511	27.22	0.18651	0.66625	1.28729	
08/08/20	14:13	0545-2511	28.23	0.17518	0.64674	1.31641	
09/25/20	10:50	0545-2511	27.39	0.17590	0.68944	1.33067	
10/12/20	09:34	0545-2511	28.25	0.17370	0.67522	1.31141	
11/17/20	14:49	0545-2511	28.91	0.17618	0.68068	1.33743	
12/22/20	16:09	0538-4971	28.05	0.18843	0.68448	1.34325	
12/22/20	16:17	0538-4971	26.11	0.18375	0.68758	1.33520	
12/31/20	10:32	0549-4971	27.06	0.17322	0.67848	1.31950	
12/31/20	11:47	0538-4971	28.08	0.20218	0.68700	1.31768	

UNIT SECT

Touch a field to select data for display.

FIELD

OK
PAGE

Loc
Page: 1 of 1

Field

PRINTER RESOURCE # space unavailable

Serial ID# 800 01 21

Low Library

Low DPE

Page 2 of 4

THE ENVIRONMENTAL

RESOURCES

00000000

WORLDWIDE DISTRIBUTION HISTORY
Resource Data

Page 01 of 01

Date	Time	Lot No	Lot	Lot	Lot	Lot	Lot
08/06/21	11:30	0605-3153	26.74	0.54133	0.28518	0.16-632	
09/21/20	19:13	0612-4181	28.13	0.58270	0.29145	0.17327	
10/17/18	10:15	0612-4081	28.40	0.52887	0.28294	0.14104	
01/03/21	09:26	0603-7178	29.81	0.58604	0.28031	0.19034	
05/25/21	11:07	0617-4100	27.39	0.58214	0.27804	0.19112	

11: SE#00

Storage 2: Error: No select data for display.

PREV PAGE
1 of 1
Page 01 of 01
PRINT

THIRTEEN MONTHS * See e-mail...

Serial No: 500-03-21
Date: 12/1/10
Page: 12 of 13

MAC DRIVER NOW DISABLED

17/01/2007
RESIDUAL COMPLETE

QUALITY CONTROL - Calibration History
Response Data

Date	Time	Lot Data	Temp	Page 01 of 01			
				1	2	3	4
08/21/14	09:01	0928-8807	25.51	0.60311	0.85587	1.01187	
09/02/14	15:32	0928-8807	23.86	0.60236	0.85506	1.01162	
01/21/20	23:34	0928-8807	27.53	0.60732	0.85789	1.01381	
02/07/20	02:40	0928-8807	26.21	0.60602	0.85610	1.02513	
09/21/20	14:13	0927-6073	28.73	0.60982	0.84701	1.01906	
10/14/20	10:16	0927-6073	28.40	0.60489	0.85375	1.00527	
11/27/20	17:03	0927-6073	28.78	0.60742	0.84233	1.01147	
08/25/21	11:08	0928-6591	27.38	0.60270	0.85887	1.04365	
01/21/19	4:23	0923-3462	26.79	0.60330	0.84800	0.92091	
08/26/19	12:09	0913-3-61	25.36	0.60219	0.80805	0.93881	

APP: SUPD4

*Touch a target to select date for display.

08/21/14	09:01	0928-8807	25.51	0.60311	0.85587	1.01187	
09/02/14	15:32	0928-8807	23.86	0.60236	0.85506	1.01162	
01/21/20	23:34	0928-8807	27.53	0.60732	0.85789	1.01381	
02/07/20	02:40	0928-8807	26.21	0.60602	0.85610	1.02513	
09/21/20	14:13	0927-6073	28.73	0.60982	0.84701	1.01906	
10/14/20	10:16	0927-6073	28.40	0.60489	0.85375	1.00527	
11/27/20	17:03	0927-6073	28.78	0.60742	0.84233	1.01147	
08/25/21	11:08	0928-6591	27.38	0.60270	0.85887	1.04365	
01/21/19	4:23	0923-3462	26.79	0.60330	0.84800	0.92091	
08/26/19	12:09	0913-3-61	25.36	0.60219	0.80805	0.93881	

FRIBTER CONTROL # (see Federal of Time)

100-10000

9000100

1000000

THE ENGINE NOT DISABLED

FRIBTER CONTROL #

100-100

1000000

FRIBTER CONTROL # (see Federal of Time)
Restriction Data

Date	Time	Lat	Long	Alt	Roll	Yaw	Rate of Roll
06/02/20	13:46	0333-6482	27.06	0.64035	1.07207	1.29505	0
06/08/20	14:14	0333-6482	28.23	0.64213	1.07605	1.29607	0
11/27/20	15:45	0333-6482	28.71	0.64207	1.08407	1.30005	0
12/17/20	14:13	0333-6482	29.09	0.64738	1.04507	1.27405	0
10/17/20	09:34	0339-6787	28.23	0.63024	1.06491	1.10000	0
10/17/20	11:40	0339-6787	28.35	0.63006	1.06133	1.10000	0
12/22/20	16:10	0339-6787	28.27	0.63122	1.07205	1.10000	0
12/22/20	16:53	0339-6787	26.27	0.62382	1.07907	1.29505	0
03/28/21	10:46	0339-6787	29.36	0.63776	1.07490	1.30004	0
01/14/21	11:19	0347-6887	27.78	0.62327	1.05614	1.31005	0

Use SE, 000

Touch a target to select date for display

FRIBTER CONTROL # (see Federal of Time) Alt Roll Yaw Rate of Roll

Date	Time	Lot No.	Loan	1	2	3	4
01/25/21	11:09	0203-2050	27.39	0.412543	0.472843	0.49079	0.81127
01/21/20	23:34	0245-2149	27.29	0.428387	0.46904	0.46909	0.88153
02/02/20	02:41	0248-2147	26.44	0.412758	0.44155	0.49155	0.82452
09/21/20	19:14	0242-2959	26.21	0.40400	0.48428	0.50793	0.67156
10/17/20	10:17	0247-2959	28.23	0.41881	0.47907	0.52187	0.64135
11/27/20	17:00	0247-2959	28.74	0.43004	0.47291	0.47296	0.82867
01/23/19	02:06	0236-3919	26.74	0.43304	0.46040	0.48638	0.81275
06/19/19	13:01	0235-3919	25.51	0.44122	0.46612	0.49057	0.88170
08/26/19	17:10	0235-3919	25.51	0.43926	0.46657	0.49177	0.83970

...
 ...
 ...

ITEM	DESCRIPTION	DATE	AMOUNT	...

Date	Time	Lot #	Rate	Instrument		Response	
				1	2	1	2
05/25/21	11:09	0203-5050	27.37	0.12341	0.08097	0.81767	0.67143
04/21/20	23:31	0203-2149	27.22	0.12838	0.48908	0.80136	0.64958
07/07/20	02:41	0203-2149	26.43	0.12438	0.49105	0.8773	0.56155
09/21/20	19:03	0247-4959	28.53	0.10200	0.50790	0.81136	0.58101
10/17/20	10:17	0247-6959	28.28	0.11941	0.52187	0.80736	0.59207
11/27/19	17:00	0247-6959	28.78	0.15081	0.49548	0.87707	0.62291
05/24/19	07:06	0236-3919	26.74	0.13301	0.48635	0.80736	0.59207
02/19/19	13:41	0236-3919	25.51	0.19110	0.49397	0.83411	0.66613
08/26/19	17:10	0236-3919	25.51	0.18924	0.49115	0.83971	0.67555

RE: SEARCH
 Search a target to select data for analysis.

MS EXPR DR ROM disabled

END OF FILE

17:45:37

02/12/87

UTILITY PROGRAM - Calibration History Response Data

Page 01 of 03

Date	Time	Lot No	Temp	Instrument			Response
				1	2	3	
01/21/20	23:23	1434-0725	27.22	0.39936	0.39820	0.39843	
02/07/20	06:40	1434-0725	26.43	0.39956	0.39997	0.39834	
07/21/20	19:12	1437-2220	28.05	0.46310	0.39603	0.42021	
10/17/20	10:16	1437-2220	28.25	0.47178	0.41344	0.32737	
12/29/20	17:35	1437-2606	28.91	0.48100	0.39047	0.49446	
01/05/21	08:26	1437-2606	29.61	0.46973	0.49344	0.41305	
02/13/21	20:21	1437-2304	29.46	0.48073	0.39800	0.39847	
04/25/21	11:08	1433-3474	27.39	0.47325	0.39491	0.37321	

CELL SERIAL

*Touch a bar on the screen to select data for display

RE: 0004

RE: 0004

RE: 0004

RE: 0004

RE: 0004

QUALITY CONTROL & communication error

00011147

Sep 03 1994

END CORRECT

001 OFF

17:45:56

THE ENGINE ROOM DISABLED

TESTING COMPLETE

QUALITY CONTROL - Calibration History

001

001

Response Data

Date	Time	Lot No	Temp	Response			
				1	2	3	4
02/13/91	20:28	7325-2467	29.46	0.00009	0.00048	0.00043	
05/10/91	12:53	7327-4374	27.19	0.00021	0.00044	0.00046	
05/20/91	11:32	7327-4374	27.19	0.00021	0.00048	0.00042	
07/18/90	21:55	7325-3274	24.74	0.00007	0.00069	0.00057	
10/17/90	10:04	7326-4074	26.23	0.00002	0.00052	0.00053	
12/17/90	19:12	7325-4096	24.69	0.00035	0.00050	0.00040	

END SCREEN

Touch a digit to select data for update

MAILBOX CONTROL * communications error

Serial 1153

Serial 03 20

Log Control

On Off

17:44:13

THE FUTURE FOR DISASTER

TESTING COMPLETE

0.000000

GLOBAL TV CONTROL - California History

0.000000

Response Data

Page 04 of 04

Date	Time	Log Num	Temp	1	2	3	4
06/30/19	10:23	5602-6373	23.05	0.10653	0.11378	0.13091	
05/28/19	21:17	5602-6373	28.91	0.10657	0.11378	0.13091	
03/12/20	10:59	5602-6373	26.90	0.10658	0.11378	0.13091	
06/12/20	11:44	5602-6373	27.39	0.10658	0.11378	0.13091	
09/12/20	21:43	5602-6373	25.81	0.10658	0.11378	0.13091	
00/00/00	00:00	5602-4781	0.00				
08/18/20	11:42	5602-4780	26.40	0.10657	0.11378	0.13091	
10/17/20	16:04	5602-4780	26.23	0.10657	0.11378	0.13091	
02/13/21	20:28	5602-4780	29.46	0.10657	0.11378	0.13091	
05/19/21	12:19	5602-4323	27.39	0.10657	0.11378	0.13091	

ALTY SERVO

Touch a target to select data for analysis.

RETRIEVE

NEXT PAGE

Exit CONTINUE

HELP

HIGHLIGHT CONTROL * recalibration error

DATE TIME 8:00 03 17:43:06
PAGE 1 OF 1

THE EDGE: NOT DISABLED

TOP COPPER
RENTING CORP. LLC

QUALITY CONTROL - Calibration History
Response Det.

Page 01 of 01

Date	Time	Lot Num	Temp	1	2	3
01/14/20	02:48	6523-9301	28.56	0.00070	0.00513	0.00225
01/14/20	02:55	6523-9301	28.56	0.00073	0.00514	0.00334
11/20/20	18:23	6523-5358	29.27	0.00054	0.00350	0.00376
02/13/20	21:56	6524-3423	26.73	0.00051	0.00357	0.00271
11/17/20	10:05	6524-3423	28.43	0.00052	0.00352	0.00320
11/27/20	16:03	6524-3423	26.71	0.00075	0.00354	0.00488
01/10/21	12:53	6523-3419	27.39	0.00077	0.00375	0.00474

END OF REPORT

Touch screen to select data for display.

RETD:01 0.001 0.001
 PRE:01 0.001 0.001
 0.001 0.001 0.001

NOIP FOR CONTROL & communications error

Test: 10000%
TESTING COMPLETE

Serial 114- 100 01
00 100 124 24
1997 10 10

THE EM/IR NOW DISABLED

NOIP CONTROL -- Calibration History
Q-Source Data

Date	Time	Full Bus	Temp	Instrument	Response	Page 01 of 01
11/05/19	22:03	0242-9131	27.08	0.81201	1.13741	4
01/22/20	27:08	0242-9131	27.35	0.81201	1.28034	4
02/03/20	21:08	0242-9131	27.49	0.81201	1.59114	4
10/17/20	09:10	0243-2833	28.53	0.81201	1.52522	4
12/22/20	17:41	0245-6124	27.02	0.81201	1.51416	4
04/03/21	11:12	0245-6124	18.28	0.81201	1.62906	4
06/22/19	14:21	0241-3792	25.24	0.81201	1.86124	4

IRIS SCRIPT

>touch - barrel to select date for display.

IRIS SCRIPT

TEST
PASS

CAL.
PASS/FAIL

10/10

LINE ENVELOPE NOW DISABLED

JOB CODE: 0831183 COMPLETE

SAMPLE TIME: 001 001

DATE: 01/17/83

QUALITY CONTROL - Calibration History
Business Date

09.7 0.6
0.01

Date	Time	Lot No.	Temp.
08/26/79	19:57	0843-5863	23.54	0.56242	0.53014	1.20207	4
09/03/79	08:59	0843-5863	23.84	0.54123	0.51318	1.19800	
11/06/79	22:02	0843-5863	27.32	0.51254	0.48411	1.18913	
01/22/80	22:54	0843-0945	27.79	0.49707	0.46867	1.18624	
02/03/80	21:05	0843-0945	27.82	0.47724	0.44927	1.18110	
08/29/80	10:48	0843-0945	28.23	0.39946	0.38832	1.20112	
10/17/80	07:49	0843-0124	28.23	0.45236	0.41309	1.20507	
05/03/79	08:54	0819-3046	26.90	0.48734	0.46788	1.23355	
06/22/79	14:09	0819-3046	26.09	0.50568	0.47054	1.24453	
07/01/79	21:20	0819-3046	27.06	0.48759	0.46973	1.26130	

END SCREEN

Touch a target in select list for display.

RETURN

11:11
5.4

CAL
PARAMETERS

INC EMER POLI DISABLED

IMMEDIATELY CONFIDENTIAL - Calibration History
Response Data

Date	Time	Lot No	Temp	1	2	3	Page 01 of 01
12/19/20	07:46	1125-6685	23.40	0.37231	0.59264	0.85815	
01/18/21	11:43	1125-6686	28.13	0.35021	0.60186	0.86021	
05/25/21	12:00	1129-4918	27.19	0.31886	0.59013	0.85852	
05/26/21	11:11	1129-4918	26.06	0.34061	0.59016	0.84579	
06/11/20	17:15	1123-9823	25.81	0.39357	0.51942	1.00007	
08/24/20	11:41	1123-9823	26.05	0.39026	0.60708	0.86771	
10/12/20	10:26	1123-1137	29.41	0.31137	0.60340	0.87838	

UNIT SERIAL

Touch a screen to select a file for display.

RETURN

OK
HOME

CALL
FOR HELP

HELP