



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CERTIFICATE NO. : CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER	:-	Page No.	:- 1 of 1
sadhbhavan pathology lab		Certificate No.	:- GTS/210526/01-003
Shop no 3, Trimurti Complex, Opp to SBI,		Date of Received	:- 26.05.2021
Pune Hwy, next to Sharad Sahakari Bank,		Date of Calibration	:- 26.05.2021
Bail Bazar, Manchar, Maharashtra, 410503.		Next Calibration Due On	:- 25.05.2022
Ambient Temp. (°C)	:- 28.7	Issue Date	:- 28.05.2021
Relative Humidity (%RH)	:- 49	Calibration method No.	:- MECH-WI-07
Location of calibration	:- On Site	ULR No	:- CC295721000004074F
Condition of Item	:- Ok		

2. Description of Item

Name	:- Centrifuge	Range	:- 0-3500 Rpm
Sr No	:- CF-01	Least Count	:- -
Make	:- Labron	Location	:- PATHOLOGY
Type	:- Knob		

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Digital Tachometer	NI/2006/002/002	Nishitronics Instrumentation	GTS/TM/01	01.06.2021

*Mechanical Calibration

4.Calibration Results :-

Sr No	Standard Reading RPM	Knob No on UUC	Expanded Uncertainty in ± %
1	847.8	1	1.20
2	1443	2	1.20
3	2415	3	1.20
4	2804	4	1.20
5	3445	5	1.20

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
- 6) Conformity With Requirements :-

Calibrated By

Calibration Engineer
D Ballal



Approved By

Technical Manager
Swapnil Bhagwat

End of Certificate

RF-55/01