

STATE REFERENCE LABORATORY  
Department of Microbiology  
Naga Hospital Authority Kohima, Nagaland

To,  
The Medical Officer  
i/c ICTC,  
C.H.C Changtongya

Subject: EQAS PT report for HIV SEROLOGY conducted by SRL, Kohima under EQAS program 2020-21(Round- 1) panels distributed on 17th November 2020.,

Madam/Sir,

I am here by reporting the result of the 4 blind panels submitted by your ICTC.

Sample	Test 1	Test II	Test III	ICTC Result	SRL Sample	Remarks
	Method: Immunodot Test Kit: Comb Aids	Immunochromat ography Meriscreen	Immunoconcentrati on Tredro			
MK-CT-1	Reactive	Reactive (HIV-1)	Reactive (HIV-1)	Positive	Positive	Tally
MK-CT-2	Reactive	Reactive (HIV-1)	Reactive (HIV-1)	Positive	Positive	Tally
MK-CT-3	Non Reactive	-----	-----	Negative	Negative	Tally
MK-CT-4	Non Reactive	-----	-----	Negative	Negative	Tally


Comment: Performance is Satisfactory

This is for your information and necessary action.

Amended report Date : 31<sup>st</sup> Aug 2021

Yours' Faithfully,



  
(Dr. Margaret Yhome),  
i/c State Reference Laboratory  
Naga Hospital Authority Kohima

The Laboratory technician has to maintain this letter as a feedback from SRL, NHAK.

STATE REFERENCE LABORATORY  
Department of Microbiology  
Naga Hospital Authority Kohima, Nagaland

To,  
The Medical Officer  
i/c ICTC,  
C.H.C Changtongya

Subject: EQAS PT report for HIV SEROLOGY conducted by SRL, Kohima under EQAS program 2020-21(Round- 2) panels distributed on 24th March 2021.

Madam/Sir,

I am here by reporting the result of the 4 blind panels submitted by your ICTC.

Sample	Test 1	Test II	Test III	ICTC Result	SRL Sample	Remarks
	Method: Immunodot Test Kit: Comb Aids	Immunochromat ography Meriscreen	Immunoconcentrati on Tredro			
MK-CT-1	Non Reactive	-----	-----	Negative	Negative	Tally
MK-CT-2	Non Reactive	-----	-----	Negative	Negative	Tally
MK-CT-3	Reactive	Reactive (HIV-1)	Reactive (HIV-1)	Positive	Positive	Tally
MK-CT-4	Reactive	Reactive (HIV-1)	Reactive (HIV-1)	Positive	Positive	Tally

Comment: Performance is Satisfactory

This is for your information and necessary action.

Amended report Date : 31<sup>st</sup> Aug 2021



Yours' Faithfully.

(Dr. Margaret Yhome),  
i/c State Reference Laboratory  
Naga Hospital Authority Kohima

The Laboratory technician has to maintain this letter as a feedback from SRL, NHAK.