

CERTIFICATE OF CALIBRATION

Instrument : Mispa Ace Fully Automated Clinical Chemistry Analyser
 Customer Name : Meditech Diagnostic Centre, West Mambalam, Chennai
 Date of Installation : 09/July/2020 S/N:WK-95106413

Alignment Check

Motion Check	Checked	Remarks
Sample Probe	<input checked="" type="checkbox"/>	OK
Reagent Probe	<input checked="" type="checkbox"/>	OK
Mixing Bar	<input checked="" type="checkbox"/>	OK
Reaction Disk	<input checked="" type="checkbox"/>	OK
Reagent Disk	<input checked="" type="checkbox"/>	OK
Sample Disk	<input checked="" type="checkbox"/>	OK
Lamp	<input checked="" type="checkbox"/>	OK

System	Checked	Remark
Main unit	<input checked="" type="checkbox"/>	OK
Temp. unit	<input checked="" type="checkbox"/>	OK
Mixing unit	<input checked="" type="checkbox"/>	OK
Reaction unit	<input checked="" type="checkbox"/>	OK
Reagent unit	<input checked="" type="checkbox"/>	OK
Sample unit	<input checked="" type="checkbox"/>	OK

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattinattom (PO), Dist., Ernakulam, Kerala - 681 562, India.
 Tel: +91 484 285 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 1/500-CO, Block No. 32, KINFRA Small Industrial Park,
 Nellikud, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
 406, Merlin Matrix, Plot No-10, Block-0N, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
 401 & 402, 4th Floor, Jai Singh Business Centre, 119,
 Sahar Road, Park Road, Andheri (East), Mumbai - 400 099, India.
 Tel: +91 22 4300 8000 | Email: mumbai@agappe.in

DELHI OFFICE
 D5M 540, 5th Floor, DLF Tower, Shivaji Marg,
 New Delhi - 110 015, India.
 Tel: +91 11 4538 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
 5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
 Bangalore - 560 001. Tel: +91 80 2228 8288
 Email: bangaloreoffice@agappe.in

**OPTICAL SECTIONS**

Filter	Dark Current	Background
340	97	46099
405	100	45980
450	95	47234
510	92	49007
546	98	51085
578	94	51256
630	97	53013
670	91	55121
Reference		

Power Supply : 5V 12V 24V

The equipment is calibrated for the different parameters by using Multi

Calibrator Lot. No: 31010077 Exp: 08-2022

This is to certify that the above mentioned product has been calibrated according to the standard operating procedure given by the MINDRAY Medical Systems.

Date of Calibration : **07/July/2021**

Next Due on : **06/Jan/2022**

For Agappe Diagnostics Ltd,

M.T.G.
Authorised signatory

Raja M

Engineer - Customer Service

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Oud. Ernakulam, Kerala - 683 962, India.
Tel: +91 484 285 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
K/SEB-CB, Block No. 72, KINFRA Small Industrial Park,
Nellai, Cochin, Kerala, India - 689 721. Tel: +91 464 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DK, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0411 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jai Singh Business Centre, 119,
Sabar Road, Parsipada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbai@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhi@agappe.in

BANGALORE OFFICE
5-6, 1st Floor, Red Cross Bhawan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8388
Email: bangaloreoffice@agappe.in