

STATE REFERENCE LABORATORY  
Department of Microbiology  
Naga Hospital Authority Kohima, Nagaland

Naga Hospital Authority Kohima

To,  
The Medical Officer  
i/c ICTC,  
DH Wokha

Subject: EQAS PT report for HIV SEROLOGY conducted by SRL, Kohima under EQAS program 2020-21(Round- 1) panels distributed on 17th November 2020..

Madam/Sir,

I am here by reporting the result of the 4 blind panels submitted by your ICTC.

Sample	Test 1	Test II	Test III	ICTC Result	SRL Sample	Remarks
	Method:	Immunochromatography	Immunoconcentration			
	Test Kit: Comb Aids	Meriscreen	Tredro			
WK-DH-1	Non Reactive	-----	-----	Negative	Negative	Tally
WK-DH-2	Reactive	Reactive (HIV-1)	Reactive (HIV-1)	Positive	Positive	Tally
WK-DH-3	Reactive	Reactive (HIV-1)	Reactive (HIV-1)	Positive	Positive	Tally
WK-DH-4	Non Reactive	-----	-----	Negative	Negative	Tally

Comment: Performance is Satisfactory

This is for your information and necessary action.

Amended report Date : 31<sup>st</sup> Aug 2021

Yours Faithfully.



(Dr. Margaret Yhome), i/c  
State Reference Laboratory  
Naga Hospital Authority Kohima

The Laboratory technician has to maintain this letter as a feedback from SRL, NHAK.

STATE REFERENCE LABORATORY  
Department of Microbiology  
Naga Hospital Authority Kohima, Nagaland

To,  
The Medical Officer  
i/c ICTC,  
DH Wokha

Subject: EQAS PT report for HIV SEROLOGY conducted by SRL, Kohima under EQAS program 2020-21(Round- 2) panels distributed on 24th March 2021.

Madam/Sir,

I am here by reporting the result of the 4 blind panels submitted by your ICTC.

Sample	Test 1	Test II	Test III	ICTC Result	SRL Sample	Remarks
	Method: Immunodot	Immunochromatography	Immunoconcentration			
	Test Kit: Comb Aids	Meriscreen	Tredro			
WK-DH-1	Reactive	Reactive (HIV-1)	Reactive (HIV-1)	Positive	Positive	Tally
WK-DH-2	Reactive	Reactive (HIV-1)	Reactive (HIV-1)	Positive	Positive	Tally
WK-DH-3	Non Reactive	-----	-----	Negative	Negative	Tally
WK-DH-4	Non Reactive	-----	-----	Negative	Negative	Tally

Comment: Performance is Satisfactory

This is for your information and necessary action.

Amended report Date : 31<sup>st</sup> Aug 2021

Yours Faithfully.



(Dr. Margaret Yhome),  
i/c State Reference Laboratory  
Naga Hospital Authority Kohima

The Laboratory technician has to maintain this letter as a feedback from SRL, NHAK.