



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT

PT SCHEME: SEROLOGY



PC-1047

Cycle No	C3
Ref.No.	NEUQAP109
Sample ID	NEUQAP SERO SERUM/2021/7/C3/S3

Report Date : 10/08/2021

Sample : July 2021

All Methods

Parameters	Lab Report	No of Participants	Consensus Report	%	Remarks
Brucella	Not Recvd.	24	Non Reactive - 24		
Brucella Antigen	Not Recvd.	3	Non Reactive - 3		
Brucella IgG	Not Recvd.	17	Non Reactive - 17		
Brucella IgM	Not Recvd.	18	Non Reactive - 18		
Chikungunya	Not Recvd.	42	Negative - 41 Positive - 1		
CMV IgG	Not Recvd.	42	Reactive - 42		
CMV IgM	Not Recvd.	43	Non Reactive - 43		
HBsAg	Reactive	91	Non Reactive - 4 Reactive - 87	96	Within Consensus
HCV	Reactive	91	Non Reactive - 16 Reactive - 75	82	Within Consensus
HIV-1	Reactive	91	Non Reactive - 1 Reactive - 90	99	Within Consensus
RPR	Reactive	88	Non Reactive - 68 Reactive - 20		Out of Consensus
Rubella IgG	Not Recvd.	41	Reactive - 41		
Rubella IgM	Not Recvd.	39	Non Reactive - 39		
Toxoplasma IgG	Not Recvd.	40	Non Reactive - 1 Reactive - 39		
Toxoplasma IgM	Not Recvd.	39	Non Reactive - 39		
Weil Felix OX19	Not Recvd.	36	Negative - 35 Positive - 1		

Authorised Signatory

Dr. Sujay Prasad

Technical Manager and Program coordinator

Page 1 of 3

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560001
+91-9902277355 · qap@naalm.com · www.naalm.com
CIN: U85300KA2018PTC115147



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT

PT SCHEME: SEROLOGY



PC-1047

Cycle No	C3
Ref.No.	NEUQAP109
Sample ID	NEUQAP SERO SERUM/2021/7/C3/S3

Report Date : 10/08/2021

Sample : July 2021

Peer group

Parameters	Lab Report	No of Participants	Consensus Report	%	Remarks
HBsAg (Card Method)	Reactive	13	Non Reactive - 2	85	Within Consensus
			Reactive - 11		
HCV (Card method)	Reactive	15	Non Reactive - 10		Out of Consensus
			Reactive - 5		
HIV-1 (Card method)	Reactive	11	Non Reactive - 1	91	Within Consensus
			Reactive - 10		
RPR (--OTHERS--)	Reactive	2	Reactive - 2	100	Within Consensus
Widal (Slide agglutination)	Negative	20	Negative - 20	100	Within Consensus

Note: Sample generation, homogeneity and stability testing are subcontracted to accredited laboratories

**End of report **

Authorised Signatory

Dr. Sujay Prasad

Technical Manager and Program coordinator

This report is for use only by the intended participant.

Page 3 of 3



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560001
+91-9902277355 - qap@naalm.com - www.naalm.com
CIN: U85300KA2018PTC115147

RCA FORM OF ILC / PT / EQAS OUTLIERS

Department: SEROLOGY

Date: 31/08/2024

1. Proficiency test exception for: RPR / VDRL
2. Proficiency test provider: Non SAP
3. Proficiency test analyte group: PCR group
4. Cause for PT exception:

Random error

5. How did the section investigate the cause:

Satisfactory in
peer group - and
inter lab comparison with card
method was satisfactory

6. What was the status of the internal QC on the day PT initially analysed:

7. Category into which the cause will fit into:


- A. Method
- B. Technical
- C. Clerical
- D. Problem with PT material Other
- E. No explanation after investigation

8. Evidence that the problem was corrected successfully:


9. Specific corrections taken to prevent the recurrence if possible:

10. Signatures :

Quality manager:



Pathologist:



Troubleshooting Guidelines

Method	Technical	Clerical	Problem with PT material	No explanation after investigation
Equipment function checks	Misinterpretation / Wrong identification / Wrong labeling	Transcription error	Leaked / broken vial / not fit for analysis	Use this choice only when the investigation has yielded no satisfactory explanation
Scheduled maintenance not carried out or out of acceptable range	Dilution error / incorrect pipetting	Registration of wrong method or method change is not updated	Bacterial contamination	
Problem with data processing functions	Time delay between reconstitution and analysis		Perceived survey bias / inappropriate target value	
Faulty standard or other reagent	Calculation error			
Incorrect calibration	Analysis accepted in nonlinear range		Unstable material	
Carry over from previous specimen	Analysis done even though controls were out of range or controls not assayed		Matrix effect incompatible with method	
Result close to the detection limit of method	QC data within acceptable limits but showed trend suggestive of problem with the assay		No comparable peer group	
			Acceptable range too low	
	Sample mix up		Late shipment	
Other method related problem	Other technical problem		Improper package and temperature control	

Note: When all identifiable sources of error have been excluded, a single unacceptable result may be attributed to random error, particularly when the result of repeat analysis is acceptable. In such cases, no corrective action should be taken; as such an action may actually increase the probability of a future unacceptable result.

RCA FORM OF ILC / PT / EQAS OUTLIERS

Department: *SEROLOGY*Date: *31/08/2021*

1. Proficiency test exception for: *HCV*
2. Proficiency test provider: *NEU-QAP*
3. Proficiency test analyte group: *ALL METHODS*
4. Cause for PT exception:

5. How did the section investigate the cause:

-ELISA, CMIA and ELFA methods which are highly sensitive whereas Immochromatography card method can be known to give false negative results (AS from NEU-QAP)

6. What was the status of the internal QC on the day PT initially analysed:

Satisfactory in all methods and also inter lab comparison satisfactory

7. Category into which the cause will fit into:

- A. Method
- B. Technical
- C. Clerical
- D. Problem with PT material Other
- E. No explanation after investigation

8. Evidence that the problem was corrected successfully:

9. Specific corrections taken to prevent the recurrence if possible:

10. Signatures :

Quality manager: *[Signature]*Pathologist: *[Signature]*

Troubleshooting Guidelines

Method	Technical	Clerical	Problem with PT material	No explanation after investigation
Equipment function checks	Misinterpretation / Wrong identification / Wrong labeling	Transcription error	Leaked / broken vial / not fit for analysis	Use this choice only when the investigation has yielded no satisfactory explanation
Scheduled maintenance not carried out or out of acceptable range	Dilution error / incorrect pipetting	Registration of wrong method or method change is not updated	Bacterial contamination	
Problem with data processing functions	Time delay between reconstitution and analysis		Perceived survey bias / inappropriate target value	
Faulty standard or other reagent	Calculation error			
Incorrect calibration	Analysis accepted in nonlinear range		Unstable material	
Carry over from previous specimen	Analysis done even though controls were out of range or controls not assayed		Matrix effect incompatible with method	
Result close to the detection limit of method	QC data within acceptable limits but showed trend suggestive of problem with the assay		No comparable peer group	
			Acceptable range too low	
	Sample mix up		Late shipment	
Other method related problem	Other technical problem		Improper package and temperature control	

Note: When all identifiable sources of error have been excluded, a single unacceptable result may be attributed to random error, particularly when the result of repeat analysis is acceptable. In such cases, no corrective action should be taken; as such an action may actually increase the probability of a future unacceptable result.



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT

PT SCHEME: SEROLOGY



PC-1047

Cycle No	C3
Ref.No.	NEUQAP109
Sample ID	NEUQAP SERO SERUM/2021/4/C3/S2

Report Date : 23/05/2021

Sample : April 2021

All Methods

Parameters	Lab Report	No of Participants	Consensus Report	%	Remarks
Brucella	Not Recvd.	25	Non Reactive - 25		
Brucella Antigen	Not Recvd.	3	Non Reactive - 3		
Brucella IgG	Not Recvd.	14	Non Reactive - 14		
Brucella IgM	Not Recvd.	15	Non Reactive - 15		
Chikungunya	Not Recvd.	37	Negative - 37		
CMV IgG	Not Recvd.	37	Non Reactive - 1 Reactive - 36		
CMV IgM	Not Recvd.	38	Equivocal - 1 Non Reactive - 37		
HBsAg	Reactive	83	Reactive - 83	100	Within Consensus
HCV	Reactive	84	Non Reactive - 7 Reactive - 77	92	Within Consensus
HIV-1	Reactive	84	Non Reactive - 6 Reactive - 78	93	Within Consensus
RPR	Reactive	78	Non Reactive - 59 Reactive - 19		Out of Consensus
Rubella IgG	Not Recvd.	37	Non Reactive - 1 Reactive - 36		
Rubella IgM	Not Recvd.	35	Non Reactive - 34 Reactive - 1		
Toxoplasma IgG	Not Recvd.	34	Non Reactive - 1 Reactive - 33		
Toxoplasma IgM	Not Recvd.	34	Non Reactive - 34		

Authorised Signatory

Dr. Sujay Prasad
Technical Manager and Program coordinator

Kindly refer last page for a note on VDRL TEST

Page 1 of 4

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560001
+91-9902277355 · qap@naalm.com · www.naalm.com
CIN: U85300KA2018PTC115147



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT

PT SCHEME: SEROLOGY



PC-1047

Cycle No	C3
Ref.No.	NEUQAP109
Sample ID	NEUQAP SERO SERUM/2021/4/C3/S2

Report Date : 23/05/2021

Sample : April 2021

Peer group

Parameters	Lab Report	No of Participants	Consensus Report	%	Remarks
HBsAg (Card Method)	Reactive	12	Reactive - 12	100	Within Consensus
HCV (Card method)	Reactive	13	Non Reactive - 7 Reactive - 6		Out of Consensus
HIV-1 (Card method)	Reactive	11	Non Reactive - 5 Reactive - 6	55	Within Consensus
RPR (--OTHERS--)	Reactive	3	Non Reactive - 1 Reactive - 2	67	Within Consensus
Widal (Slide agglutination)	Positive	18	Negative - 17 Positive - 1		Out of Consensus

Note: Sample generation, homogeneity and stability testing are subcontracted to accredited laboratories

**End of report **

Authorised Signatory

Dr. Sujay Prasad

Technical Manager and Program coordinator

Kindly refer last page for a note on VDRL TEST

Page 3 of 4

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560001
+91-9902277355 · qap@naalm.com · www.naalm.com
CIN: U85300KA2018PTC115147




Name : Mr. INTERLAB
COMPARTION -2
Age / Sex : 23 Year(s) / Male
Patient ID : PDY052063
Visit No. : 21YP99164
Ref. Doctor : DR. SELF
Client : Yelahanka

Registered Date : 26/08/2021 02:46 PM
Collected On : 26/08/2021 03:20 PM
Received On : 26/08/2021 03:21 PM
Reported On : 26/08/2021 06:46 PM



Report Type : Final

Test	Results	Units	Biological Reference Range
HEMATOLOGY			
<u>CBC - Complete Blood Count</u>			
<u>CBC</u>			
Haemoglobin - Whole Blood Photometric (SLS Hemoglobin)	15.3	g/dL	13.5-18.0
Red blood cell count - Whole Blood Electrical impedance	4.97	million/cu.mm	4.2-6.5
Hematocrit(PCV) - Whole Blood Electrical impedance	46.1	%	39-54
Mean Corpuscular Volume(MCV) - Whole Blood Calculated from HCT and RBC	92.8	fL	75-95
Mean Corpuscular Hemoglobin(MCH) - Whole Blood Calculated from RBC and Hb	30.8	pg	26-32
Mean Corpuscular Hemoglobin Concentration (MCHC) - Whole Blood Calculated from HCT and Hb	33	g/dL	30-35
Red cell distribution width (RDW)-CV - Whole Blood Calculated	14.3	%	11.6-14.6
Platelet count - Whole Blood Electrical impedance	1.88	lakh/cu.mm	1.50-4.40
Total WBC count - Whole Blood	9400	cells/cu.mm	4000-11000
Absolute Basophil Count - Whole Blood Flow Cytometry	<u>0.10</u>	10 ³ / µl	0.01-0.05
Absolute Eosinophil Count - Whole Blood Flow Cytometry	<u>1.20</u>	10 ³ / µl	0.02-0.50
Absolute Lymphocyte Count - Whole Blood	2.40	10 ³ / µl	1-3


Dr. Shreya Prabhu,
Pathologist
KMC-No:114274

A UNIT OF BHIMA LIFESCIENCES LLP

4, "THERINA BLOCK", 17th A Cross, 3 rd Block, Jayanagar, Bangalore - 560 011

☎ 080 46382333 | 4638 2222 ✉ admin@primadiagnostics.com


Home collection : +91 78294 90000 + 91 78294 93333, Web : www.primadiagnostics.com



Name : Mr. INTERLAB
COMPARTION -2
Age / Sex : 23 Year(s) / Male
Patient ID : PDY052063
Visit No. : 21YP99164

Ref. Doctor : DR. SELF
Client : Yelahanka

Registered Date : 26/08/2021 02:46 PM
Collected On : 26/08/2021 03:20 PM
Received On : 26/08/2021 03:21 PM
Reported On : 26/08/2021 06:46 PM



Report Type : Final

Flow Cytometry

Absolute Neutrophil Count - Whole Blood 4.90 $10^3 / \mu\text{l}$ 2-7
Flow Cytometry
Absolute Monocyte Count - Whole Blood 0.70 $10^3 / \mu\text{l}$ 0.2-1.0
Flow Cytometry

DIFFERENTIAL COUNT - DC - Whole Blood
Flow Cytometry/Microscopy

Neutrophils 52.5 % 40-75
Lymphocytes 26.1 % 20-45
Monocytes 7.7 % 1-10
Eosinophils **13.0** % 1-6
Basophils 0.7 % 0-1

IMMUNOLOGY

T3 (TOTAL) - Serum 1.59 ng/mL
CLIA
Newborn : 0.7-2.0
< 1 Year: 1.0-2.4
1 - 5 Years : 1.0-2.4
6 - 10 Years: 0.9-2.4
11 - 50 Years: 0.7-2.0
> 50 Years: 0.4-1.8
First Trimester: 0.8-1.9
Second Trimester: 1.0-2.6

SEROLOGY

RAPID PLASMA REAGIN (RPR) / VDRL - Serum
Slide Flocculation

RAPID PLASMA REAGIN RPR / VDRL Non Reactive

Dr. Shreya Prabhu,
Pathologist
KMC No: 114274

A UNIT OF BHIMA LIFESCIENCES LLP

4, "THERINA BLOCK", 17th A Cross, 3 rd Block, Jayanagar, Bangalore - 560 011

☎ 080 46382333 | 4638 2222 ✉ admin@primadiagnostics.com

Home collection : +91 78294 90000 + 91 78294 93333, Web : www.primadiagnostics.com



Name : Mr. INTERLAB
COMPARTION -2
Age / Sex : 23 Year(s) / Male
Patient ID : PDY052063
Visit No. : 21YP99164

Ref. Doctor : DR. SELF
Client : Yelahanka

Registered Date : 26/08/2021 02:46 PM
Collected On : 26/08/2021 03:20 PM
Received On : 26/08/2021 03:21 PM
Reported On : 26/08/2021 06:46 PM



Report Type : Final

Test	Results	Units	Biological Reference Range
BIOCHEMISTRY			
Alanine Transaminase(ALT/ SGPT) - Serum IFCC without P-5-P	14.0	U/L	0-55
Alkaline Phosphatase - Serum IFCC PNPP	51.0	U/L	Male: 13 - 15 Year(s) : < 750 Both: 0 - 12 Year(s) : < 500 Both: 16 - 120 Year(s) : 40 - 150
Cholesterol Total - Serum CHOD-POD	158.00	mg/dl	Desirable: < 200 Borderline High: 200-239 High: > 240
Gamma-Glutamyl Transferase (GGT) - Serum Gamma--Glutamyl-p-nitroanilide IFCC	<u>9.0</u>	U/L	12-64
HDL Cholesterol - Serum Accelerator selective detergent	42.00	mg/dL	40-60
<u>ELECTROLYTES (Na+, K+, Cl)</u> - Serum ISE indirect			
SODIUM (Na+)	139	mmol/L	135-148
POTASSIUM (K+)	3.9	mmol/L	3.5-5.3
CHLORIDE(Cl-)	106	mmol/L	98-107

-- End of Report --

Kindly correlate clinically. If necessary discuss/repeat

Ramakrishna
Mr. Ramakrishna.B,
Biochemist



Name : Mr. INTERLAB
COMPARTION -1
Ref. Doctor : DR. SELF
Age / Sex : 22 Year(s) / Male
Client : Walk-in
Patient ID : PDY052062
/visit No. : 21YP99163

Registered Date : 26/08/2021 02:42 PM
Collected On : 26/08/2021 02:43 PM
Received On : 26/08/2021 02:43 PM
Reported On : 27/08/2021 12:08 PM



21YP99163

Report Type : Final

Test	Results	Units	Biological Reference Range
HEMATOLOGY			
<u>CBC - Complete Blood Count</u>			
<u>CBC</u>			
Haemoglobin - EDTA Whole Blood Spectrophotometry	15.8	g/dL	13.5-18
Red blood cell count - EDTA Whole Blood RBC Histogram(Volume)	4.8	million/cu.mm	Male: 4.2-6.5 Female: 3.7-5.6
Hematocrit(PCV) - EDTA Whole Blood RBC Pulse Height detection	44.7	%	Male: 39-54 Female: 34-48
Mean Corpuscular Volume(MCV) - EDTA Whole Blood Calculate From RBC Histogram	94.2	fL	75-95
Mean Corpuscular Hemoglobin(MCH) - EDTA Whole Blood Calculated From RBC and HGB	32.0	pg	26-32
Mean Corpuscular Hemoglobin Concentration (MCHC) - EDTA Whole Blood Calculated From HGB and HCT	35.0	g/dL	30-35
Red cell distribution width (RDW)-CV - EDTA Whole Blood Calculated	13.2	%	11.6-14.6
Platelet count - EDTA Whole Blood Platelet Histogram(Volume)	2.05	lakh/cu.mm	1.50-4.40
Total WBC count - EDTA Whole Blood Double Hydrodynamic Sequential System(DHSS)	8880	cells/cu.mm	4000-11000
Absolute Basophil Count - EDTA Whole Blood Flow Cytometry	0.02	10 ³ / µl	0.01-0.05
Absolute Eosinophil Count - EDTA Whole Blood	1.06	cells/cu.mm	0.01-0.50

Seema Umarji

Dr. Seema Umarji
MBBS MD, Pathologist.
KMC NO:95296

Page 1 / 4


PRIMA DIAGNOSTICS YELAHANKA

A UNIT OF GURURAJGOVIND LIFESCIENCE YELAHANKA LLP

Reg. Office : #93, 1st A Min Road, 1st Phase, A Sector, Yelahanka New Town, Bengaluru - 560064, Tel 080 - 4624 5555

(Head Office : 4/16, 9th Main Road, 3rd Block, Jnananagar, Bengaluru - 560011. Tel : 080 - 46382333)


www.primadiagnostics.com



Name : Mr. INTERLAB COMPARTION -1
 Age / Sex : 22 Year(s) / Male
 Patient ID : PDY052062
 Visit No. : 21YP99163

Ref. Doctor : DR. SELF
 Client : Walk-in

Registered Date : 26/08/2021 02:42 PM
 Collected On : 26/08/2021 02:43 PM
 Received On : 26/08/2021 02:43 PM
 Reported On : 27/08/2021 12:08 PM



Report Type : Final

Double Hydrodynamic Sequential System(DHSS)

Absolute Lymphocyte Count - EDTA Whole Blood 2.29 cells/cu.mm 1-3
 Blood
 Double Hydrodynamic Sequential System(DHSS)

Absolute Neutrophil Count - EDTA Whole Blood 4.90 $10^3 / \mu\text{l}$ 2-7
 Flow Cytometry

Absolute Monocyte Count - EDTA Whole Blood 0.61 $10^3 / \mu\text{l}$ 0.2-1.0
 Flow Cytometry

DIFFERENTIAL COUNT - DC

Neutrophils - EDTA Whole Blood 55 % 40-75
 DHSS / Microscopy

Lymphocytes - EDTA Whole Blood 26 % 20-45
 DHSS / Microscopy

Monocytes - EDTA Whole Blood 06 % 1-10
 Flow Cytometry/Microscopy

Eosinophils - EDTA Whole Blood 12 % 1-6
 DHSS / Microscopy

Basophils - EDTA Whole Blood 01 % 0-1
 DHSS / Microscopy

BIOCHEMISTRY

Alanine Transaminase(ALT/ SGPT) - Serum 15.0 U/L Adult: < 45
 UV with P-5-P Dry Chemistry

Alkaline Phosphatase - Serum 52.0 U/L 53-128
 Para Nitro Phenyl Phosphate,AMP Buffer Dry Chemistry

Cholesterol Total - Serum 153.0 mg/dl Desirable: < 200
 Cholesterol Oxidase Dry Chemistry BorderLine High: 200-239
 High: >= 240

Seema Umari

Dr. Seema Umari
 MBBS MD, Pathologist.
 KMC NO:95296

Page 2 / 4


PRIMA DIAGNOSTICS YELAHANKA

A UNIT OF GURURAJGOVIND LIFESCIENCE YELAHANKA LLP

Reg. Office : #93, 1st A Min Road, 1st Phase, A Sector, Yelahanka New Town, Bengaluru - 560064. Tel 080 - 4624 5555


(Head Office : 4/16, 9th Main Road, 3rd Block, Jayanagar, Bengaluru - 560011. Tel : 080 - 46382333)

www.primadiagnostics.com



Name : Mr. INTERLAB COMPARTION -1 Ref. Doctor : DR. SELF
 Age / Sex : 22 Year(s) / Male Client : Walk-in
 Patient ID : PDY052062
 Visit No. : 21YP99163

Registered Date : 26/08/2021 02:42 PM
 Collected On : 26/08/2021 02:43 PM
 Received On : 26/08/2021 02:43 PM
 Reported On : 27/08/2021 12:08 PM



Report Type : Final

Gamma-Glutamyl Transferase (GGT) - Serum 14.0 U/L < 55
 ?-glutamyl-p-nitroanilide Dry Chemistry
 HDL Cholesterol - Serum **43.0** mg/dL Low: < 40.0
 Colorimetric: non HDL Precipitation Method High: >= 60.0

ELECTROLYTES (Na+, K+, Cl) - Serum
 DIRECT ISE

SODIUM (Na+) 141.5 mmol/L 137-145
 POTASSIUM (K+) 3.81 mmol/L 3.5-5.1
 CHLORIDE(Cl-) 105.5 mmol/L 98-107

IMMUNOLOGY

T3 (TOTAL) - Serum 1.27 ng/mL
 Enhanced Chemiluminescence
 Newborn: 0.7-2.0
 < 1 Year: 1.0-2.4
 1 - 5 Years: 1.0-2.4
 6 - 10 Years: 0.9-2.4
 11 - 50 Years: 0.7-2.0
 > 50 Years: 0.4-1.8
 First Trimester: 0.8-1.9
 Second Trimester: 1.0-2.6

SEROLOGY

RAPID PLASMA REAGIN (RPR) / VDRL

RAPID PLASMA REAGIN RPR / VDRL - Serum, Non Reactive
 Slide Flocculation

INTERPRETATION - Serum Non Reactive

-- End of Report --

Seema Umarji

Dr. Seema Umarji
 MBBS MD, Pathologist.
 KMC NO:95296



PRIMA
DIAGNOSTICS
YELAHANKA

Name : Mr. INTERLAB
COMPARTION -1
Age / Sex : 22 Year(s) / Male
Ref. Doctor : DR. SELF
Client : Walk-in
Patient ID : PDY052062
Visit No. : 21YP99163

Registered Date : 26/08/2021 02:42 PM
Collected On : 26/08/2021 02:43 PM
Received On : 26/08/2021 02:43 PM
Reported On : 27/08/2021 12:08 PM



Report Type : Final

Kindly correlate clinically. If necessary discuss/repeat