




CALIBRATION CERTIFICATE

Page 1 of 1

1. Customer Name:	M/s. Prima Diagnostics, Yelahanka, C/o. Gururaj Govind Life Science LLP, No.93, 1st A Main Road, 1st Phase, Sec A, Yelahanka New Town, Bnagalore-560064	6. Certificate No:	RCPL/20-21/ML/S1107-01	
		7. ULR No:	CC221520000008234F	
		8. Date of Issue:	03-11-2020	
2. SRF NO:	S1107	9. Date of Calibration:	27-10-2020	
3. Date of SRF/Receipt:	27-10-2020	10. Next Cal Due:	26-10-2021	
4. Discipline:	Mechanical Calibration	11. Calibration SOP no:	RCPL/SOP/ML-05	
5. Calibrated At:	ONSITE	12. Condition of DUC on Receipt:	Good	
13. ENVIRONMENTAL CONDITION : Temperature: 25.8°C / Humidity: 56% R.H				
14. DUC DETAILS:				
Nomenclature :	Centrifuge	Serial. No:	ZEDN-15180	
Make / Model :	REMI & R-8C	Code/ID No :	PD-YH/LGI/CEN/01	
Range :	0 to 5250 RPM	Resolution:	10 RPM	
Location:	LABORATORY			
15. STANDARD INSTRUMENT USED:				
Nomenclature:	Digital Tachometer	Serial. No:	164-0109	
Make & Model :	Line Seiki & TM-5000	Cal Due Date:	23-09-2021	
Range :	6.0 to 99999.9 RPM	Traceable to:	Godrej & Boyce Mfg. Co. Ltd, CHENNAI	
Certificate No:	C-200915-4-1			
16. CALIBRATION RESULTS:				
Parameter: SPEED				
Sl. No.	Standard Reading (RPM)	DUC Set Point (RPM)	Error Observed (RPM)	Uncertainty of Measurement ± (% of reading)
1	999.8	1000.0	0.2	0.39
2	2001.1	2000.0	-1.1	0.20
3	3001.7	3000.0	-1.7	0.13
4	3999.2	4000.0	0.8	0.24
5	5001.3	5000.0	-1.3	0.19
17. REMARKS & CONCLUSION:				
<p>a. Reported Values of DUC & Standard are Average of 5 Measuring Series.</p> <p>b. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.0.</p> <p>c. Calibration certificate issued for Centrifuge is used scientific or industrial purpose only.</p> <p>d. The Std. used for calibration were calibrated and Traceable to National/International Standards.</p>				
Calibrated By  NAVEENKUMAR B CALIBRATION ENGINEER		Authorized Signatory  ARUNKUMAR QUALITY MANAGER		
				
FM-GL-85				



CALIBRATION CERTIFICATE

CC-2559

Godrej & Boyce Mfg. Co. Ltd.,
Calibration Services, No.1, SIDCO
Industrial Estate,
Ambattur., Chennai
Tamilnadu. 600098. India
Dir.:91-44-66544651/52/53

Format No.	:FR-QT-21	Page No.	: 1 of 2	
Certificate No.	:C-200915-4-1	Issue Date	:29/09/2020	
Date Of Calibration	:17/09/2020	Recom. Due Date	:17/09/2021	
Customer Details :		Calibrated at	: Lab (-)	
Reltec Calibration Pvt Ltd		Challan Date.	: 15/09/2020	
98/50, 1st Floor, 2nd Main, Industrial Town, Rajajinagar, Bangalore karnataka India 560044		Date Of Receipt	: 15/09/2020	
		Cond. On Receipt	: Satisfactory	
		Receipt No.	: C-200915-4	
Details of Unit Under Calibration :				
Name	: DIGITAL TACHOMETER	Range	: 6 to 99999 RPM	
Make	: LINE	Least Count	: 0.1 RPM	
Sr. No.	: 164-0109	Accuracy(±)	: AS PER MANUAL	
Id. No.	: RCPL/TA/65	Model No.	: TM-5000	
Operating Range	: As Per Report			
Details Of Master Equipments				
Name	Sr. No.	IDNO	Valid Upto	Traceability
DIGITAL TACHOMETER	A.E.05259	QSC/E/31	05/11/2020	Hitech ,Chennai
DIGITAL TACHOMETER	1954829	QSC/M/75	14/11/2021	FCRI, Kerala
Work Instruction No.: WI-ME-05 & WI-ME-01		Reference Standard No.: SANAS 45-02		
Environmental Details:				
Temperature: (20 ± 1) °C		Relative Humidity: (60 ± 20) %RH		
Remark				
1. The Standard reading given is the average of 3 Measurements				
Notes				
<ul style="list-style-type: none"> •Equipment used for calibration were calibrated & traceable to National & International Standards. •This certificate refers only to particular item(s) submitted for calibration. •Godrej Calibration Services is not liable for any change in calibration data & performance specification on account of malfunctioning of Standards/instruments/Equipment covered by this certificate due to damage caused to it after issuance of this certificate. •The calibration results reported are valid at the time of and under the stated conditions of the measurements. •This certificate shall not be reproduced in full/part without prior permission of Godrej Calibration Services. •Satisfactory calibration report in no ways implies that the equipment calibrated is approved by NABL. •All precautions have been taken for any error or omission while calibrating the instruments and issuing this certificate. However Godrej Calibration Services shall not be liable for any loss or liability that may be arise to any party in this regards. 				
Authorized Signatory :P Muthukumar Designation : Senior Manager Calibrated By: M Rajesh Calibration Engg ULR No. : (CC255920000016657F)				Signature valid Digitally signed by DS GODREJ AND BOYCE MANUFACTURING CO LIMITED 1 Date: 2020.10.08 18:00:26 +05:30 Reason: Calibration Certificate Location: Chennai



Godrej & Boyce Mfg. Co. Ltd.