

Test Record ID:	21153010118/ 70040487/ Lab		TEST STANDARD	Biochemistry Analyzer							
	TEST NUMBER	TEST	READING VALUE	READING UNIT	FILTER SELECTION	EO POLARITY	EO NEUTRAL	EO EARTH	APPLIED PARTS SELECTION		
			0.303	Ohms	N/A	N/A	N/A	N/A	N/A		
1	Protective Earth Resistance		6.2	uA	AC+DC	Off	N/A	N/A	//OPEN		
2	Leakage: Earth		65.1	uA	AC+DC	Normal	Open	N/A	//OPEN		
3	Leakage: Earth		24.6	uA	AC+DC	Normal	Closed	N/A	//OPEN		
4	Leakage: Earth		65.1	uA	AC+DC	Reversed	Open	N/A	//OPEN		
5	Leakage: Earth		24.7	uA	AC+DC	Reversed	Closed	N/A	//OPEN		
6	Leakage: Earth		6	uA	AC+DC	Off	N/A	Open	//OPEN		
7	Leakage: Enclosure		0.2	uA	AC+DC	Off	N/A	Closed	//OPEN		
8	Leakage: Enclosure		64.7	uA	AC+DC	Normal	Open	Open	//OPEN		
9	Leakage: Enclosure		0.3	uA	AC+DC	Normal	Closed	Closed	//OPEN		
10	Leakage: Enclosure		0.2	uA	AC+DC	Normal	Open	Closed	//OPEN		
11	Leakage: Enclosure		24.7	uA	AC+DC	Normal	Closed	Closed	//OPEN		
12	Leakage: Enclosure		0.1	A	N/A	Normal	N/A	Open	//OPEN		
13	Equipment Current		220.8	V	N/A	N/A	N/A	N/A	N/A		
14	Mains Voltage: Live to Neutral		7.9	V	N/A	N/A	N/A	N/A	N/A		
15	Mains Voltage: Neutral to Earth		8.2	V	N/A	N/A	N/A	N/A	N/A		
16	Mains Voltage: Live to Earth				N/A	N/A	N/A	N/A	N/A		

Calibration Passed	Yes	No
	✓	

