



FACTUAL LEGACY
SINCE 1964

**YORK
SCIENTIFIC INDUSTRIES
PRIVATE LIMITED**

8/6E, Site IV, Sahibabad Industrial Area,
Ghaziabad-201010, U.P.

0120-4741800 (30 Lines)

0120-4741811

ysipl@yorco.co.in, sales@yorco.co.in

CALIBRATION CERTIFICATE

Certificate No. : YSI / QC /21I\2K753
Date of Calibration : 24.08.2021
Item : YORCO SEROLOGICAL WATER BATH

Temp. Range : Ambient to 80°C
Make : YORCO
Model : YSI-413
S. No. : 21H4849
Size : 300 x 250 x 175 mm
Valid upto : 23.08.2022

Standard used : Digital Temperature Indicator
Make : Thermotech Model : L-2001
Sr. No. : 1374114 Certificate No.021702
Certificate by : Loctron Instruments Delhi

Test Observations

S. No.	U.U.C.*READING	ACTUAL READING ON STANDARD
1	30°C	30.3°C
2	40°C	40.2°C
3	50°C	50.3°C
4	60°C	60.3°C

U.U.C.*(UNIT UNDER CALIBRATION)

Conclusion: YORCO Water Bath test results are Satisfactory and within limits.



York Scientific Industries Pvt. Ltd

(Q. A. Engineer)

Loctron INSTRUMENTS PVT.LTD.



REGD. OFFICE: 217, ALLIED HOUSE, OLD ROHTAK ROAD, INDERLOK, DELHI -110035.
Ph. 23658589,55159087 M: 9810041363, Fax 011-23658336 E-mail: thermotechindia@rediffmail.com
thermotechdelhi@gmail.com

CALIBRATION CERTIFICATE

An ISO 9001:2000 Certified Co.

Calibrated For : - York Scientific Industries Pvt Ltd.	Certificate Number :- 021702
	Calibrated on 17/02/2021
	Calibration valid upto: 16/02/2022

Description and Identification of Instrument

Instrument	Digital Temp. Indicator	Model	E-2001
Range	0.0°C to 199.9°C	TYPE	-----
Operating range	0.0°C to 150.0°C	S. No.	1374114
Environmental Conditions	Room Temp. = 25 ± 2°C Humidity = 35 ± 10% RH	Sensor	PT-100

Standard Equipment Used

S.No.	Nomenclature	Make	S.No./ID Mark	Calibrated By	Certificate No.	Due date of Cal.
1	Universal Cali-brator	Victor 01+ ³	093381970	Indian Calibration Services (NABL Accredited)	21000001101	01-02-2022


CALIBRATION RESULTS

Sr.No.	Nominal Value in °C	Actual Display Value in °C
1	0.0	0.0
2	50.0	50.0
3	100.0	100.2
4	150.0	150.3

The above calibration has been performed at our works

Calibrated by :

Approved by:



Conditions:

1. The certificate refers only to the particular item(s) submitted for calibration.
2. The calibration results reported in this certificate are valid at the time of and under the stated condition of measurement.
3. Unit is accepted as accuracy in within the specified limit of claimed accuracy and certificate is valid upto one year from the date of issue

COMMITTED FOR THE EXCELLENCE IN INSTRUMENTATION

Thermotech®