

**Corrective action**

April 2021 External Quality assurance scheme found abnormal in Urea in previous kit and replaced by New Kit

  
Lab Incharge



AUTHORISED SIGNATORY

Rec.No.