


**Corrective action**

March 2021 External Quality assurance scheme found abnormal in T.Bilirubin, Albumin, Cholesterol and Triglycerides in previous kit and replaced by New Kit and Fully auto Biochemistry Instrument.

  
Lab Incharge



Virudhai  
Medical Centre

AUTHORISED SIGNATORY

Rec.No.

PARTIAL REPRODUCTION OF THIS REPORT IS NOT PERMITTED  
Note: This report is subject to the terms and conditions overleaf

மு் வீருதை மெடிக்கல் சென்டர்,  
தென்கோட்டை வீதி,  
விருத்தாசலம்-606 001.