

**REMI GROUP**

ESTD. 1960

**REMI SALES & ENGINEERING LIMITED**

B- 25, 2nd Floor, B Block Community  
Center, Janakpuri, New Delhi - 110 058  
Phone : 011 - 28502666  
Fax : 011 - 25573170  
E-mail : delhi@remigroup.com

**CALIBRATION**

CUSTOMER / BRANCH: M/s. Blood Bank , Fatima Hospital, Gorakhpur

DISCRIPTION OF MODEL: PLASMA FREEZER

MODEL: RPF-224 SR.NO: ZDBP-04788

CERTIFICATE NO: REMI/0214/2021-22

CALIBRATION DATE: 08.08.2021

CALIBRATION VALIDITY: 07.02.2022

**DETAILS OF EQUIPMENT USED FOR CALIBRATION**

Description	Certified by	Certificate No.	Calibration Date	Validity
Digital Temperature Indicator	Classic Instrumentation Pvt. Ltd.	CI/2K20-21/11/289	24/11/2020	23/11/2021

**OBSERVATIONS**

Sr. No.	Instrument Temp. Displayed (°C)	Indicator Temp. (°C)	Temperature Deviation	Calibration Date	Remark
1.	-40°C	-40.4°C	0.4°C	08.08.2021	O.K

- Remark:**
1. Calibration Traceable to national standards.
  2. Overall Deviation is within specified accuracy.
  3. Result are related only to the item calibrated.
  4. Maximum tolerance permissible is  $\pm 1^\circ\text{C}$  &  $\pm 5^\circ\text{C}$  whichever is applicable & RPM 1.5% - 2% FSD. &  $\pm 5\text{g}$ . to set weight.

Calibrated by: Mr. Vineet Kumar

**SALES & SERVICE BRANCHES :**

Ahmedabad, Bengaluru, Kolkata, Chennai, Hyderabad, Nagpur, New Delhi, Kochi, Kanpur, Indore, Goa

# Fatima Hospital

Mother Teresa Road, Padri Bazar P.O., Gorakhpur - 273 014  
 Ph: (0551) - 2281862, 2282602, E-mail - fatimagnkp@gmail.com  
 Visit us: www.fatimahospitalgkp.com



## Refrigerator Temperature Monitoring Check List (Freezer)

Month September

Year 2021

Department Lab

Date	Morning		Evening		Night	
	Temperature level	Checked by staff name and sig.	Temperature level	Checked by staff name and sig.	Temperature level	Checked by staff name and sig.
01	-5°C	Vicky	-4.9°C	Ashu	-6°C	Ashu
02	-4°C	Vicky	-6.1	Vicky	-6.5°C	Ashu
03	-3°C	Vicky	-5.1	Ashu	-6.1°C	Ashu
04	-5.1°C	Vicky	-4.9	Ashu	-7.1°C	Ashu
05	-4.9°C	Vicky	-4.3	Ashu	-7.9	Ashu
06	-3.5°C	Vicky	-4.6	Ashu	-8.1	Ashu
07	-4.1°C	Vicky	-3.7	Ashu	-8.1	Ashu
08	-4.2°C	Vicky	-3.5	Ashu	-6.9°C	CW
09	-3.1	Vicky	-4.1	Ashu	-5.9°C	CW
10	-4.1	Ashu	-4.0	Ashu	-5.1°C	CW
11	-3.9	Ashu	-5.1°C	Disha	-6.2°C	CW
12	-4.7	Ashu	-4.9°C	Disha	-5.3°C	CW
13	-4.3	Ashu	-4.8°C	Disha	-5.1°C	CW
14	-4.3	Ashu	-3.5°C	Disha	-4.9°C	CW
15	-4.2	Ashu	-4.2°C	Disha	-5.5°C	Ashu
16	-4.0	Ashu	-4.0	Disha	-5.1°C	Ashu
17	-3.9	Ashu	-4.7	Disha	+8	
18						
19						
20						
21						
22						
23						
24						
25						
26						
27						

CUSTOMER / BRANCH: M/s. Fatima Laboratory , Fatima Hospital, Gorakhpur

DISCRIPTION OF MODEL: FREEZER

MODEL: RPF-224 SR.NO: ZDBP-04788

CERTIFICATE NO: REMI/0182/2020-21

CALIBRATION DATE: 08/02/2021

CALIBRATION VALIDITY: 07/08/2021

**DETAILS OF EQUIPMENT USED FOR CALIBRATION**

Description	Certified by	Certificate No.	Calibration Date	Validity
Digital Temperature Indicator	Classic Instrumentation Pvt. Ltd.	CI/2K19-2K20/05/TH005	09/07/2020	08/07/2021

**OBSERVATIONS**

Sr. No.	Instrument Temp. Displayed (°C)	Indicator Temp. (°C)	Temperature Deviation	Calibration Date	Remark
1.	-40°C	-40.4°C	0.4°C	08/02/2021	O.K

- Remark:
1. Calibration Traceable to national standards.
  2. Overall Deviation is within specified accuracy.
  3. Result are related only to the item calibrated.
  4. Maximum tolerance permissible is  $\pm 1^{\circ}\text{C}$  &  $\pm 5^{\circ}\text{C}$  whichever is applicable & RPM 1.5% - 2% FSD. &  $\pm 5\text{g.}$  to set weight.

Calibrated by: Mr. Vineet Kumar

**SALES & SERVICE BRANCHES :**

Ahmedabad, Bengaluru, Kolkata, Chennai, Hyderabad, Nagpur, New Delhi, Kochi, Kanpur, Indore, Goa