




Calibration Certificate

Customer Name	Dr. SR Pathology Lab
Address	Nawashehar (Punjab)
Contact Person	Dr Rahul Goyal
Contact No.	+91 8146630668
Instrument	Haematology Analyzer
Model	BC-5000
Serial No.	SS-0B007535
Instrument Status	Under Warranty
Calibration Done On	22.09.2021
Next Calibration Due On	21.09.2022

For Aspen Diagnostic Pvt. Ltd.


Nitish Mahajan
Service Engineer




Rajaram Krishnamurthy
National Service Manager



Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



Certificate No.: 30903/A/0001/UK/En
ISO 9001:2015
CERTIFIED COMPANY

Calibrator Calibration

Instrument: Haematology Analyzer		Model: BC -5000		Serial No.: SS-0B007535		Date of Cal. 22.09.2021	
Mode: OV-WB		Calibrator Lot No. PLUS0921			Exp.Date:05.10.2021		
Parameter	WBC	RBC	HGB	MCV	PLT		
TARGET	9.05	4.54	13.5	89.4	252		
1	9.00	4.51	13.4	89.8	234		
2	8.93	4.55	13.4	89.1	234		
3	8.97	4.56	13.4	89.6	226		
4	9.02	4.55	13.4	89.1	237		
5	8.91	4.51	13.3	89.5	231		
6	9.24	4.52	13.4	89.3	236		
7	9.12	4.58	13.5	89.4	236		
8	9.20	4.54	13.4	89.7	235		
9	9.11	4.54	13.5	88.9	222		
10	8.89	4.56	13.4	89.1	232		
Mean	9.04	4.54	13.4	89.4	232		
CV(%)	1.4	0.5	0.3	0.3	2.1		
New Cal. Factor (%)	100.12	99.97	100.68	100.05	108.53		
Old Cal. Factors (%)	100.00	97.00	95.00	100.00	100.00		

For Aspen Diagnostic Pvt. Ltd.

Nitish Mahajan

Nitish Mahajan
Service Engineer



Rajaram Krishnamurthy

Rajaram Krishnamurthy
National Service Manager