



# APEX QUALITY CERTIFICATION SERVICES PVT. LTD.

Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019  
Ph.: 0141-4036776, 9414079483, 9636293333  
Email: apex.qcs@gmail.com



NABL ACCREDITED LAB  
Cert. No.: CC-2298

## CALIBRATION CERTIFICATE

ULR: CC229821000001533F	DISCIPLINE	Mechanical	Page 01 of 01
<b>CUSTOMER'S DETAILS</b>	<b>CERTIFICATE NO.:</b>	AQCS/21061401.001	
M/S DR. B LAL CLINICAL LABORATORY PVT. LTD. (ALW)	Date of Issue of Certificate	17/06/2021	
House No. 12, Near Ambedkar Circle, Lajpat Nagar, Alwar-301001 (Rajasthan)	Date of Calibration	14/06/2021	
	Recommended due date	13/06/2022	

Service Request No.	210612/02	<b>INSTRUMENT DESCRIPTION</b>	
Service Request Date	12/06/2021	Nomenclature	Centrifuge Machine
DUC Receipt Date	14/06/2021	Make	Remi
DUC Condition	Ok	Model	R-8C BL
Cal. Performed At	Site	Sr. No. / Ident. No.	ZCGN-20563 / BLAL/AL/CFG/01
Procedure Followed	CP/RM/01	Range	0-6000 RPM
<b>Environmental Condition</b>		Least Count	10 RPM
Temperature	30±10°C	Location	--
Rel. Hum.	50±10%		

### STANDARD EQUIPMENT USED:

Nomenclature	Range	Uncertainty (±)	Identification No.	Traceability Cert No.	Valid up to
Digital Tachometer	10 to 99,999	5 to 26	S999205	NCL/B/2021/34/01	25-02-2022


### CALIBRATION RESULT

Measured Value of RPM on Std.	Measured Value of RPM on DUC	Error in RPM	Expanded Uncertainty (± % of rdg.) in RPM
710.3	710	-0.3	2.37
1251	1250	-1	2.37
2551	2550	-1	2.37
3582	3580	-2	1.04
6002	6000	-2	1.04

### REMARKS :

- Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor k = 2 is as given above.
- The result reported in this certificate is valid at the time of and under the stated conditions of Measurement for the particular item and instrument identified above.
- This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur
- Format Number.: F-06/00

Calibrated by  
  
(V. B. Mathew)



Approved by  
  
(R. R. Yadav)  
Quality Manager



# APEX QUALITY CERTIFICATION SERVICES PVT. LTD.

Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019  
Ph.: 0141-4036776, 9414079483, 9636293333  
Email: apex.qcs@gmail.com



NABL ACCREDITED LAB  
Cert. No.: CC-2298

## CALIBRATION CERTIFICATE

ULR: CC229821000001534F	DISCIPLINE Mechanical	Page 01 of 01
<b>CUSTOMER'S DETAILS</b> M/S DR. B LAL CLINICAL LABORATORY PVT. LTD. (ALW) House No. 12, Near Ambedkar Circle, Lajpat Nagar, Alwar-301001 (Rajasthan)	<b>CERTIFICATE NO.:</b> Date of Issue of Certificate Date of Calibration Recommended due date	AQCS/21061401.002 17/06/2021 14/06/2021 13/06/2022

Service Request No.	210612/02	<b>INSTRUMENT DESCRIPTION</b>	
Service Request Date	12/06/2021	Nomenclature	Centrifuge Machine
DUC Receipt Date	14/06/2021	Make	Remi
DUC Condition	Ok	Model	R-8C PLUS
Cal. Performed At	Site	Sr. No. / Ident. No.	ZEHN-26433 / BLAL/AL/CFG/02
Procedure Followed	CP/RM/01	Range	0-6000 RPM
<b>Environmental Condition</b>		Least Count	10 RPM
Temperature	30±10°C	Rel. Hum.	50±10%
		Location	--

### STANDARD EQUIPMENT USED:

Nomenclature	Range	Uncertainty (±)	Identification No.	Traceability Cert No.	Valid up to
Digital Tachometer	10 to 99,999	5 to 26	S999205	NCL/B/2021/34/01	25-02-2022

### CALIBRATION RESULT


Measured Value of RPM on Std.	Measured Value of RPM on DUC	Error in RPM	Expanded Uncertainty (± % of rdg.) in RPM
799.4	800	0.6	2.37
1499	1500	1	2.37
2598	2600	2	2.37
3998	4000	2	1.04
5987	5990	3	1.04

### REMARKS :

- Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor  $k = 2$  is as given above.
- The result reported in this certificate is valid at the time of and under the stated conditions of Measurement for the particular item and instrument identified above.
- This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur
- Format Number.: F-06/00

Calibrated by

(V. B. Mathew)



Approved by

(R. R. Yadav)  
Quality Manager



# APEX QUALITY CERTIFICATION SERVICES PVT. LTD.



Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019  
Ph.: 0141-4036776, 9414079483, 9636293333  
Email: apex.qcs@gmail.com

## CALIBRATION CERTIFICATE

Page 01 of 01

CERTIFICATE NO.: AQCS/21061401.10015		Date of Calibration : 14/06/2021	Recommended due date : 13/06/2022																		
<b>CUSTOMER'S DETAILS</b>		<b>INSTRUMENT DESCRIPTION</b>																			
M/s DR. B LAL CLINICAL LABORATORY PVT. LTD. (ALW) House No. 12, Near Ambedkar Circle, Lajpat Nagar, Alwar - 301001 (Rajasthan)		Nomenclature : Centrifuge Machine Make / Model : Remi /R-303 Sr. No./Ident. No. : ZEFN-20008 /BLAL/AL/CFG/03 Range : 0-3800 RPM Least Count : NA Location : No Cal. Performed At : Site																			
Service Request No. : 2106112/02 Dated: 12/06/2021																					
<b>STANDARD EQUIPMENT USED:</b>																					
Nomenclature	Identification No.	Traceability Certificate No.	Valid up to																		
Digital Tachometer	S999205	NCL/B/2021/34/01	25-02-2022																		
<b>ENVIRONMENTAL CONDITIONS:</b>		<b>PROCEDURE FOLLOWED:</b>																			
Temperature : 23 ± 5°C Relative Humidity : 50± 10%		CP/RM-01																			
<b>CALIBRATION RESULT</b>																					
<table border="1"><thead><tr><th>S. No.</th><th>Set Knob on DUC</th><th>Measured Value on Standard RPM</th></tr></thead><tbody><tr><td>1</td><td>1</td><td>1120</td></tr><tr><td>2</td><td>2</td><td>1920</td></tr><tr><td>3</td><td>3</td><td>2690</td></tr><tr><td>4</td><td>4</td><td>3170</td></tr><tr><td>5</td><td>5</td><td>3720</td></tr></tbody></table>				S. No.	Set Knob on DUC	Measured Value on Standard RPM	1	1	1120	2	2	1920	3	3	2690	4	4	3170	5	5	3720
S. No.	Set Knob on DUC	Measured Value on Standard RPM																			
1	1	1120																			
2	2	1920																			
3	3	2690																			
4	4	3170																			
5	5	3720																			
<b>REMARKS :</b>																					
<ol style="list-style-type: none"><li>Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor k = 2 is ± 2.37% of reading.</li><li>The result reported in this certificate is valid at the time of and under the started conditions of Measurement for the particular item and instrument identified above.</li><li>This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur</li><li>Format Number.: F-06/00</li></ol>																					
Calibrated by  (V. B. Mathew)			Approved by  (R. R. Yadav) Quality Manager																		