

VIEW LAB MONTHLY SUMMARY

Lab Name **NEW GREEN LAB AND BIOPSY CENTRE**[Details About Robust Analysis](#)Lab No **13661**[Detail About Monthly Summary](#)Month **March**[Detail about SDI](#)Year **2021**Constituent Group **Chemistry II**[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)Date of Result Entered : **19/03/2021**Date of Report Published : **01/04/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Transasia / Erba	791	160.91	10.47	16.84	146.2 mg/dl	-0.87	1.20
2	UREA II	Urease UV / GLDH II	Transasia / Erba	605	42.26	12.78	5.40	38.3 mg/dl	-0.73	0.44
3	CREATININE II	Enzymatic Colorimetric II	Transasia / Erba	146	1.66	8.62	0.14	1.56 mg/dl	-0.70	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	658	3.60	13.31	0.48	4.15 mg/dl	1.15	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	595	5.89	9.22	0.54	6 g/dl	0.20	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	599	3.52	9.28	0.33	3.2 g/dl	-0.98	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	668	7.46	13.58	1.01	7.2 mg/dl	-0.26	0.08
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	717	118.97	12.88	15.32	109.2 mg/dl	-0.64	1.14
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	665	105.23	19.47	20.49	88.1 mg/dl	-0.84	1.59

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out