



Lab Name **CENTRE**
 Month **April**
 Constituent Group **Chemistry II**

Lab No **1000**
 Year **2021**



Date of Result Entered : 20/04/2021

Date of Report Published : 01/05/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	713	318.19	10.10	32.15	304.9 mg/dl	-0.41	2.41
2	UREA II	Urease UV / GLDH II	Transasia / Erba	522	27.05	13.69	3.70	24.2 mg/dl	-0.77	0.32
3	CREATININE II	Enzymatic Colorimetric II	Transasia / Erba	142	7.19	8.81	0.63	6.23 mg/dl	-1.52	0.11
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	676	3.28	15.32	0.50	2.9 mg/dl	-0.76	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	537	5.61	8.16	0.46	6.5 g/dl	1.94	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	550	3.32	9.16	0.30	4 g/dl	2.24	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	601	4.45	17.19	0.76	3.9 mg/dl	-0.72	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	677	113.01	11.66	13.18	99 mg/dl	-1.06	1.01
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	629	214.19	12.53	26.84	188.7 mg/dl	-0.95	2.14

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****