



Lab Name **NEW GREEN LAB AND BIOPSY CENTRE** Lab No **13661**
 Month **May** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 19/05/2021

Date of Report Published : 05/06/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	670	132.95	10.22	13.59	136.5 mg/dl	0.26	1.05
2	UREA II	Urease UV / GLDH II	Transasia / Erba	562	63.66	12.61	8.03	64.4 mg/dl	0.09	0.68
3	CREATININE II	Enzymatic Colorimetric II	Transasia / Erba	146	1.84	9.05	0.17	1.83 mg/dl	-0.06	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	634	2.47	17.92	0.44	2.46 mg/dl	-0.02	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	588	5.74	9.09	0.52	5.91 g/dl	0.33	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	567	3.39	9.40	0.32	3.34 g/dl	-0.16	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	636	7.08	13.17	0.93	7.4 mg/dl	0.34	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	671	109.08	12.80	13.96	100.1 mg/dl	-0.64	1.08
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	615	208.59	12.05	25.13	201.6 mg/dl	-0.28	2.03

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

ALL PARAMETERS
LENT

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****