

**CALIBRATION REPORT**

ADL/CO/SER/857/20-21

Customer: Green Life Diagnostic Care, Ramanathapuram

Serial No. : 212030355476

Model : Mispa Viva

DOI: 12/09/2020

Date of Calibration: 12/09/2020

INSTRUMENT CONTROL

Visual Control: PASSED

Zero Drift after 1 Hour of zero setting (WL=340nm) 0.001

PHOTOMETRIC LINEARITY

Solution	WL (nm)	100%	75%	50%	25%
Potassium dichromate	340nm	100.2	74.6	51.8	25.2
P-Nitrophenol	405nm	100.4	74.5	50.3	25.8
Cobaltous Ammonium Sulphate	505nm	100.8	76.9	50.1	24.7

CHEMISTRY CONTROL

	TEST	REAGENT	CONTROL SERUM	THEORETICAL VALUE	READ VALUE	RANGE
END POINT			Lot No: Expiry :			
KINETIC			Lot No: Expiry :			
FIXED TIME			Lot No: Expiry :			

Validity: 6 months from the date of calibration.

For, Agappe Diagnostics Ltd,

Authorized Signatory

Senthilkumar R

Engineer-Customer Service

**AGAPPE DIAGNOSTICS LTD.**

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellai, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkatoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in