



## Calibration Certificate

Certificate No.

STC/CAL/00115

User Name :

M/s Dr. SR Pathology lab  
Chandigarh Road, Opp. K sera Miniplex  
Near ITI, Nawanshahr, Pb- 144514

Page No.: 1 of 1

Format No. STC/7.8F-01

|                           |   |
|---------------------------|---|
| Equip. For Calibration    | Digital Temperature Sensor With Indicator |
| Make/Model No.            | N/A                                       |
| Range                     | -50 to 80°C                               |
| Least Count               | 0.1°C                                     |
| Accuracy                  | N/A                                       |
| ID No./Serial No.         | DTIS-01                                   |
| SRF No. & Date of Receipt | 2021/07/022-14/07/2021                    |
| Visual Inspection         | Satisfactory                              |
| Date of Calibration       | 15/07/2021                                |
| Suggested Due Date        | 14/07/2022                                |
| Date of Issue             | 17/07/2021                                |
| Calibrated at             | Site                                      |
| Location                  | ---                                       |
| Environmental Condition   |   |
| Temperature               | 25 ± 15°C                                 |
| Humidity                  | 50 ± 20%RH                                |

| Equipment & Master Used For Calibration |                           |
|---|---------------------------|
| 01.                                     | 4 Wire RTD with Indicator |
| Make                                    | Fluke/Tempsens            |
| Serial No.                              | 3812025/26097, STC/RTD/01 |
| Certificate No.                         | TL/021/120.2.1            |
| Next Due Date                           | 01/02/2022                |
| Traceability                            | Tempsens, Raj.            |

Discipline: Thermal Calibration

### Calibration Results

| Sr. No. | Unit Under Calibration in °C | Standard Reading in °C | Expanded Uncertainty |
|---------|------------------------------|------------------------|----------------------|
| 1       | -25.1                        | -25.04                 | ±0.31°C              |
| 2       | 40.2                         | 40.35                  |                      |
| 3       | 60.3                         | 60.51                  | ±0.22 °C             |
| 4       | 80.4                         | 80.65                  |                      |

Reference Document For Calibration : STC/SOP/T-05, IS: DKD-R 5-1

Uncertainty of Measurement : The Reported expended uncertainty of measurement is stated as the standard uncertainty of mesurment multiplied by the coverage factor k=2 such the coverage probability corresponds to approximately 95% confidence level.

Note:

1. The calibration results reported in this certificate are valid at the time of and under the stated condition of measurement.
2. This Report should not be reproduced except in full without our prior permission in writing.
3. Calibration Certificate without signature are not valid.

Calibrated By  
Vivek Safi  
Calibration Engineer



Authorized Signatory  
Krishan Pal  
(Branch Head)



## Calibration Certificate

Page No.: 1 of 1

Certificate No. **STC/CAL/00109**  
 User Name: M/s Dr. SR Pathology lab  
 Chandigarh Road, Opp. K sera Miniplex  
 Near ITI, Nawanshahr, Pb- 144514

Format No. **STC/7.8F-01**

|                                |                           |
|--------------------------------|---------------------------|
| Equip. For Calibration         | Digital Thermo Hygrometer |
| Make/Model No.                 | HTC-1                     |
| Range                          | -10 to 50 °C/10 to 99% RH |
| Least Count                    | 0.1 °C/1% RH              |
| Accuracy                       | N/A                       |
| Id No./Serial No.              | DTHM-01                   |
| SRF No. & Date of Receipt      | 2021/07/022-14/07/2021    |
| Visual Inspection              | Satisfactory              |
| Date of Calibration            | 15/07/2021                |
| Due Date                       | 14/07/2022                |
| Date of Issue                  | 17/07/2021                |
| Calibrated at                  | Lab                       |
| Location                       | --                        |
| <b>Environmental Condition</b> |                           |
| Temperature                    | 25 ± 4°C                  |
| Humidity                       | 30 to 75% RH              |

| Equipment & Master Used For Calibration |                                  |
|---|----------------------------------|
| 01.                                     | : 4 Wire RTD with Indicator      |
| Make                                    | : Fluke/Tempsens                 |
| Serial No./Id No.                       | : 3812025/26097, STC/RTD/01      |
| Certificate No.                         | : TL/021/120.2.1                 |
| Next Due Date                           | : 01/02/2022                     |
| Traceability                            | : Tempsens, Raj.                 |
| 02.                                     | : Humidity Indicator With Sensor |
| Make                                    | : HPG Systems/G-1105             |
| Serial No./Id No.                       | : 61319530, STC/DRHIS/01         |
| Certificate No.                         | : C&IJ/CAL/21-01/133             |
| Next Due Date                           | : 30/01/2022                     |
| Traceability                            | : C and I, Jaipur                |

Discipline: Thermal Calibration

### CALIBRATION RESULTS

| S. No. | Unit Under Calibration in °C@50%RH | Standard Reading in °C@50%RH | Expanded Uncertainty |
|--------|------------------------------------|------------------------------|----------------------|
| 1.     | 15.1                               | 15.04                        | ±0.63°C              |
| 2.     | 25.2                               | 25.10                        |                      |
| 3.     | 35.1                               | 34.97                        |                      |
| 4.     | 45.4                               | 45.11                        |                      |
|        | (%RH) @ 25 °C                      | (%RH) @ 25 °C                | ± 2.8 %RH            |
| 1.     | 20                                 | 20.38                        |                      |
| 2.     | 40                                 | 40.62                        |                      |
| 3.     | 60                                 | 60.79                        |                      |
| 4.     | 81                                 | 80.96                        |                      |

Reference Documents For Calibration : STC/SOP/T-02, DKD-R5-7

Uncertainty of Measurement : The Reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2 such the coverage probability corresponds to approximately 95% confidence level.

**Note:**

- The calibration results reported in this certificate are valid at the time of and under the stated condition of measurement.
- This Report should not be reproduced except in full without our prior permission in writing.
- Calibration Certificate without signature are not valid.

Calibrated By  
Vivek Sati  
Calibration Engineer



Authorized Signatory  
Kishan pal  
(Branch Head)



## Calibration Certificate

Certificate No. **STC/CAL/00110**  
 User Name: M/s Dr. SR Pathology lab  
 Chandigarh Road, Opp. K sera Miniplex  
 Near ITI, Nawanshahr, Pb- 144514

Page No.: 1 of 1

Format No. **STC/7.8F-01**

|                                |                        |
|--------------------------------|------------------------|
| Equip. For Calibration         | Centrifuge Machine     |
| Make/Model No.                 | Laby™                  |
| Range                          | 0 to 3500 rpm          |
| Least Count                    | 10 rpm                 |
| Accuracy                       | N/A                    |
| Id No./Serial No.              | CM/01                  |
| SRF No. & Date of Receipt      | 2021/07/022-14/07/2021 |
| Visual Inspection              | Satisfactory           |
| Date of Calibration            | 15/07/2021             |
| Suggested Due Date             | 14/07/2022             |
| Date of Issue                  | 17/07/2021             |
| Calibrated at                  | Site                   |
| Location                       | --                     |
| <b>Environmental Condition</b> |                        |
| Temperature                    | 25 ± 15°C              |
| Humidity                       | 50 ± 20% RH            |

| Equipment & Master Used For Calibration |                       |
|---|-----------------------|
| 01.                                     | : Digital Tachometer  |
| Make                                    | : Prova               |
| Serial No./Id No.                       | : 10300455/STC/DTM/01 |
| Certificate No.                         | : 2106/0099/01        |
| Next Due Date                           | : 25/05/2022          |
| Traceability                            | : Excellent Services. |

Discipline: RPM Calibration

### CALIBRATION RESULTS

| S. No. | Unit Under Calibration in (RPM) | Standard Reading in (RPM) | Expanded Uncertainty |
|--------|---------------------------------|---------------------------|----------------------|
| 1.     | 1000                            | 1002.4                    | ±7.26 RPM            |
| 2.     | 2000                            | 2003.3                    |                      |
| 3.     | 3000                            | 3004.2                    |                      |
| 4.     | 3500                            | 3505.1                    |                      |

**Reference Documents For Calibration :** STC/SOP/M-09, IS: 12508

**Uncertainty of Measurement :** The Reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2 such that the coverage probability corresponds to approximately 95% confidence level.

**Note:**

1. The calibration results reported in this certificate are valid at the time of and under the stated condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid.

Calibrated by  
 Vivek Sab  
 Calibration Engineer



Authorized Signatory  
 Kishan Pal  
 (Branch Head)



## Calibration Certificate

Certificate No. **STC/CAL/00114**  
 User Name: **M/s Dr. SR Pathology lab**  
**Chandigarh Road, Opp. K sera Miniplex**  
**Near ITI, Nawanshahr, Pb- 144514**

Page No.: 1 of 1

Format No. **STC/7.8F-01**

|                                |                        |
|--------------------------------|------------------------|
| Equip. For Calibration         | Centrifuge Machine     |
| Make/Model No.                 | Laby™                  |
| Range                          | 0 to 3500 rpm          |
| Least Count                    | 10 rpm                 |
| Accuracy                       | N/A                    |
| Id No./Serial No.              | CM/02                  |
| SRF No. & Date of Receipt      | 2021/07/022-14/07/2021 |
| Visual Inspection              | Satisfactory           |
| Date of Calibration            | 15/07/2021             |
| Suggested Due Date             | 14/07/2022             |
| Date of Issue                  | 17/07/2021             |
| Calibrated at                  | Site                   |
| Location                       | --                     |
| <b>Environmental Condition</b> |                        |
| Temperature                    | 25 ± 15°C              |
| Humidity                       | 50 ± 20% RH            |

| Equipment & Master Used For Calibration |                       |
|---|-----------------------|
| 01.                                     | : Digital Tachometer  |
| Make                                    | : Prova               |
| Serial No./Id No.                       | : 10300455/STC/DTM/01 |
| Certificate No.                         | : 2106/0099/01        |
| Next Due Date                           | : 25/05/2022          |
| Traceability                            | : Excellent Services. |

Discipline: RPM Calibration

### CALIBRATION RESULTS

| S. No. | Unit Under Calibration in (RPM) | Standard Reading in (RPM) | Expanded Uncertainty |
|--------|---------------------------------|---------------------------|----------------------|
| 1.     | 1000                            | 1001.2                    | ±7.26 RPM            |
| 2.     | 2000                            | 2002.5                    |                      |
| 3.     | 3000                            | 3003.1                    |                      |
| 4.     | 3500                            | 3504.4                    |                      |

Reference Documents For Calibration : STC/SOP/M-29, IS: 12508

Uncertainty of Measurement : The Reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2 such that the coverage probability corresponds to approximately 95% confidence level.

Note:

1. The calibration results reported in this certificate are valid at the time of and under the stated condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid.

Calibrated By  
  
 Vivek Saini  
 Calibration Engineer



Authorized Signatory  
  
 Kishan Pal  
 (Branch Head)



## Calibration Certificate

Certificate No. STC/CAL/00113

Page No.: 1 of 1

User Name: M/s Dr. SR Pathology lab  
Chandigarh Road, Opp. K sera Miniplex  
Near ITI, Nawanshahr, Pb- 144514

Format No. STC/7.8F-01

|                                |                        |
|--------------------------------|------------------------|
| Equip. For Calibration         | Micro pipette          |
| Make/Model No.                 | Erba                   |
| Range                          | 100 to 1000 µl         |
| Least Count                    | N/A                    |
| Accuracy                       | N/A                    |
| Id No./Serial No.              | OC500083               |
| SRF No.& Date of Receipt       | 2021/07/022-14/07/2021 |
| Visual Inspection              | Satisfactory           |
| Date of Calibration            | 15/07/2021             |
| Suggested Due Date             | 14/07/2022             |
| Date of Issue                  | 17/07/2021             |
| Calibrated at                  | Lab                    |
| Location                       | --                     |
| <b>Environmental Condition</b> |                        |
| Temperature                    | 25 ± 3°C               |
| Humidity                       | 50 ± 20 % RH           |

| Equipment & Master Used For Calibration |                       |
|---|-----------------------|
| 01.                                     | : Std. Weight Box     |
| Make                                    | : Weightronics        |
| Serial No./Id No.                       | : STC/ASW/01          |
| Certificate No.                         | : WMCL/E/2020-02/3262 |
| Next Due Date                           | : 06/02/2023          |
| Traceability                            | : WMCL                |

Discipline: Mechanical Calibration (Mass & Volume)

### CALIBRATION RESULTS

| S. No. | Unit Under Calibration in (µl) | Standard Reading in (µl) | Expanded Uncertainty |
|--------|--------------------------------|--------------------------|----------------------|
| 1.     | 200                            | 200.21                   | ± 35 µl              |
| 2.     | 400                            | 400.36                   |                      |
| 3.     | 600                            | 600.42                   |                      |
| 4.     | 800                            | 800.51                   |                      |

Reference Documents For Calibration :STC/SOP/M-07, IS :8655-6

Uncertainty of Measurement :The Reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2such the coverage probability corresponds to approximately 95% confidence level.

Note:

- 1.The calibration results reported in this certificate are valid at the time of and under the stated condition of measurement.
- 2.This Report should not be reproduced except in full without our prior permission in writing.
- 3.Calibration Certificate without signature are not valid.

Calibrated By  
Vivek Sati  
Calibration Engineer



Authorized Signatory  
Kishan Pal  
(Branch Head)



## Calibration Certificate

Page No.: 1 of 1

Certificate No. **STC/CAL/00111**  
 User Name: **M/s Dr. SR Pathology lab**  
 Chandigarh Road, Opp. K sera Miniplex  
 Near ITI, Nawanshahr, Pb- 144514

Format No. **STC/7.8F-01**

|                                |                         |
|--------------------------------|-------------------------|
| Equip. For Calibration         | Micro pipette           |
| Make/Model No.                 | Erba                    |
| Range                          | 10 to 100 $\mu$ l       |
| Least Count                    | N/A                     |
| Accuracy                       | N/A                     |
| Id No./Serial No.              | YE177AF0040510          |
| SRF No.& Date of Receipt       | 2021/07/022-14/07/2021  |
| Visual Inspection              | Satisfactory            |
| Date of Calibration            | 15/07/2021              |
| Suggested Due Date             | 14/07/2022              |
| Date of Issue                  | 17/07/2021              |
| Calibrated at                  | Lab                     |
| Location                       | --                      |
| <b>Environmental Condition</b> |                         |
| Temperature                    | 25 $\pm$ 3 $^{\circ}$ C |
| Humidity                       | 50 $\pm$ 20 % RH        |

| Equipment & Master Used For Calibration |                       |
|---|-----------------------|
| 01.                                     | : Std. Weight Box     |
| Make                                    | : Weightronics        |
| Serial No./Id No.                       | : STC/ASW/01          |
| Certificate No.                         | : WMCL/E/2020-02/3262 |
| Next Due Date                           | : 06/02/2023          |
| Traceability                            | : WMCL                |

**Discipline: Mechanical Calibration (Mass & Volume)**

### CALIBRATION RESULTS

| S. No. | Unit Under Calibration in ( $\mu$ l) | Standard Reading in ( $\mu$ l) | Expanded Uncertainty |
|--------|--------------------------------------|--------------------------------|----------------------|
| 1.     | 20                                   | 20.02                          | $\pm$ 7.93 $\mu$ l   |
| 2.     | 40                                   | 40.10                          |                      |
| 3.     | 60                                   | 60.16                          |                      |
| 4.     | 80                                   | 80.22                          |                      |

Reference Documents For Calibration :STC/SOP/M-07, IS :8655-6

**Uncertainty of Measurement :**The Reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2such the coverage probability corresponds to approximately 95% confidence level.

**Note:**

- 1.The calibration results reported in this certificate are valid at the time of and under the stated condition of measurement.
- 2.This Report should not be reproduced except in full without our prior permission in writing.
- 3.Calibration Certificate without signature are not valid.

Calibrated By  
**Vivek Sati**  
 Calibration Engineer



Authorized Signatory  
**Kishan Pal**  
 (Branch Head)



## Calibration Certificate

Certificate No. **STC/CAL/00112**  
 User Name: **M/s Dr. SR Pathology lab**  
 Chandigarh Road, Opp. K sera Miniplex  
 Near ITI, Nawanshahr, Pb- 144514

Page No.: 1 of 1

Format No. **STC/7.8F-01**

|                                |                         |
|--------------------------------|-------------------------|
| Equip. For Calibration         | Micro pipette           |
| Make/Model No.                 | --                      |
| Range                          | 5 to 50 $\mu$ l         |
| Least Count                    | N/A                     |
| Accuracy                       | N/A                     |
| Id No./Serial No.              | YE192AL0264125          |
| SRF No.& Date of Receipt       | 2021/07/022-14/07/2021  |
| Visual Inspection              | Satisfactory            |
| Date of Calibration            | 15/07/2021              |
| Suggested Due Date             | 14/07/2022              |
| Date of Issue                  | 17/07/2021              |
| Calibrated at                  | Lab                     |
| Location                       | --                      |
| <b>Environmental Condition</b> |                         |
| Temperature                    | 25 $\pm$ 3 $^{\circ}$ C |
| Humidity                       | 50 $\pm$ 20 % RH        |

| Equipment & Master Used For Calibration |                       |
|---|-----------------------|
| 01.                                     | : Std. Weight Box     |
| Make                                    | : Weightronics        |
| Serial No./Id No.                       | : STC/ASW/01          |
| Certificate No.                         | : WMCL/E/2020-02/3262 |
| Next Due Date                           | : 06/02/2023          |
| Traceability                            | : WMCL                |

**Discipline: Mechanical Calibration (Mass & Volume)**

### CALIBRATION RESULTS

| S. No. | Unit Under Calibration in ( $\mu$ l) | Standard Reading in ( $\mu$ l) | Expanded Uncertainty |
|--------|--------------------------------------|--------------------------------|----------------------|
| 1.     | 5                                    | 5.01                           | $\pm$ 7.93 $\mu$ l   |
| 2.     | 10                                   | 10.06                          |                      |
| 3.     | 30                                   | 30.13                          |                      |
| 4.     | 50                                   | 50.20                          |                      |

**Reference Documents For Calibration :**STC/SOP/M-07, IS :8655-6

**Uncertainty of Measurement :**The Reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$  such the coverage probability corresponds to approximately 95% confidence level.

**Note:**

- 1.The calibration results reported in this certificate are valid at the time of and under the stated condition of measurement.
- 2.This Report should not be reproduced except in full without our prior permission in writing.
- 3.Calibration Certificate without signature are not valid.

Calibrated By  
 Vivek Sahi  
 Calibration Engineer



Authorized Signatory  
 Kishan Pal  
 (Branch Head)