

Patient Name: Ms ANURADHA PATIL

**SPLIT TESTING RECORD**

ABNORMAL SAMPLE

| Date     | Patients ID | Test Parameter | SPIT 1<br>Expected Result | SPIT 2<br>Observed Result | Acceptable CV % | Observed CV% | Comment       | Section Head Sign/Date |
|----------|-------------|----------------|---------------------------|---------------------------|-----------------|--------------|---------------|------------------------|
| 27/12/21 | 46502       | Quantity       | 20 ml                     | 25 ml                     | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | Colour         | yellow                    | yellow.                   | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | Appearance     | cloudy                    | cloudy                    | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | PH             | 6                         | 6                         | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | Sp             | 1.030                     | 1.030                     | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | Protein        | Absent                    | Absent                    | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | Sugar          | +++                       | +++                       | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | Ketone         | Present                   | Present                   | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | Urobilino-gen  | Normal                    | Normal                    | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | DB             | Negative                  | Negative                  | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | White          | Negative                  | Negative                  | NA              | NA           | Satisfactory  | <u>[Signature]</u>     |
|          |             | Bile-salt.     | Negative                  | Negative                  | NA              | NA           | Satisfactory. | <u>[Signature]</u>     |

Acceptable Criteria: As mentioned in QSP

Prepared by: MANIKA JOSHI And ANURITA

Checked by: KARUNA

|                                  |                                     |
|----------------------------------|-------------------------------------|
| Document No: GEN-69              | Document Name: Split Testing Record |
| Issue No: 01                     | Issue Date: 01.02.2020              |
| Amendment No: 00                 | Amendment Date: NA                  |
| ISSUED BY : QA DEPARTMENT        |                                     |
| APPROVED BY : HEAD OF DEPARTMENT |                                     |
| Status : Controlled              | Page 1 of 1                         |

**SPLIT TESTING RECORD**

*NORMAL SAMPLE*

Patient Name: *MR SURAJ K*

| Date   | Patients ID | Test Parameter | Expected Result | Observed Result | Acceptable CV % | Observed CV% | Comment      | Section Head Sign/Date |
|--------|-------------|----------------|-----------------|-----------------|-----------------|--------------|--------------|------------------------|
| 2/7/21 | 46500       | QUANTITY       | Normal          | Normal          | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | COLOUR         | Pale yellow     | Pale yellow     | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | APPEARANCE     | Clear           | Clear           | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | PH             | 6.5             | 6.5             | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | SG             | 1.000           | 1.000           | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | PROTEIN        | Absent          | Absent          | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | SUGAR          | Absent          | Absent          | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | KETONE         | Absent          | Absent          | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | URIBILINURIN   | Normal          | Normal          | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | BIS            | Negative        | Negative        | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | NITRITE        | Negative        | Negative        | NA              | NA           | Satisfactory | <i>[Signature]</i>     |
|        |             | BILE SALT      | Negative        | Negative        | NA              | NA           | Satisfactory | <i>[Signature]</i>     |

Acceptable Criteria: As mentioned in QSP

PERFORMED BY: *MAHESH*

APPROVED BY: *[Signature]*

CHECKED BY: *KARUNA*

|                          |                                     |                                 |
|--------------------------|-------------------------------------|---------------------------------|
| Document No: GEN-69      | Document Name: Split Testing Record | Status: Controlled              |
| Issue No: 01             | Issue Date: 01.02.2020              | Page 1 of 1                     |
| Amendment No: 00         | Amendment Date: NA                  | APPROVED BY: HEAD OF DEPARTMENT |
| ISSUED BY: QA DEPARTMENT |                                     |                                 |

July - 2021.

Normal.

Split - 1 - Mahira

DekaPHAN LAURA  
Seq.No:0013  
ID:SWAPNIL  
COLOR:PALE-Y  
CLARITY:CLEAR  
08.07.2021 17:36

BLD NEG  
LEU NEG  
BIL NEG  
UBG NORM  
KET NEG  
GLU NEG  
PRO NEG  
pH 6  
NIT NEG  
SG 1.030

Mali  
8/07/2021

Normal

Split - 2 - Ananta

DekaPHAN LAURA  
Seq.No:0014  
ID:SWAPNIL-2  
COLOR:PALE-Y  
CLARITY:CLEAR  
08.07.2021 17:39

BLD NEG  
LEU NEG  
BIL NEG  
UBG NORM  
KET NEG  
GLU NEG  
PRO NEG  
pH 6  
NIT NEG  
SG 1.030

Mali  
8/07/2021

Abnormal

Split 1 - Mahira

DekaPHAN LAURA  
Seq.No:0001  
ID:ANURADHA  
COLOR:YELLOW  
CLARITY:CLOUDY  
08.07.2021 16:49

BLD NEG  
\*LEU 25 Leu/ $\mu$ l 1+  
BIL NEG  
UBG NORM  
\*KET 5.2 mg/dl +-  
\*GLU 300 mg/dl 3+  
PRO NEG  
pH 6  
NIT NEG  
SG 1.030

Mali  
8/07/2021

Abnormal

Split - 2 - Ananta

DekaPHAN LAURA  
Seq.No:0006  
ID:ANURADHA  
COLOR:YELLOW  
CLARITY:CLOUDY  
08.07.2021 17:01

BLD NEG  
\*LEU 25 Leu/ $\mu$ l 1+  
BIL NEG  
UBG NORM  
\*KET 5.2 mg/dl +-  
\*GLU 1000 mg/dl 4+  
PRO NEG  
pH 6  
NIT NEG  
SG 1.030

Mali  
8/07/2021