

CERTIFICATE OF CALIBRATION

CUSTOMER DETAILS:
NEYVELI DIAGONSTIC CENTRE

NO.2, Dr.Radhakrishanan salai,block 11

Neyveli-T.s,Cuddalore,

Tamilnadu.

Calibrated at : Onsite

Date of Receipt :13.6.21

Cond. On Receipt:Satisfactory

Date of issue :14.6.21

Certificate No : NDC/2021-22/CAL/1

Date Of calibration : 13.6.21

Asset No: NDC/2021-22/AN/1

Recom. Due Date:14.6.21

DETAILS OF UUC:

Description : CENTRIFUGE

Identification No : CEN -01

Range : 500 to 4000 rpm

Make : REMI

Least Count: -

Serial No: -

Location : -

Accuary: -

Cal\Working Range : -


Model/Type: -

Environmental Details : Temperature 25±5°C Relative Humidity : 40-70%RH


Discipline: Mechanical **parameter:** RPM

CALIBRATION RESULT FOR RPM				
SL.NO	TEST READING (RPM)	STANDARD READING (RPM)	ERROR (RPM)	EXPANDED UNCERTAINTY ±(RPM)
1	500	499.1	0.9	6
2	1000	998.4	1.6	
3	2000	1996	4	
4	4000	3992	8	


The above mentioned instrument has been calibrated using standards manufacturer recommended protocols, using equipments having traceability to National/International Standards . Devices for which there are no National / International Standards calibration standards are measured against in-house performance standard using accepted test procedures. Partial reproduction of the certificate is not permitted.



Calibration Done by



Approved By

 **SAMMSE**