

CERTIFICATE OF CALIBRATION

CUSTOMER DETAILS:

NEYVELI DIAGONSTIC CENTRE
 NO.2, Dr.Radhakrishanan salai,block 11
 Neyveli-T.s,Cuddalore,
 Tamilnadu.

Calibrated at : Lab
 Date of Receipt :13.6.21
 Cond. On Receipt:Satisfactory
 Date of issue :14.6.21

Certificate No : NDC/2021-22/CAL/3
 Asset No: NDC/2021-22/AN/3

Date Of calibration : 13.6.21
 Recom. Due Date:14.6.21

DETAILS OF UUC:

Description : **REFRIGERATOR**
 Range : -
 Least Count: -
 Location : -
 Cal\Working Range : 2 to 8

Identification No : RL/0001
 Make : WHIRLPOOL
 Serial No: -
 Accuary : -
 Model/Type: DC205


Environmental Details : Temperature 25±5°C Relative Humidity : 40-70%RH

THERMAL CALIBRATION


CALIBRATION RESULT

SL.NO	TEST VALUE	STANDARD VALUE (°c)	UNCERTAINTY ±(°c)
1	Refrigerator(2 to 8°c)	3.23	0.19


The above mentioned instrument has been calibrated using standards manufacturer recommended protocols, using equipments having traceability to National/International Standards. Devices for which there are no National / International Standards calibration standards are measured against in-house performance standard using accepted test procedure. Partial reproduction of the certificate is not permitted.



Calibration Done by



SAMMSE
Approved By



SAMMSE