

Patient Name: Mr. Sushant L

SPLIT TESTING RECORD

NORMAL SAMPLES

Date	Patients ID	Test Parameter	Expected Result	Observed Result	Acceptable CV %	Observed CV %	Comment	Section Head Sign/Date
27/11/21	46500	Quantity	Low	Low	NA	NA	Satisfactory	[Signature]
		Colour	Pale yellow	Pale yellow	NA	NA	Satisfactory	[Signature]
		Appearance	Clear	Clear	NA	NA	Satisfactory	[Signature]
		RH	6.5	6.5	NA	NA	Satisfactory	[Signature]
		SH	1.000	1.000	NA	NA	Satisfactory	[Signature]
		PROTEIN	Absent	Absent	NA	NA	Satisfactory	[Signature]
		SUGAR	Absent	Absent	NA	NA	Satisfactory	[Signature]
		KETONE	Absent	Absent	NA	NA	Satisfactory	[Signature]
		URICACID	Normal	Normal	NA	NA	Satisfactory	[Signature]
		OB	Negative	Negative	NA	NA	Satisfactory	[Signature]
		URICATE	Negative	Negative	NA	NA	Satisfactory	[Signature]
		BILE SALT	Negative	Negative	NA	NA	Satisfactory	[Signature]

Acceptable Criteria: As mentioned in QSP

PERFORMED BY: MANISHA [Signature] APPROVED BY: ANKITA [Signature] CHECKED BY: KARUNA [Signature]

Document No: GEN-69	Document Name: Split Testing Record	Status: Controlled
Issue No: 01	Issue Date: 01.02.2020	Page 1 of 1
Amendment No: 00	Amendment Date: NA	
ISSUED BY: QA DEPARTMENT APPROVED BY: HEAD OF DEPARTMENT		

**SPLIT TESTING RECORD**

Patient Name: **Ms ANURADHA ANANDKAR**

**MS ANURADHA ANANDKAR**

Date	Patients ID	Test Parameter	SPUT 1 Expected Result	SPUT 2 Observed Result	Acceptable CV %	Observed CV%	Comment	Section Head Sign/Date
2/7/21	46502	Quantity	20 ml	25 ml	NA	NA	Satisfactory	[Signature]
		Colour	yellow	yellow.	NA	NA	Satisfactory	[Signature]
		Appearance	cloudy	cloudy	NA	NA	Satisfactory	[Signature]
		PH	6	6	NA	NA	Satisfactory	[Signature]
		Sp	1.030	1.030	NA	NA	Satisfactory	[Signature]
		protein	Absent	Absent	NA	NA	Satisfactory	[Signature]
		Sugar	+++	+++	NA	NA	Satisfactory	[Signature]
		ketone	Present	Present	NA	NA	Satisfactory	[Signature]
		Urobilinogen	Normal	Normal	NA	NA	Satisfactory	[Signature]
		OB	Negative	Negative	NA	NA	Satisfactory	[Signature]
		Nitrite	Negative	Negative	NA	NA	Satisfactory	[Signature]
		Bile salt	Negative	Negative	NA	NA	Satisfactory.	[Signature]

Acceptable Criteria: As mentioned in QSP Performed by: **MANIKA JOSHI** And **ANVITA AND**

Checked by: **KRISHNA KUMAR**

Document No: GEN-69	Document Name: Split Testing Record	Status: Controlled
Issue No: 01	Issue Date: 01.02.2020	Page 1 of 1
Amendment No: 00	Amendment Date: NA	APPROVED BY: HEAD OF DEPARTMENT
ISSUED BY: QA DEPARTMENT		

July - 2021.

Normal.  
Split - 1 - Mahira

DekaPHAN LAURA  
Seq.No:0013  
ID:SWAPNIL  
COLOR:PALE-Y  
CLARITY:CLEAR  
08.07.2021 17:36

BLD NEG  
LEU NEG  
BIL NEG  
UBG NORM  
KET NEG  
GLU NEG  
PRO NEG  
pH 6  
NIT NEG  
SG 1.030

Performed by  
Mahi  
8/07/2021

Normal  
Split - 2 - Ankita

DekaPHAN LAURA  
Seq.No:0014  
ID:SWAPNIL-2  
COLOR:PALE-Y  
CLARITY:CLEAR  
08.07.2021 17:39

BLD NEG  
LEU NEG  
BIL NEG  
UBG NORM  
KET NEG  
GLU NEG  
PRO NEG  
pH 6  
NIT NEG  
SG 1.030

Performed by  
AN  
8/07/2021

Abnormal  
Split 1 - mahira

DekaPHAN LAURA  
Seq.No:0001  
ID:ANURADHA  
COLOR:YELLOW  
CLARITY:CLOUDY  
08.07.2021 16:49

BLD NEG  
\*LEU 25 Leu/ $\mu$ l 1+  
BIL NEG  
UBG NORM  
\*KET 5.2 mg/dl +-  
\*GLU 300 mg/dl 3+  
PRO NEG  
pH 6  
NIT NEG  
SG 1.030

Performed by  
ANURADHA  
8/07/2021

Abnormal  
Split - 2 - Ankita

DekaPHAN LAURA  
Seq.No:0006  
ID:ANURADHA  
COLOR:YELLOW  
CLARITY:CLOUDY  
08.07.2021 17:01

BLD NEG  
\*LEU 25 Leu/ $\mu$ l 1+  
BIL NEG  
UBG NORM  
\*KET 5.2 mg/dl +-  
\*GLU 1000 mg/dl 4+  
PRO NEG  
pH 6  
NIT NEG  
SG 1.030

Performed by  
AN  
08/07/2021

PERFORMED BY : MAHIRA

CHECKED BY : KARUNA

PERFORMED BY : ANKITA

CHECKED BY : KARUNA