



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF REFRIGERATOR

Calibration Certificate No.	21100010085	Issued On:	10/08/2021
ULR No.	CC214621100010085F	Calibration Date:	07/08/2021
Job Identification No.	ICS/C/T/08/10085	Next Calibration Date:	06/08/2022
Ref. No.	SRF- Dated: 04/08/2021		

CALIBRATED FOR:	Dr. Dham's Diagnostic Centre 3856/23, Reghar Pura, Karol Bagh, New Delhi-110005.
------------------------	---

EQUIPMENT DESCRIPTION

Name	Refrigerator		
Make/Model	Celfrost / FKG320	Visual Inspection	Ok
Serial No.	FKG320160600256	I.D. No.	---
Range	2°C to 8°C	Least Count	---
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National Standard (NPL, New Delhi)

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Multi-Channel Data-Logger with RTD Sensor	Countron	CC28882000000956F	Sigma, Delhi	17/09/2020	17/09/2021

USED EQUIPMENT DETAILS	Not Applicable
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-TH04 (ITS-90) (Measurement Method)

RESULTS: Thermal Calibration

Location Details	Corrected Value of Standard (Average)	Uncertainty at Approx 95% C.L and Coverage Factor K=2	Location Diagram
Location - 1 (L - 1)	2.4 °C	± 1.2°C	
Location - 2 (L - 2)	2.7 °C	± 1.2°C	
Location - 3 (L - 3)	3.4 °C	± 1.2°C	
Location - 4 (L - 4)	3.6 °C	± 1.2°C	

REMARKS: The recommended date for next validation is mentioned, as asked by the customer.

*** END OF REPORT ***



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF REFRIGERATOR

Calibration Certificate No.	21100010086	Issued On:	10/08/2021
ULR No.	CC214621100010086F	Calibration Date:	07/08/2021
Job Identification No.	ICS/C/T/08/10086	Next Calibration Date:	06/08/2022
Ref. No.	SRF- Dated: 04/08/2021		

CALIBRATED FOR: Dr. Dham's Diagnostic Centre
 3856/23, Reghar Pura,
 Karol Bagh, New Delhi-110005.

EQUIPMENT DESCRIPTION

Name	Refrigerator		
Make/Model	Whirlpool / DC215F/2017	Visual Inspection	Ok
Serial No.	---	I.D. No.	---
Range	2°C to 8°C	Least Count	---
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National Standard (NPL, New Delhi)

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Multi-Channel Data-Logger with RTD Sensor	Countron	CC28882000000956F	Sigma, Delhi	17/09/2020	17/09/2021

USED EQUIPMENT DETAILS	Not Applicable
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-TH04 (ITS-90) (Measurement Method)

RESULTS: Thermal Calibration

Location Details	Corrected Value of Standard (Average)	Uncertainty at Approx 95% C.L and Coverage Factor K=2	Location Diagram
Location - 1 (L - 1)	2.4 °C	± 1.2°C	
Location - 2 (L - 2)	2.6 °C	± 1.2°C	
Location - 3 (L - 3)	3.1 °C	± 1.2°C	
Location - 4 (L - 4)	3.4 °C	± 1.2°C	

REMARKS: The recommended date for next validation is mentioned, as asked by the customer.

*** END OF REPORT ***

1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

Authorised Signatory
RAJESH KAPOOR
 Operation-Head